

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-25.18 revision 0  
Project number 3909768  
Page 1 of 2

Issuing authority

NMi Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant and  
Manufacturer

SELVAS Healthcare, Inc  
155, Sinseong-ro, Yuseong-gu  
Daejeon 34109  
South-Korea

Identification of the  
certified type

**A Non-automatic weighing instrument**  
Type : ACCUNIQ BC380 series and ACCUNIQ BC390 series

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76-1:2006** for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
17 November 2025

Certification Board

NMi Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-25.18 revision 0  
Project number 3909768  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-3909768-01 dated 17 November that includes 44 pages;
- No. NMI-3909768-02 dated 17 November that includes 19 pages.

### Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	Max $\leq$ 250 kg
Verification scale interval	e $\geq$ 100 g
Weighing range	Single interval
Maximum number of scale intervals	n $\leq$ 2500
Temperature range	- 10 °C / + 40 °C
Power supply voltage	12 V DC from 100-240 V AC 50/60 Hz plugin power adapter
Application	Intended to be used for weighing patients
Software identification	ACCUNIQ BC380 Series : BC380.m.x.y.z ACCUNIQ BC390 Series : BC390.m.x.y.z (m = A ... ZZZ, x, y, z = 0 .. 99)

The software identification is displayed at start-up.

### Revision History

Revision	Date	Change(s)
0	2025-11-17	Initial issue.