

OIML Member State
The Netherlands

Number R76/2006-A-NL1-25.18 revision 1
Project number 3788911
Page 1 of 3

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

Teraoka Seiko Co., Ltd.
5-13-12, Kugahara, Ohta-ku
146-8580 Tokyo
Japan

Identification of the
certified type

A Non-automatic weighing instrument
Type : DPS-6600 series

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
23 April 2026

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



OIML Member State
The Netherlands

Number R76/2006-A-NL1-25.18 revision 1
Project number 3788911
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-3788911-01 dated 25 September 2025 that includes 35 pages;
- No. NMI-3788911-03 dated 25 September 2025 that includes 22 pages;
- No. NMI-3788911-05 dated 25 September 2025 that includes 12 pages;
- No. NMI-3788911-07 dated 25 September 2025 that includes 20 pages;
- No. NMI-3788911-06 dated 23 April 2026 that includes 30 pages;
- No. NMI-3788911-08 dated 23 April 2026 that includes 16 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	(III)	
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 75 \text{ kg}$	
Verification scale interval	$e \geq 1 \text{ g}$	
Weighing range(s)	Single interval Multi-interval	
Maximum number of scale intervals (one weighing range)	$n \leq 3000$	
Maximum number of scale intervals (multi-interval)	$n \leq 3000$ (per partial weighing range)	
Maximum number of partial weighing ranges	2	
Tare	$T \leq -\text{Max}$ for single interval instruments $T \leq -\text{Max}_1 + e_1$ for multi-interval instruments	
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$	
Power supply voltage	100 –240V AC 50/60 Hz	
Application	Intended to be used for the making-up of prepackages	
Software identification	Measurement software	01.xx (x = 0... 9 for non-legally relevant changes)
	A/D software	03.xx (x = 0... 9 for non-legally relevant changes)

The software identification is displayed at the start up or after selecting: 'SETUP' → 'SOFTWARE VERSION'.

OIML Member State
The Netherlands

Number R76/2006-A-NL1-25.18 revision 1
Project number 3788911
Page 3 of 3

Revision History

+ This revision replaces the previous version.

Revision	Date	Change(s)
0	2025-09-25	Initial issue.
1	2026-04-23	Changing verification scale interval to $e \geq 1$ g, maximum capacity to 3 kg, and adding alternative mainboard.