



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-25.22 revision 0
Project number 2597918
Page 1 of 3

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

SENSOCAR S.A.U.
Pol. Ind. Can Parellada -c/ Géminis, 77
08228 Terrassa - Barcelona
Spain

Identification of the
certified type

An **Indicator**
Type : KRONOS

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class (III) or (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
7 November 2025

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



OIML Member State
The Netherlands

Number R76/2006-A-NL1-25.22 revision 0
Project number 2597918
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-2597918-01 dated 7 November 2025 that includes 40 pages;
- No. NMI-2597918-02 dated 7 November 2025 that includes 16 pages.

Characteristics of the indicator:

Accuracy class	III or IIII
Weighing range(s)	Single interval Multi-interval
Maximum number of scale intervals (one weighing range)	$n \leq 6000$
Maximum number of scale intervals (multi-interval)	$n \leq 3000$ (per partial weighing range)
Maximum number of partial weighing ranges	2
Load cell excitation voltage	5 V DC
Minimum signal input voltage	$U_{\min} = 0 \text{ mV}$
Minimum input voltage per verification scale interval	0,5 μV
Minimum load cell resistance	40 Ω
Maximum load cell resistance	1050 Ω
Fraction of the maximum permissible error	0,5
Load cell interface	6-wire with sense technology
Maximum value of the cable length per cross wire section between the indicator and the junction box or load cells	1956 m/mm ²
Temperature range	-10 °C / +40 °C
Electromagnetic environment class	E2
Power supply voltage	8 - 28 V DC
Software identification	YYXXX-XXXXY-001, where Y = A...Z and X = 0...9

Software:

- The identification number can be checked in the "VERSIO" option, which is found in the "DISPLAY" submenu of the settings menu;
- Alternatively, the identification number will be displayed at start-up;
- The indicator has embedded software.



OIML Member State
The Netherlands

OIML Certificate

Number R76/2006-A-NL1-25.22 revision 0
Project number 2597918
Page 3 of 3

Revision History

Revision	Date	Change(s)
0	2025-11-07	Initial issue.