



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-25.23 revision 0
Project number 3968614
Page 1 of 2

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Manufacturer

Dibal S.A.
Astintze Kalea, 26 - Pol. Ind. Neinver
48160 Derio, Vizcaya
Spain

Identification of the
certified type

A Non-automatic weighing instrument
Type : NGDPOS

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class (III) or (IIII)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
2 December 2025

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



OIML Member State
The Netherlands

Number R76/2006-A-NL1-25.23 revision 0
Project number 3968614
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated report(s):

- No. NMI-3968614-01 dated 20 November 2025 that includes 43 pages;
- No. NMI-3968614-02 dated 20 November 2025 that includes 21 pages;
- No. NMI-3968614-03 dated 20 November 2025 that includes 19 pages;
- No. NMI-3968614-04 dated 20 November 2025 that includes 19 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III or IIII
Verification scale interval	$e \geq 2 \text{ g}$
Weighing ranges	Single interval Multiple range
Maximum number of scale intervals (one weighing range)	$n \leq 3000$
Maximum number of scale intervals (multiple range)	$n \leq 3000$ (per weighing range)
Maximum number of weighing ranges	2
Temperature range	-10 °C / 40 °C
Power supply voltage	100 – 240 V AC 50/60 Hz to supply 5 V DC (DPOS Format), or 12 V DC (G Format). 6 V DC (G Format battery) (not suitable for a road vehicle power supply)
Application	Intended to be used for direct sales to the public
Software identification	V-1.00

Software:

- The identification number will be displayed at start-up;
- The indicator has embedded software.

Revision History

Revision	Date	Change(s)
0	2025-12-02	Initial issue.