

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-26.27 revision 0  
Project number 4045467  
Page 1 of 2

Issuing authority

NMi Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant and  
Manufacturer

KERN & SOHN GmbH  
Ziegelei 1  
72336 Balingen-Frommern  
Germany

Identification of the  
certified type

**A Non-automatic weighing instrument**  
Type : TPGJ-A

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76-1:2006** for accuracy class **II**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
16 February 2026

Certification Board

NMi Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-26.27 revision 0  
Project number 4045467  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-13200783-01 dated 24 April 2014 that includes 43 pages;
- No. NMI-13200783-02 dated 24 April 2014 that includes 20 pages;
- No. NMI-13200783-03 dated 24 April 2014 that includes 30 pages;
- No. NMI-13200783-04 dated 12 March 2014 that includes 17 pages;
- No. NMI-15200108-01 dated 2 April 2015 that includes 12 pages;
- No. NMI-15200108-02 dated 2 April 2015 that includes 16 pages;
- No. NMI-3753602-01 dated 20 September 2024 that includes 20 pages;
- No. NMI-3753602-02 dated 20 September 2024 that includes 9 pages;
- No. NMI-4045467-01 dated 16 February 2026 that includes 10 pages.

### Characteristics of the non-automatic weighing instrument:

Accuracy class	II
Maximum capacity	620 g ≤ Max ≤ 3200 g or 3100 ct ≤ Max ≤ 16000 ct
Verification scale interval	e = 0,1 g / 0,01 g or e = 1 ct / 0,1 ct
Actual scale interval	e = d
Weighing range	Single interval
Maximum number of scale intervals	n ≤ 62000
Tare	T ≤ -Max
Temperature range	+5 °C / +40 °C
Power supply voltage	6 V DC through 100 – 240 V 50/60 Hz AC/DC adapter or by 4x 1,5 V AA size battery
Application	Intended to be used for direct sales to the public
Software identification	Checksum: 5A8b

### Revision History

Revision	Date	Change(s)
0	2026-02-16	Initial issue.