

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-26.28 revision 0  
Project number 4039408  
Page 1 of 3

Issuing authority

NMi Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant

Mettler-Toledo GmbH  
Im Langacher 44  
8606 Greifensee,  
Switzerland

Manufacturer

Mettler-Toledo (Changzhou) Measurement Technology Ltd.  
No.111, West TaiHu Road,  
Changzhou,  
Jiangsu, 213125,  
China

Identification of the certified type

**A Non-automatic weighing instrument**  
Type : VTS..., VTC..., VRS...

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76-1:2006** for accuracy class  $\textcircled{\text{III}}$  or  $\textcircled{\text{III}}$

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

This certificate and supporting reports comply with the requirements of OIML-CS-PD-07 clause 6.2.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
16 April 2026

Certification Board

NMi Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-26.28 revision 0  
Project number 4039408  
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated reports:

**Indicator/Terminal:**

Type IND246 with OIML certificate R76/2006-A-NL1-23.23R1:

- No. NMI-11200016-01 dated 18 May 2011 that includes 47 pages;
- No. NMI-11200016-02 dated 18 May 2011 that includes 38 pages;
- No. NMI-11200016-03 dated 18 May 2011 that includes 13 pages;
- No. NMI-13200507-01 dated 06 November 2013 that includes 27 pages;
- No. NMI-15200477-01 dated 29 October 2015 that includes 7 pages;
- No. NMI-2372754-01 dated 12 December 2019 that includes 9 pages;
- No. NMI-3756754-02 dated 06 June 2024 that includes 17 pages;
- No. NMI-3756754-04 dated 06 June 2024 that includes 9 pages;

**Analog Load Cell:**

Type: 0782 with OIML certificate R60/2000-NL1-01.04R1:

- No. NMI-219471-02 dated 09 August 2011 that includes 31 pages.

### Characteristics of the non-automatic weighing instrument:

Accuracy class	III or IIII
Maximum capacity	10 t ≤ Max ≤ 400 t
Verification scale interval	e ≥ 5 kg
Weighing ranges	Single interval Multiple range
Maximum number of scale intervals (one weighing range)	n ≤ 3000
Maximum number of scale intervals (multiple range)	n ≤ 3000 (per weighing range)
Maximum number of weighing ranges	2
Tare	T ≤ -100% for instruments with one weighing range T ≤ -Max <sub>2</sub> for multi-range instruments
Temperature range	-10 °C / 40 °C
Power supply voltage	100 – 240 V AC 50/60 Hz
Software identification	Version number RainBow V1.6 1.xx Where xx can be a number between 0 and 99.
	Type Analog 72257764



# OIML Certificate

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-26.28 revision 0  
Project number 4039408  
Page 3 of 3

## Revision History

Revision	Date	Change(s)
0	2026-04-16	Initial issue.