

OIML Member State
The Netherlands

Number R76/2006-A-NL1-26.29 revision 0
Project number 4039408
Page 1 of 3

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant

Mettler-Toledo GmbH
Im Langacher 44
8606 Greifensee,
Switzerland

Manufacturer

Mettler-Toledo (Changzhou) Measurement Technology Ltd.
No.111, West TaiHu Road,
Changzhou,
Jiangsu, 213125,
China

Identification of the certified type

A Non-automatic weighing instrument
Type : VTS..., VTC..., VRS...

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **(III)** or **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
16 April 2026

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



OIML Member State
The Netherlands

Number R76/2006-A-NL1-26.29 revision 0
Project number 4039408
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated reports:

Indicator IND570 with OIML certificate R76/2006-A-NL1-20.27R3:

- No. NMI-13200606-01 dated 17 April 2014 that includes 53 pages;
- No. NMI-13200606-02 dated 17 April 2014 that includes 17 pages;
- No. NMI-15200584-01 dated 16 June 2016 that includes 24 pages;
- No. NMI-2372754-01 dated 12 December 2019 that includes 9 pages;
- No. NMI-2395635-01 revision 1 dated 17 September 2020 that includes 20 pages.

Digital load cell GDD/SLC720 with OIML Certificate: R60/2000-A-NL1-22.17R0:

- No. NMI-10201060-01 dated 4 December 2012 that includes 69 pages;
- No. NMI-2611823-01 dated 13 May 2022 that includes 23 pages.

Digital load cell PDX/SLC820 with OIML Certificate: R60/2000-A-NL1-19.07R1:

- No. R60/2000-NL1-09.08 dated 11 June 2009 that includes 53 pages;
- No. NMI-15200551-01 dated 15 April 2016 that includes 15 pages;
- No. NMI-2611823-02 dated 13 May 2022 that includes 23 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	Ⓜ or ⓂⓂ
Maximum capacity	10 t ≤ Max ≤ 400 t
Verification scale interval	e ≥ 5 kg
Weighing ranges	Single interval Multiple Range
Maximum number of scale intervals (one weighing range)	n ≤ 6000
Maximum number of partial weighing ranges	2
Maximum number of scale interval (Multiple range)	n ≤ 6000 (per weighing range)
Temperature range	Matching the temperature range in the test reports involved
Power supply voltage	100 – 240 V AC 50/60 Hz 24 V DC
Software identification Version number (POWERCELL Scale)	1.xx.yyyy 2.xx.yyyy 3.xx.yyyy 4.xx.yyyy
	xx is a number between 00 and 99 representing major updates of the non-legally relevant part of the software and yyyy is a number between 0000 and 9999 and represents minor updates of the non-legally relevant part of the software.



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-26.29 revision 0
Project number 4039408
Page 3 of 3

Revision History

Revision	Date	Change(s)
0	2026-04-16	Initial issue.