

OIML Member State
The Netherlands

Number R76/2006-A-NL1-26.32 revision 0
Project number 3945425
Page 1 of 2

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

CAS Corporation
#262, Geurugogae-ro,
Gwangjeok-myeon
Yangju-si, Gyeonggi-do, 11415
Republic of Korea

Identification of the
certified type

A Non-automatic weighing instrument
Type : PR2N - xxyyy with xx = capacity of the scale
and yyy = CB, CP, CU, EB, EP, EU, CD, ED, EDU (model
variant)

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
16 April 2026

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
the Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



OIML Member State
The Netherlands

Number R76/2006-A-NL1-26.32 revision 0
Project number 3945425
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-3945425-01 dated 2 March 2026 that includes 50 pages;
- No. NMI-3945425-02 dated 2 March 2026 that includes 36 pages;
- No. NMI-3945425-03 dated 2 March 2026 that includes 14 pages;
- No. NMI-3945425-04 dated 2 March 2026 that includes 12 pages;
- No. NMI-3945425-05 dated 2 March 2026 that includes 11 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$3 \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 0,5$
Weighing ranges	Single interval Multi-interval
Maximum number of scale intervals (one weighing range)	$n \leq 3000$
Maximum number of scale intervals (multi-interval)	$n \leq 3000$ (per partial weighing range)
Maximum number of weighing ranges	2
Tare	$T \leq -\text{Max}_2$ for instruments with one weighing range $T \leq -\text{Max}_1 - e_1$ for multi-interval instruments
Temperature range	$-10 \text{ }^\circ\text{C} / + 40 \text{ }^\circ\text{C}$
Power supply voltage	100 – 240 V AC 50/60 Hz to 6 V DC by A/C adapter Integrated Pb 4V/4Ah battery 3 x 1.5 V dry battery (D type)
Application	Intended to be used for direct sales to the public
Software identification Version number	V 5xx or yy5xx with non legally relevant parts: xx = any character reflecting software changes, and yy = any character reflecting the country code (y,x = 0...99)

The software identification is displayed at start-up.

Revision History

Revision	Date	Change(s)
0	2026-04-16	Initial Issue