OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-26.42 revision 0
Project number 4151770

Issuing authority	NMi Certin B.V. Person responsible: M.Ph.D. Schmidt
Applicant	Yamato Scale Co., Ltd. 5 – 22 Saenba-cho Akashi, 673-8688 Japan
Manufacturer	Shanghai Yamato Scale Co., Ltd. 368, Qingda Road, Heqing Industrial Field, Pudong, Shanghai 201201, China
Identification of the certified type	A **Non-automatic weighing instrument** Type : DP-6701 series
Characteristics	See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
8 October 2026

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.

OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-26.42 revision 0
Project number 4151770

The conformity was established by the results of tests and examinations provided in the associated reports:
- No. NMi-2551350-01 dated 7 June 2021 that includes 58 pages.

Characteristics of the non-automatic weighing instrument:

The construction of the non-automatic weighing instrument is recorded in the Documentation folder no. R76-26.42-1.

Accuracy class		(III)	
Maximum capacity		30 kg	32 kg ≤ Max ≤ 150 kg
Verification scale interval		e ≥ 0,01 kg	
Weighing range		Single interval	
Maximum number of scale intervals (one weighing range)		n ≤ 3000	n ≤ 3200
Tare		T ≤ -Max	
Temperature range		-10 °C / +40 °C	
Power supply voltage		3,7 V – 6 V DC by non-rechargeable battery, or 100 – 240 V AC/DC to 5,9 V DC by plug-in adapter	
Software identification	Version number	V1.xx. (x = 0… 99)	

The software identification is displayed at start-up.

Revision History

Revision	Date	Change(s)
0	2026-10-08	Initial issue.