OIML Certificate

OIML Member State
The Netherlands

Number R99/2008-A-NL1-26.01 revision 0
Project number 4022686

Issuing authority	NMi Certin B.V.
Person responsible	M.Ph.D. Schmidt
Applicant and Manufacturer	Test Equipment Nederland B.V. Rijksstraatweg 45 1396 JD Baambrugge The Netherlands
Identification of the certified type	**Instrument for measuring vehicle exhaust emissions** Type : INNOVA 600 INNOVA 6000
Characteristics	See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 99-1 & 2 - Edition 2008 (E)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
8 October 2026

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.

OIML Certificate

OIML Member State
The Netherlands

Number R99/2008-A-NL1-26.01 revision 0
Project number 4022686

The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Report(s):
- No. NMi-3788687-01 Revision 1 dated 18 December 2025 that includes 38 pages;
- No. NMi-4022686-01 dated 8 October 2026 that includes 24 pages;
- No. NMi-4022686-02 dated 8 October 2026 that includes 18 pages.

Characteristics of the instrument for measuring vehicle exhaust emissions:

Accuracy class		0, I	
Electromagnetic environment class		E2	
Mechanical environment class		M1	
Climatic environment	temperature range	+5 °C / +40 °C	
	humidity	non-condensing	
	location	closed	
Warm-up time		≤ 15 min	
Minimal flow		3,5 L/min	
Power supply voltage		230 V_{AC} 50 Hz	
Adjustment period		12 months	
Indication range and resolution	CO	0 – 10 % vol	0,01 % vol
	CO_2	0 – 20 % vol	0,1 % vol
	O_2	0 – 4 % vol	0,01 % vol
		4 - 25 % vol	0,1 % vol
	HC	0 - 10000 ppm vol	1 ppm vol
	lambda	0,600 – 9,999	0,001

Software	INNOVA 600	INNOVA 6000	
		Measuring unit	Universal computer
part	firmware	firmware	SSM32.dll
version	1.00	1.00	1.00
identification code	AC48	AC48	FFD8...092D
display	User menu / info. sys.	Benzine menu / Systeeminfo	

The construction is described in Documentation folder number T13310-1.

Certificate history:

Revision	Date	Description of the modification
0	8 October 2026	Initial issue.