

**ERRATUM
(2010.08.09)**

OIML R 51-1
Edition 2006 (E)

**Automatic catchweighing instruments.
Part 1: Metrological and technical requirements - Tests**

Instruments de pesage trieurs-étiqueteurs à fonctionnement automatique.
Partie 1 : Exigences métrologiques et techniques - Essais



**ORGANISATION INTERNATIONALE
DE MÉTROLOGIE LÉGALE**

**INTERNATIONAL ORGANIZATION
OF LEGAL METROLOGY**

The published version of OIML R 51-1 (Edition 2006), currently reads:

A.5.6.2.1 Static tare

Allow the tare device to operate, then increment the tare load by using change point weights until the indication has definitely changed by one scale interval. Verify by the method of A.3.10.2.1 that the tare setting accuracy is better than $\pm 0.25 e$ with a deviation of not more than $0.25 e$.

This should be corrected to read:

A.5.6.2.1 Static tare

Allow the tare device to operate, then increment the tare load by using change point weights until the indication has definitely changed by one scale interval. Verify by the method of A.3.10.2.1 that the tare setting accuracy is better than $\pm 0.25 e$.