

AMENDMENT
(2019-02-07)

OIML R 60-3

Edition 2017 (E)

Metrological regulation for load cells

Part 3: Test report format

Réglementation métrologique des cellules de pesée

Partie 3 : Format du rapport d'essais



Amendment (2019) to OIML R 60-3:2017
Metrological regulation for load cells.
Part 3: Test report format

Form 6.16.1 Electrostatic discharge – direct application

Form 6.16.1 of the published version of OIML R 60-3:2017 (page 85) currently reads:

- Notes:*
- 1) If the load cell fails, the test point at which this occurs shall be recorded.
 - 2) IEC Publication 61000-4-2 (2008) Ed 1.1 Consolidated edition specifies that the test be conducted with the most sensitive polarity.

The reference in Note 2 should be corrected to read:

- Notes:*
- 1) If the load cell fails, the test point at which this occurs shall be recorded.
 - 2) IEC 61000-4-2 Ed. 2.0 (2008-12) Consolidated edition specifies that the test be conducted with the most sensitive polarity.

Form 6.16.2 Electrostatic discharge – indirect application

Form 6.16.2 of the published version of OIML R 60-3:2017 (page 86) currently reads:

- Notes:*
- 1) If the load cell fails, the test point at which this occurs shall be recorded.
 - 2) IEC Publication 61000-4-2 (1999-05) Ed 1.1 Consolidated edition specifies that the test be conducted with the most sensitive polarity.

The reference in Note 2 should be corrected to read:

- Notes:*
- 1) If the load cell fails, the test point at which this occurs shall be recorded.
 - 2) IEC 61000-4-2 Ed. 2.0 (2008-12) Consolidated edition specifies that the test be conducted with the most sensitive polarity.