

ORGANISATION INTERNATIONALE
DE MÉTROLOGIE LÉGALE



INTERNATIONAL RECOMMENDATION

Nonautomatic weighing instruments
Part 1: Metrological and technical requirements - Tests

Instruments de pesage à fonctionnement non automatique
Partie 1: Exigences métrologiques et techniques - Essais

OIML R 76-1

Edition 1992 (E)

Note: This pdf document INCLUDES:

- Amendment 1 (1994)
- the 1992 erratum
- the Index of Terms

CONTENTS

Foreword		4
T Terminology (terms and definitions)		5
T.1 General definitions		5
T.2 Construction of an instrument		6
T.3 Metrological characteristics of an instrument		10
T.4 Metrological properties of an instrument		12
T.5 Indications and errors		12
T.6 Influences and reference conditions		16
T.7 Performance test		16
1 Scope		18
2 Principles of the Recommendation		18
2.1 Units of measurement		18
2.2 Principles of the metrological requirements		18
2.3 Principles of the technical requirements		18
2.4 Application of requirements		19
2.5 Terminology		19
3 Metrological requirements		19
3.1 Principles of classification		19
3.2 Classification of instruments		20
3.3 Additional requirements for a multi-interval instrument		21
3.4 Auxiliary indicating devices		22
3.5 Maximum permissible errors		23
3.6 Permissible differences between results		25
3.7 Verification standards		26
3.8 Discrimination		26
3.9 Variations due to influence quantities and time		27
3.10 Pattern evaluation tests		29
4 Technical requirements for a self- or semi-self-indicating instrument		30
4.1 General requirements of construction		30
4.2 Indication of weighing results		31
4.3 Analogue indicating devices		32
4.4 Digital indicating and printing devices		35
4.5 Zero-setting and zero-tracking devices		36
4.6 Tare devices		37
4.7 Preset tare devices		39
4.8 Locking positions		40
4.9 Auxiliary verification devices (removable or fixed)		40
4.10 Selection of weighing ranges on a multiple range instrument		40
4.11 Devices for selection (or switching) between various load receptors-load transmitting devices and various load measuring devices		41
4.12 Requirements for load cells		41
4.13 "Plus" and "minus" comparator instrument		42
4.14 Instrument for direct sales to the public		43
4.15 Additional requirements for an instrument for direct sales to the public with price indication		45
4.16 Instrument similar to one normally used for direct sales to the public		47
4.17 Price-labelling instrument		47
4.18 Mechanical counting instrument with unit-weight receptor		48

5	Requirements for electronic instruments	48
5.1	General requirements	48
5.2	Acting upon significant faults	49
5.3	Functional requirements	49
5.4	Performance and span stability tests	50
6	Technical requirements for a non-self-indicating instrument	51
6.1	Minimum sensitivity	51
6.2	Acceptable solutions for indicating devices	51
6.3	Conditions of construction	53
6.4	Simple equal arm beam	54
6.5	Simple 1/10 ratio beam	54
6.6	Simple sliding poise instrument (steelyard)	54
6.7	Roberval and Beranger instruments	56
6.8	Instruments with ratio platforms	56
6.9	Instrument with a load-measuring device with accessible sliding poises (of the steelyard type)	57
7	Marking of an instrument	58
7.1	Descriptive markings	58
7.2	Verification marks	61
8	Metrological controls	61
8.1	Liability to metrological controls	61
8.2	Pattern approval	62
8.3	Initial verification	63
8.4	Subsequent metrological control	64
Annex A	Testing procedures for nonautomatic weighing instruments	65
A.1	Administrative examination	65
A.2	Compare construction with documentation	65
A.3	Initial examination	65
A.4	Performance tests	65
A.5	Influence factors	73
A.6	Endurance test	76
Annex B	Additional tests for electronic instruments	78
B.1	General requirements for electronic instruments under test (EUT)	78
B.2	Performance tests for influence factors	78
B.3	Performance tests for disturbances	79
B.4	Span stability test	81
	Bibliography	83
	Index of terms defined in the terminology	see separate sheet

FOREWORD

The Organisation Internationale de Métrologie Légale (OIML) is a world-wide, intergovernmental organization whose main task is that of harmonizing the regulations and metrological controls applied by the national metrological services, or related organizations, of its Member States.

The two main categories of OIML publications are:

- International Recommendations (OIML R), which are model regulations generally establishing the metrological characteristics required of the measuring instruments concerned and specifying methods and equipment for checking their conformity; the OIML Member States shall implement these Recommendations as far as possible,
- International Documents (OIML D), of an informative nature, to help and improve the work of the metrological services.

Draft Recommendations and Documents of OIML are developed by Pilot Secretariats and Reporting Secretariats formed of Member States, in consultation with the international institutions (world-wide and regional) concerned.

The cooperative agreements between OIML and certain institutions (particularly ISO and IEC) aim at avoiding contradictory requirements so that manufacturers and users of measuring instruments, test laboratories etc., may apply simultaneously the OIML publications and the publications of other institutions.

International Recommendations and International Documents are published in French (F) and English (E) and are subject to periodic revision. The words "Edition..." refer to the year in which the document is printed.

OIML publications may be obtained from:

Bureau International de Métrologie Légale
11, rue Turgot - 75009 Paris - France
Telephone: 33 (1) 48 78 12 82 and 42 85 27 11
Fax: 33 (1) 42 82 17 27
Telex: 234 444 SVP SERV F ATTN OIML

*
* *

This publication in two parts – references OIML R 76-1, edition 1992 (E) and OIML R 76-2, edition 1992 (E) – was developed by the Reporting Secretariat SP 7-Sr 4 "Nonautomatic weighing instruments", the Pilot Secretariat SP 7 "Measurement of mass" and by the "Nordic Countries Working Group", attached to SP 7 and author of Annex A and the Pattern Evaluation Report (R 76-2). It was approved for final publication by the International Committee of Legal Metrology in 1991 and will be submitted to the International Conference of Legal Metrology in 1992 for formal sanction. It supersedes the previous edition dated 1988.

TERMINOLOGY

(terms and definitions)

The terminology used in this Recommendation conforms to the "International Vocabulary of basic and general terms in metrology" (1984 edition) and the "Vocabulary of legal metrology" (1978 edition). In addition, for the purposes of this Recommendation, the following definitions apply. An index of all the terms defined below is published as a separate sheet to help in finding the corresponding definitions.

T.1 General definitions

T.1.1 Weighing instrument

Measuring instrument that serves to determine the mass of a body by using the action of gravity on this body.

The instrument may also be used to determine other quantities, magnitudes, parameters or characteristics related to mass.

According to its method of operation, a weighing instrument is classified as an automatic or nonautomatic instrument.

T.1.2 Nonautomatic weighing instrument

Instrument that requires the intervention of an operator during the weighing process, for example to deposit on or remove from the receptor the load to be measured and also to obtain the result.

The instrument permits direct observation of the weighing results, either displayed or printed; both possibilities are covered by the word "indication".

Note: Terms such as "indicate", "indicating component" and their derivatives do not include printing.

A nonautomatic weighing instrument may be:

- graduated or non-graduated,
- self-indicating, semi-self-indicating or non-self-indicating.

Note: In this Recommendation a nonautomatic weighing instrument is called an "instrument".

T.1.2.1 Graduated instrument

Instrument allowing the direct reading of the complete or partial weighing result.

T.1.2.2 Non-graduated instrument

Instrument not fitted with a scale numbered in units of mass.

- T.1.2.3 Self-indicating instrument
Instrument in which the position of equilibrium is obtained without the intervention of an operator.
- T.1.2.4 Semi-self-indicating instrument
Instrument with a self-indication weighing range, in which the operator intervenes to alter the limits of this range.
- T.1.2.5 Non-self-indicating instrument
Instrument in which the position of equilibrium is obtained entirely by the operator.
- T.1.2.6 Electronic instrument
Instrument equipped with electronic devices.
- T.1.2.7 Instrument with price scales
Instrument that indicates the price to pay by means of price charts or scales related to a range of unit prices.
- T.1.2.8 Price-computing instrument
Instrument that calculates the price to pay on the basis of the indicated mass and the unit price.
- T.1.2.9 Price-labelling instrument
Price-computing instrument that prints the weight value, unit price and price to pay for prepackages.
- T.1.2.10 Self-service instrument
Instrument that is intended to be operated by the customer.
- T.1.3 Indications provided by an instrument
- T.1.3.1 Primary indications
Indications, signals and symbols that are subject to requirements of this Recommendation.
- T.1.3.2 Secondary indications
Indications, signals and symbols that are not primary indications.
- T.2 Construction of an instrument
In this Recommendation the term "device" is used for any means by which a specific function is performed, irrespective of the physical realization, e.g. by a mechanism or a key initiating an operation; the device may be a small part or a major portion of an instrument.

- T.2.1 Main devices
- T.2.1.1 Load receptor
Part of the instrument intended to receive the load.
- T.2.1.2 Load-transmitting device
Part of the instrument for transmitting the force produced by the load acting on the load receptor, to the load-measuring device.
- T.2.1.3 Load-measuring device
Part of the instrument for measuring the mass of the load by means of an equilibrium device for balancing the force coming from the load transmitting device, and an indicating or printing device.
- T.2.2 Module
Part of an instrument which performs a specific function, can be examined separately and is subject to specified partial error limits.
- T.2.3 Electronic parts
- T.2.3.1 Electronic device
A device employing electronic sub-assemblies and performing a specific function. An electronic device is usually manufactured as a separate unit and can be independently tested.

Note: An electronic device, as defined above, may be a complete instrument (e.g. instrument for direct sales to the public) or part of an instrument (e.g. printer, indicator).
- T.2.3.2 Electronic sub-assembly
A part of an electronic device, employing electronic components and having a recognizable function of its own.

Examples: A/D converter, display matrix,...
- T.2.3.3 Electronic component
The smallest physical entity that uses electron or hole conduction in semi-conductors, gases or in a vacuum.
- T.2.4 Indicating device (of a weighing instrument)
Part of the load measuring device on which the direct reading of the result is obtained.

- T.2.4.1 Indicating component
Component indicating the equilibrium and/or the result.
On an instrument with one position of equilibrium it indicates only the equilibrium (so-called zero).
On an instrument with several positions of equilibrium it indicates both the equilibrium and the result. On an electronic instrument, this is the display.
- T.2.4.2 Scale mark
A line or other mark on an indicating component corresponding to a specified value of mass.
- T.2.4.3 Scale base
An imaginary line through the centres of all the shortest scale marks.
- T.2.5 Auxiliary indicating devices
- T.2.5.1 Rider
Detachable poise of small mass that may be placed and moved either on a graduated bar integral with the beam or on the beam itself.
- T.2.5.2 Device for interpolation of reading (vernier or nonius)
Device connected to the indicating element and sub-dividing the scale of an instrument, without special adjustment.
- T.2.5.3 Complementary indicating device
Adjustable device by means of which it is possible to estimate, in units of mass, the value corresponding to the distance between a scale mark and the indicating component.
- T.2.5.4 Indicating device with a differentiated scale division
Digital indicating device of which the last figure after the decimal sign is clearly differentiated from other figures.
- T.2.6 Extended indicating device
A device temporarily changing the actual scale interval (d) to a value less than the verification scale interval (e) following a manual command.
- T.2.7 Supplementary devices
- T.2.7.1 Levelling device
Device for setting an instrument to its reference position.

T.2.7.2 Zero-setting device

Device for setting the indication to zero when there is no load on the load receptor.

T.2.7.2.1 Nonautomatic zero-setting device

Device for setting the indication to zero by an operator.

T.2.7.2.2 Semi-automatic zero-setting device

Device for setting the indication to zero automatically following a manual command.

T.2.7.2.3 Automatic zero-setting device

Device for setting the indication to zero automatically without the intervention of an operator.

T.2.7.2.4 Initial zero-setting device

Device for setting the indication to zero automatically at the time the instrument is switched on and before it is ready for use.

T.2.7.3 Zero-tracking device

Device for maintaining the zero indication within certain limits automatically.

T.2.7.4 Tare device

Device for setting the indication to zero when a load is on the load receptor:

- without altering the weighing range for net loads (additive tare device), or
- reducing the weighing range for net loads (subtractive tare device).

It may function as:

- a nonautomatic device (load balanced by an operator),
- a semi-automatic device (load balanced automatically following a single manual command),
- an automatic device (load balanced automatically without the intervention of an operator).

T.2.7.4.1 Tare-balancing device

Tare device without indication of the tare value when the instrument is loaded.

T.2.7.4.2 Tare-weighing device

Tare device that stores the tare value and is capable of indicating or printing it whether or not the instrument is loaded.

- T.2.7.5 Preset tare device
- Device for subtracting a preset tare value from a gross or net weight value and indicating the result of the calculation. The weighing range for net loads is reduced accordingly.
- T.2.7.6 Locking device
- Device for immobilizing all or part of the mechanism of an instrument.
- T.2.7.7 Auxiliary verification device
- Device permitting separate verification of one or more main devices of an instrument.
- T.2.7.8 Selection device for load receptors and load-measuring devices
- Device for attaching one or more load receptors to one or more load measuring devices, whatever intermediate load-transmitting devices are used.
- T.2.7.9 Indication stabilizing device
- Device for maintaining a stable indication under given conditions.
- T.3 Metrological characteristics of an instrument
- T.3.1 Weighing capacity
- T.3.1.1 Maximum capacity (Max)
- Maximum weighing capacity, not taking into account the additive tare capacity.
- T.3.1.2 Minimum capacity (Min)
- Value of the load below which the weighing results may be subject to an excessive relative error.
- T.3.1.3 Self-indication capacity
- Weighing capacity within which equilibrium is obtained without the intervention of an operator.
- T.3.1.4 Weighing range
- Range between the minimum and maximum capacities.

- T.3.1.5 Extension interval of self-indication
Value by which it is possible to extend the range of self-indication within the weighing range.
- T.3.1.6 Maximum tare effect (T = + ..., T = - ...)
Maximum capacity of the additive tare device or the subtractive tare device.
- T.3.1.7 Maximum safe load (Lim)
Maximum static load that can be carried by the instrument without permanently altering its metrological qualities.
- T.3.2 Scale divisions
- T.3.2.1 Scale spacing (instrument with analogue indication)
Distance between any two consecutive scale marks, measured along the scale base.
- T.3.2.2 Actual scale interval (d)
Value expressed in units of mass of:
– the difference between the values corresponding to two consecutive scale marks, for analogue indication, or
– the difference between two consecutive indicated values, for digital indication.
- T.3.2.3 Verification scale interval (e)
Value, expressed in units of mass, used for the classification and verification of an instrument.
- T.3.2.4 Scale interval of numbering
Value of the difference between two consecutive numbered scale marks.
- T.3.2.5 Number of verification scale intervals (single-interval instrument)
Quotient of the maximum capacity and the verification scale interval:
$$n = \text{Max}/e$$
- T.3.2.6 Multi-interval instrument
Instrument having one weighing range which is divided into partial weighing ranges each with different scale intervals, with the weighing range determined automatically according to the load applied, both on increasing and decreasing loads.

- T.3.2.7 Multiple range instrument
Instrument having two or more weighing ranges with different maximum capacities and different scale intervals for the same load receptor, each range extending from zero to its maximum capacity.
- T.3.3 Reduction ratio R
The reduction ratio of a load transmitting device is:
$$R = FM/FL$$
where:
FM: force acting on the load measuring device,
FL: force acting on the load receptor.
- T.4 Metrological properties of an instrument
- T.4.1 Sensitivity
For a given value of the measured mass, the quotient of the change of the observed variable ℓ and the corresponding change of the measured mass M:
$$k = \Delta\ell/\Delta M$$
- T.4.2 Discrimination
Ability of an instrument to react to small variations of load.
The discrimination threshold, for a given load, is the value of the smallest additional load that, when gently deposited on or removed from the load receptor, causes a perceptible change in the indication.
- T.4.3 Repeatability
Ability of an instrument to provide results that agree one with the other when the same load is deposited several times and in a practically identical way on the load receptor under reasonably constant test conditions.
- T.4.4 Durability
Ability of an instrument to maintain its performance characteristics over a period of use.
- T.4.5 Warm-up time
The time between the moment power is applied to an instrument and the moment at which the instrument is capable of complying with the requirements.
- T.5 Indications and errors
- T.5.1 Methods of indication
- T.5.1.1 Balancing by weights
Value of metrologically controlled weights that balance the load (taking into account the reduction ratio of the load).

- T.5.1.2 Analogue indication
Indication enabling the evaluation of the equilibrium position to a fraction of the scale interval.
- T.5.1.3 Digital indication
Indication in which the scale marks are composed of a sequence of aligned figures that do not permit interpolation to fractions of the scale interval.
- T.5.2 Weighing results
Note: The following definitions apply only when the indication has been zero before the load has been applied to the instrument.
- T.5.2.1 Gross value (G or B)
Indication of the weight of a load on an instrument, with no tare or preset tare device in operation.
- T.5.2.2 Net value (N)
Indication of the weight of a load placed on an instrument after operation of a tare device.
- T.5.2.3 Tare value (T)
The weight value of a load, determined by a tare weighing device.
- T.5.3 Other weight values
- T.5.3.1 Preset tare value (PT)
Numerical value, representing a weight, that is introduced into the instrument.
"Introduced" includes procedures such as: keying in, recalling from a data storage, or inserting via an interface.
- T.5.3.2 Calculated net value
Value of the difference between a gross or net weight value and a preset tare value.
- T.5.3.3 Calculated total weight value
Calculated sum of more than one weight value and/or calculated net value.
- T.5.4 Reading
- T.5.4.1 Reading by simple juxtaposition
Reading of the weighing result by simple juxtaposition of consecutive figures giving the weighing result, without the need of calculation.

T.5.4.2 Overall inaccuracy of reading

The overall inaccuracy of reading of an instrument with analogue indication is equal to the standard deviation of the same indication, the reading of which is carried out under normal conditions of use by several observers.

It is customary to make at least ten readings of the result.

T.5.4.3 Rounding error of digital indication


Difference between the indication and the result the instrument would give with analogue indication.

T.5.4.4 Minimum reading distance

The shortest distance that an observer is able freely to approach the indicating device to take a reading under normal conditions of use.

This approach is considered to be free for the observer if there is a clear space of at least 0.8 m in front of the indicating device (see Figure 1).

Figure 1



- T.5.5 Errors
(See Figure 2 for illustration of certain terms used)
- T.5.5.1 Error (of indication)
The indication of an instrument minus the (conventional) true value of the mass.
- T.5.5.2 Intrinsic error
The error of an instrument under reference conditions.
- T.5.5.3 Initial intrinsic error
The intrinsic error of an instrument as determined prior to the performance and span stability tests.
- T.5.5.4 Maximum permissible error
Maximum difference, positive or negative, allowed by regulation between the indication of an instrument and the corresponding true value, as determined by reference standard masses, with the instrument being at zero at no-load, in the reference position.
- T.5.5.5 Fault
The difference between the error of indication and the intrinsic error of an instrument.
Note: Principally, a fault is the result of an undesired change of data contained in or flowing through an electronic instrument.
- T.5.5.6 Significant fault
A fault greater than e .
Note: For a multi-interval instrument, the value of e is that appropriate to the partial weighing range.
The following are not considered to be significant faults, even when they exceed e :
 - faults arising from simultaneous and mutually independent causes in the instrument,
 - faults implying the impossibility to perform any measurement,
 - faults being so serious that they are bound to be noticed by all those interested in the result of measurement,
 - transitory faults being momentary variations in the indication which cannot be interpreted, memorized or transmitted as a measuring result.
- T.5.5.7 Durability error
The difference between the intrinsic error over a period of use and the initial intrinsic error of an instrument.
- T.5.5.8 Significant durability error
A durability error greater than e .
Note 1: A durability error can be due to mechanical wear and tear or due to drift and ageing of electronic parts. The concept of significant durability error applies only to electronic parts.

Note 2: For a multi-interval instrument, the value of e is that appropriate to the partial weighing range.

The following are not considered to be significant durability errors, even when they exceed e :

errors occurring after a period of instrument use that are clearly the result of a failure of a device/component, or of a disturbance and for which the indication:

- cannot be interpreted, memorized, or transmitted as a measurement result, or
- implies the impossibility to perform any measurement, or
- is so obviously wrong that it is bound to be noticed by all those interested in the result of measurement.

T.5.5.9 Span stability

The capability of an instrument to maintain the difference between the indication of weight at maximum capacity and the indication at zero over a period of use within specified limits.

T.6 Influences and reference conditions

T.6.1 Influence quantity

A quantity that is not the subject of the measurement but which influences the values of the measurand or the indication of the instrument.

T.6.1.1 Influence factor

An influence quantity having a value within the specified rated operating conditions of the instrument.

T.6.1.2 Disturbance

An influence quantity having a value within the limits specified in this Recommendation, but outside the specified rated operating conditions of the instrument.

T.6.2 Rated operating conditions

Conditions of use, giving the range of values of influence quantities for which the metrological characteristics are intended to lie within the specified maximum permissible errors.

T.6.3 Reference conditions

A set of specified values of influence factors fixed to ensure valid inter-comparison of the results of measurements.


T.6.4 Reference position

Position of the instrument at which its operation is adjusted.

T.7 Performance test

A test to verify whether the equipment under test (EUT) is capable of performing its intended functions.

Figure 2
Illustration of certain terms used



- M = mass to be measured
- E = error of indication (T.5.5.1)
- MPE₁ = maximum permissible error on initial verification
- MPE₂ = maximum permissible error in service
- C = characteristic under reference conditions
- C₁ = characteristic due to influence factor or disturbance (*)
- E_{SP} = error of indication evaluated during span stability test
- I = intrinsic error (T.5.5.2)
- V = variation in the errors of indication during span stability test

Situation 1 – shows the error E₁ of an instrument due to an influence factor or a disturbance. I₁ is the intrinsic error. The fault (T.5.5.5) due to the influence factor or disturbance applied equals E₁ - I₁.

Situation 2 – shows the average value E_{SP1av} of the errors at the first measurement of the span stability test, some other errors E_{SPi} and E_{SPk}, and the extreme values of the errors, E_{SPm} and E_{SPn}, all these errors being evaluated at different moments during the span stability test. The variation V in the errors of indication during the span stability test equals E_{SPm} - E_{SPn}.

(*) For the purposes of this illustration it is supposed that the influence factor or the disturbance has an influence on the characteristic which is not erratic.

NONAUTOMATIC WEIGHING INSTRUMENTS

1 Scope

This Recommendation specifies the metrological and technical requirements for nonautomatic weighing instruments that are subject to official metrological control.

It is intended to provide standardized requirements and testing procedures to evaluate the metrological and technical characteristics in a uniform and traceable way.

2 Principles of the Recommendation

2.1 Units of measurement

The units of mass to be used on an instrument are the kilogram (kg), the milligram (mg), the gram (g), and the ton (t).

For special applications, e.g. trade with precious stones, the metric carat (1 carat = 0.2 g) may be used as unit of measurement. A symbol for the carat shall be ct.

2.2 Principles of the metrological requirements

The requirements apply to all instruments irrespective of their principles of measurement.

Instruments are classified according to:

- the verification scale interval, representing absolute accuracy, and
- the number of verification scale intervals, representing relative accuracy.

The maximum permissible errors are in the order of magnitude of the verification scale interval.

A minimum capacity (Min) is specified to indicate that use of the instrument with light loads is likely to give rise to excessive relative errors.

2.3 Principles of the technical requirements

General technical requirements apply to all types of instruments, whether mechanical or electronic, and are completed or modified with additional requirements for instruments used for specific applications or designed for a special technology. They are intended to specify the performance, not the design of an instrument, so that technical progress is not impeded.

In particular, functions of electronic instruments not covered by this Recommendation should be allowed provided that they do not interfere with the metrological requirements.

Testing procedures are provided to establish conformity of instruments with the requirements of this Recommendation. They should be applied, and the Pattern Evaluation Report (R 76-2) should be used, to facilitate exchange and acceptance of test results by metrological authorities.

2.4 Application of requirements

The requirements of this Recommendation apply to all devices performing the relevant functions, whether they are incorporated in an instrument or manufactured as separate units.

Examples are: load-measuring device,
indicating device,
printing device,
preset tare device,
price-calculating device.

However, devices that are not incorporated in the instrument may, by national legislation, be exempted from the requirements for special applications.

2.5 Terminology

The terminology given in pages 5 to 17 shall be considered as a part of this Recommendation.

3 Metrological requirements

3.1 Principles of classification

3.1.1 Accuracy classes

The accuracy classes for instruments and their symbols (*) are given in Table 1.

Table 1

special accuracy	Ⓘ
high accuracy	Ⓜ
medium accuracy	Ⓜ
ordinary accuracy	Ⓜ

(*) Ovals of any shape, or two horizontal lines joined by two half-circles are permitted. A circle shall not be used because, in conformity with the International Recommendation OIML R 34 "Accuracy classes of measuring instruments", it is used for the designation of accuracy classes of measuring instruments of which the maximum permissible errors are expressed by a constant relative error in %.

3.1.2 Verification scale interval

The verification scale interval for different types of instruments is given in Table 2.

Table 2

Type of instrument	Verification scale interval
Graduated, without auxiliary indicating device	$e = d$
Graduated, with auxiliary indicating device	e is chosen by the manufacturer according to requirements in 3.2 and 3.4.2
Non-graduated	e is chosen by the manufacturer according to requirements in 3.2

3.2 Classification of instruments

The verification scale interval, number of verification scale intervals and the minimum capacity, in relation to the accuracy class of an instrument, are given in Table 3.

Table 3

Accuracy class	Verification scale interval e	Number of verification scale intervals $n = \text{Max}/e$		Minimum capacity Min (Lower limit)
		minimum	maximum	
Special Ⓘ	$0.001 \text{ g} \leq e$ (*)	50 000 (**)	–	100 e
High Ⓙ	$0.001 \text{ g} \leq e \leq 0.05 \text{ g}$ $0.1 \text{ g} \leq e$	100 5 000	100 000 100 000	20 e 50 e
Medium Ⓚ	$0.1 \text{ g} \leq e \leq 2 \text{ g}$ $5 \text{ g} \leq e$	100 500	10 000 10 000	20 e 20 e
Ordinary Ⓛ	$5 \text{ g} \leq e$	100	1 000	10 e

On multiple range instruments the verification scale intervals are e_1, e_2, \dots, e_r with $e_1 < e_2 < \dots < e_r$. Min, n and Max are indexed accordingly.

(*) It is not normally feasible to test and verify an instrument to $e < 1 \text{ mg}$, due to the uncertainty of the test loads.

(**) See exception in 3.4.4.

On multiple range instruments, each range is treated basically as an instrument with one range.

For special application that are clearly marked on the instrument, an instrument may have weighing ranges in classes **I** and **II** or in classes **II** and **III**. The instrument as a whole shall then comply with the more severe requirements of 3.9 applicable to either of the two classes.

3.3 Additional requirements for a multi-interval instrument (*)

3.3.1 Partial weighing range

Each partial range (index $i = 1, 2 \dots$) is defined by:

- its verification scale interval e_i , $e_{i+1} > e_i$,
- its maximum capacity Max_i ,
- its minimum capacity $Min_i = Max_{i-1}$ (for $i = 1$ the minimum capacity is $Min_1 = Min$).

The number of verification scale intervals n_i for each partial range is equal to:

$$n_i = Max_i/e_i$$

3.3.2 Accuracy class

e_i and n_i in each partial weighing range, and Min_1 shall comply with the requirements given in Table 3 according to the accuracy class of the instrument.

3.3.3 Maximum capacity of partial weighing ranges

With the exception of the last partial weighing range, the requirements in Table 4 shall be complied with, according to the accuracy class of the instrument.

Table 4

Class	I	II	III	III
Max_i/e_{i+1}	$\geq 50\ 000$	$\geq 5\ 000$	≥ 500	≥ 50

(*) Example for a multi-interval instrument:

Maximum capacity $Max = 15\text{ kg}$ class **III**
 Verification scale intervals $e_1 = 1\text{ g}$ from 0 to 2 kg
 $e_2 = 2\text{ g}$ from 2 kg to 5 kg
 $e_3 = 10\text{ g}$ from 5 kg to 15 kg

This instrument has one Max and one weighing range from $Min = 20\text{ g}$ to $Max = 15\text{ kg}$. The partial weighing ranges are:

$Min = 20\text{ g}$, $Max_1 = 2\text{ kg}$, $e_1 = 1\text{ g}$, $n_1 = 2\ 000$
 $Min_2 = 2\text{ kg}$, $Max_2 = 5\text{ kg}$, $e_2 = 2\text{ g}$, $n_2 = 2\ 500$
 $Min_3 = 5\text{ kg}$, $Max_3 = Max = 15\text{ kg}$, $e_3 = 10\text{ g}$, $n_3 = 1\ 500$

The maximum permissible errors on initial verification (mpe) (see 3.5.1) are:

for $m = 400\text{ g} = 400e_1$ mpe = 0.5 g
 for $m = 1\ 600\text{ g} = 1\ 600e_1$ mpe = 1.0 g
 for $m = 2\ 100\text{ g} = 1\ 050e_2$ mpe = 2.0 g
 for $m = 4\ 250\text{ g} = 2\ 125e_2$ mpe = 3.0 g
 for $m = 5\ 100\text{ g} = 510e_3$ mpe = 10.0 g
 for $m = 15\ 000\text{ g} = 1\ 500e_3$ mpe = 10.0 g

Whenever the variation of the indication due to certain influence factors is limited to a fraction or multiple of e , this means, in a multi-interval instrument, that e is to be taken according to the load applied; in particular, at or near zero load $e = e_1$.

3.3.4 Instrument with a tare device

Requirements concerning the ranges of a multi-interval instrument apply to the net load, for every possible value of the tare.

3.4 Auxiliary indicating devices

3.4.1 Type and application

Only instruments of classes **I** and **II** may be fitted with an auxiliary indicating device, which shall be:

- a device with a rider, or
- a device for interpolation of reading, or
- a complementary indicating device (*), or
- an indicating device with a differentiated scale division (**).

These devices are permitted only to the right of the decimal sign.

A multi-interval instrument shall not be fitted with an auxiliary indicating device.

3.4.2 Verification scale interval

The verification scale interval e is determined by the expression:


$$d < e \leq 10 d \text{ (***)}$$

$$e = 10^k \text{ kg}$$

k being a positive or negative whole number, or zero.


This requirement does not apply to an instrument of class **I** with $d < 1 \text{ mg}$, where $e = 1 \text{ mg}$.

(*) Figure 3: example of a complementary indicating device



indication: 174.273 g
last figure: 3
 $d = 1 \text{ mg}$ $e = 10 \text{ mg}$

(**) Figure 4: examples of indicating devices each with a differentiated scale division



last differentiated figure: 5
 $d = 0.01 \text{ g}$ or 0.05 g
 $e = 0.1 \text{ g}$



last differentiated figure: 8
 $d = 0.01 \text{ g}$ or 0.02 g
 $e = 0.1 \text{ g}$

(***) The values of e , calculated following this rule, are, for example:

Table 5

$d =$	0.1 g	0.2 g	0.5 g
$e =$	1 g	1 g	1 g

3.4.3 Minimum capacity

The minimum capacity of the instrument is determined in conformity with the requirements in Table 3. However, in the last column of this Table, the verification scale interval e is replaced by the actual scale interval d .

3.4.4 Minimum number of verification scale intervals

For an instrument of class **I** with $d < 0.1$ mg, n may be less than 50 000.

3.5 Maximum permissible errors (*)

3.5.1 Values of maximum permissible errors on initial verification

The maximum permissible errors for increasing or decreasing loads are given in Table 6.

Table 6

Maximum permissible errors on initial verification	for loads m expressed in verification scale intervals e			
	class I	class II	class III	class III
$\pm 0.5 e$	$0 \leq m \leq 50\ 000$	$0 \leq m \leq 5\ 000$	$0 \leq m \leq 500$	$0 \leq m \leq 50$
$\pm 1 e$	$50\ 000 < m \leq 200\ 000$	$5\ 000 < m \leq 20\ 000$	$500 < m \leq 2\ 000$	$50 < m \leq 200$
$\pm 1.5 e$	$200\ 000 < m$	$20\ 000 < m \leq 100\ 000$	$2\ 000 < m \leq 10\ 000$	$200 < m \leq 1\ 000$

3.5.2 Values of maximum permissible errors in service

The maximum permissible errors in service shall be twice the maximum permissible errors on initial verification.

3.5.3 Basic rules concerning the determination of errors

3.5.3.1 Influence factors

Errors shall be determined under normal test conditions. When the effect of one factor is being evaluated, all other factors are to be kept relatively constant, at a value close to normal.

3.5.3.2 Elimination of rounding error

The rounding error included in any digital indication shall be eliminated if the actual scale interval is greater than $0.2 e$.

3.5.3.3 Maximum permissible errors for net values

The maximum permissible errors apply to the net value for every possible tare load, except preset tare values.

(*) An example of application to multi-interval instruments is given in the footnote to 3.3.

3.5.3.4 Tare weighing device

The maximum permissible errors for a tare weighing device are the same, for any tare value, as those of the instrument, for the same value of load.

3.5.4 Apportioning of errors

Where modules are examined separately in the process of pattern approval, the following requirements apply.

3.5.4.1 The error limits applicable to a module M_i which is examined separately are equal to a fraction p_i of the maximum permissible errors or the allowed variations of the indication of the complete instrument. The fractions for any module have to be taken for the same accuracy class and the same number of verification scale intervals, as for the complete instrument incorporating the module.

The fractions p_i shall satisfy the following equation:

$$p_1^2 + p_2^2 + p_3^2 + \dots \leq 1$$

3.5.4.2 The fraction p_i shall be chosen by the manufacturer of the module and shall be verified by an appropriate test. However, the fraction shall not exceed 0.8 and shall not be less than 0.3, when more than one module contribute to the effect in question.

Acceptable solution (see 2nd paragraph of clause 4, page 40)

For mechanical structures such as weighbridges, load transmitting devices and mechanical or electrical connecting elements evidently designed and manufactured according to sound engineering practice, an overall fraction $p_i = 0.5$ may be applied without any test, e.g. when levers are made of the same material and when the chain of levers has two planes of symmetry (longitudinal and transversal), or when the stability characteristics of electrical connecting elements are appropriate for the signals transmitted, such as load cell output, impedance,...

For instruments incorporating the typical modules (see acceptable solution in 8.2.1) the fractions p_i may have the values given in Table 7.

Table 7

Performance criteria	Load cell	Electronic indicator	Connecting elements, etc.
Combined effect (*)	0.7	0.5	0.5
Temperature effect on no load indication	0.7	0.5	0.5
Power supply variation	–	1	–
Effect of creep	1	–	–
Damp heat	0.7	0.5	0.5
(*) Combined effects: non-linearity, hysteresis, temperature effect on span. After the warm-up time specified by the manufacturer, the combined effect error fractions apply to modules. The sign "–" means "not applicable".			

3.5.5 Testing for verification

Where a load measuring device is tested separately for verification, the maximum permissible error is equal to 0.7 times the maximum permissible error for the complete instrument (this fraction includes the errors attributable to the verification devices used).

In any case the instrument submitted for verification shall be tested as a whole.

3.6 Permissible differences between results

Regardless of what variation of results is permitted, the error of any single weighing result shall by itself not exceed the maximum permissible error for the given load.

3.6.1 Repeatability

The difference between the results of several weighings of the same load shall not be greater than the absolute value of the maximum permissible error of the instrument for that load.

3.6.2 Eccentric loading

The indications for different positions of a load shall meet the maximum permissible errors, when the instrument is tested according to 3.6.2.1 through 3.6.2.4.

Note: If an instrument is designed in such a way that loads may be applied in different manners, it may be appropriate to apply more than one of the following tests.

3.6.2.1 Unless otherwise specified hereafter, a load corresponding to 1/3 of the sum of the maximum capacity and the corresponding maximum additive tare effect shall be applied.

3.6.2.2 On an instrument with a load receptor having n points of support, with $n > 4$, the fraction $1/(n-1)$ of the sum of the maximum capacity and the maximum additive tare effect shall be applied to each point of support.

3.6.2.3 On an instrument with a load receptor subject to minimal off-centre loading (e.g. tank, hopper,...) a test load corresponding to one-tenth of the sum of the maximum capacity and the maximum additive tare effect shall be applied to each point of support.

3.6.2.4 On an instrument used for weighing rolling loads (e.g. vehicle scale, rail suspension instrument) a rolling test load corresponding to the usual rolling load, the heaviest and the most concentrated one which may be weighed, but not exceeding 0.8 times the sum of the maximum capacity and the maximum additive tare effect, shall be applied at different points on the load receptor.

3.6.3 Multiple indicating devices

For a given load the difference between the indications of multiple indicating devices including tare weighing devices, shall be not greater than the absolute value of the maximum permissible error, but shall be zero between digital indicating or printing devices.

3.6.4 Different positions of equilibrium

The difference between two results obtained for the same load when the method of balancing the load is changed (in the case of an instrument fitted with a device for extending the self-indication capacity) in two consecutive tests, shall not exceed the absolute value of the maximum permissible error for the applied load.

3.7 Verification standards

3.7.1 Weights

The standard weights or standard masses used for the verification of an instrument shall not have an error greater than 1/3 of the maximum permissible error of the instrument for the applied load.

3.7.2 Auxiliary verification device

When an instrument is fitted with an auxiliary verification device, or when it is verified with a separate auxiliary device, the maximum permissible errors of this device shall be 1/3 of the maximum permissible errors for the applied load. If weights are used, the effect of their errors shall not exceed 1/5 of the maximum permissible errors of the instrument to be verified for the same load.

3.7.3 Substitution of standard weights

When testing instruments with $Max > 1$ t, instead of standard weights any other constant load may be used, provided that standard weights of at least 1 t or 50 % of Max, whichever is greater, are used. Instead of 50 % of Max, the portion of standard weights may be reduced to:

35 % of Max if the repeatability error is not greater than 0.3 e,
20 % of Max if the repeatability error is not greater than 0.2 e.

The repeatability error has to be determined with a load of about 50 % of Max which is placed 3 times on the load receptor.

3.8 Discrimination

3.8.1 Non-self-indicating instrument

An extra load equivalent to 0.4 times the absolute value of the maximum permissible error for the applied load when gently placed on or withdrawn from the instrument at equilibrium shall produce a visible movement of the indicating element.

3.8.2 Self- or semi-self-indicating instrument

3.8.2.1 Analogue indication

An extra load equivalent to the absolute value of the maximum permissible error for

the applied load when gently placed on or withdrawn from the instrument at equilibrium shall cause a permanent displacement of the indicating element corresponding to not less than 0.7 times the extra load.

3.8.2.2 Digital indication

An additional load equal to 1.4 times the actual scale interval, when gently placed on or withdrawn from the instrument at equilibrium shall change the initial indication.

3.9 Variations due to influence quantities and time

An instrument shall comply, unless otherwise specified, with 3.5, 3.6 and 3.8 under the conditions of 3.9.2 and 3.9.3 and additionally it shall comply with 3.9.1 and 3.9.4.

3.9.1 Tilting

3.9.1.1 For an instrument of class ①, ② or ③ liable to be tilted, the influence of tilting shall be determined under the effect of a lengthwise or transverse tilting equal to 2/1000 or corresponding to the limiting value of tilting marked on, or indicated by a level indicator, whichever is the greater tilt.

The absolute value of the difference between the indication of the instrument in its reference position (not tilted) and the indication in the tilted position shall not exceed:

- at no load, two verification scale intervals (the instrument having first been adjusted to zero at no load in its reference position) except instruments of class ① (however see 4.14.8),
- at self indication capacity and at maximum capacity, the maximum permissible error (the instrument having been adjusted to zero at no load both in the reference and in the tilted position).

An instrument shall be fitted with a levelling device and a level indicator fixed firmly on the instrument in a place clearly visible to the user, unless the instrument is:

- freely suspended, or
- installed in a fixed position, or
- complying with the requirements on tilting when tilted to 5 % in any direction.

The limiting value of the level indicator shall be obvious, so that tilting is easily noticed.

Note: “Limiting value of tilting”: displacement of 2 mm from a central position (regardless of the diameter of any ring used to indicate the center), lamp, or any other indication of the level indicator which shows that the maximum permissible tilt is being exceeded.

3.9.1.2 On a class ① instrument, the limiting value of tilting shall correspond to a tilt of no more than 2/1000 otherwise the instrument shall meet the requirements for instruments of class ②.

3.9.2 Temperature (*)

3.9.2.1 Prescribed temperature limits

If no particular working temperature is stated in the descriptive markings of an instrument, this instrument shall maintain its metrological properties within the following temperature limits:

$$- 10 \text{ }^{\circ}\text{C}, + 40 \text{ }^{\circ}\text{C}$$

3.9.2.2 Special temperature limits

An instrument for which particular limits of working temperature are stated in the descriptive markings shall comply with the metrological requirements within those limits.

The limits may be chosen according to the application of the instrument.

The ranges within those limits shall be at least equal to:

- 5 °C for instruments of class **I**,
- 15 °C for instruments of class **II**,
- 30 °C for instruments of classes **III** and **III**.

3.9.2.3 Temperature effect on no-load indication

The indication at zero or near zero shall not vary by more than one verification scale interval for a difference in ambient temperature of 1 °C for instruments of class **I** and 5 °C for other classes.

For multi-interval instruments and for multiple range instruments this applies to the smallest verification scale interval of the instrument.

3.9.3 Mains power supply

An instrument operated from a mains power supply shall comply with the metrological requirements if the power supply varies:

- in voltage from – 15 % to + 10 % of the value marked on the instrument,
- in frequency from – 2 % to + 2 % of the value marked on the instrument, if AC is used.

3.9.4 Time

Under reasonably constant environmental conditions, an instrument of class **II**, **III** or **III** shall meet the following requirements.

3.9.4.1 When any load is kept on an instrument, the difference between the indication obtained immediately after placing a load and the indication observed during the following 30 minutes, shall not exceed 0.5 e. However, the difference between the indication obtained at 15 minutes and that at 30 minutes shall not exceed 0.2 e.

(*) Tolerances for the temperature values are given in the testing procedures, Annexes A and B.

If these conditions are not met, the difference between the indication obtained immediately after placing a load on the instrument and the indication observed during the following four hours shall not exceed the absolute value of the maximum permissible error at the load applied.

3.9.4.2 The deviation on returning to zero as soon as the indication has stabilized, after the removal of any load which has remained on the instrument for one half hour, shall not exceed $0.5 e$.

For a multi-interval instrument, the deviation shall not exceed $0.5 e_1$.

On a multiple range instrument, the deviation on returning to zero from Max_i shall not exceed $0.5 e_i$. Furthermore, after returning to zero from any load greater than Max_1 and immediately after switching to the lowest weighing range, the indication near zero shall not vary by more than e_1 during the following 5 minutes.

3.9.4.3 The durability error due to wear and tear shall not be greater than the absolute value of the maximum permissible error.

Adherence to this requirement is assumed if the instrument has passed the endurance test specified in A.6, which shall be performed only for instruments with $Max \leq 100$ kg.

3.9.5 Other influence quantities and restraints

Where other influences and restraints, such as:

- vibrations,
- precipitations and draughts,
- mechanical constraints and restrictions,

are a normal feature of the intended operating environment of the instrument, the instrument shall comply with the requirements of clauses 3 and 4 under those influences and restraints, either by being designed to operate correctly in spite of these influences, or by being protected against their action.

Example: Instruments installed outdoors without suitable protection against atmospheric conditions may normally not comply with the requirements of clauses 3 and 4 if the number of verification scale intervals n is too great. (A value of $n = 3\ 000$ should not be exceeded. Furthermore for road or rail weighbridges the verification scale interval should not be less than 10 kg).

This limit should also apply to each weighing range of combinations of instruments or of multiple range instruments or to each partial weighing range of multi-interval instruments.

3.10 Pattern evaluation tests

Upon pattern evaluation, the tests given in Annexes A and B shall be performed, to verify adherence to the requirements in 3.5, 3.6, 3.8, 3.9.1, 3.9.2, 3.9.3, 3.9.4, 4.5, 4.6, 5.3 and 6.1. The endurance test (A.6) shall be performed after all other tests in Annexes A and B.

Peripheral devices that only perform digital functions, e.g. printers or additional displays, only need to be tested for correct functioning and submitted to the disturbance tests in B.3.

4 Technical requirements for a self- or semi-self-indicating instrument

The following requirements relate to the design and the construction of instruments which are suitable to give correct and unambiguous weighing results, under normal conditions of use and proper handling by unskilled users. They are not intended to prescribe solutions, but to define appropriate functioning of the instrument.

Certain solutions that have been tried over a long period have become accepted; these solutions are marked "acceptable solution"; while it is not necessary to adopt them, they are considered to comply with the requirements of the applicable provision.

4.1 General requirements of construction

4.1.1 Suitability

4.1.1.1 Suitability for application

An instrument shall be designed to suit its intended purpose of use.

4.1.1.2 Suitability for use

An instrument shall be solidly and carefully constructed in order to ensure that it maintains its metrological qualities during a period of use.

4.1.1.3 Suitability for verification

An instrument shall permit the tests set out in this Recommendation to be performed.

In particular, load receptors shall be such that the standard masses can be deposited on them easily and in total safety. If weights cannot be placed, an additional support may be required.

It must be possible to identify devices that have been subject to a separate type examination procedure (e.g. load cells, printers,...).

4.1.2 Security

4.1.2.1 Fraudulent use

An instrument shall have no characteristics likely to facilitate its fraudulent use.

4.1.2.2 Accidental breakdown and maladjustment

An instrument shall be so constructed that an accidental breakdown or a maladjustment of control elements likely to disturb its correct functioning cannot take place without its effect being evident.

4.1.2.3 Controls

Controls shall be so designed that they cannot normally come to rest in positions other than those intended by design, unless during the manoeuvre all indication is made impossible. Keys shall be marked unambiguously.

4.1.2.4 Securing (sealing) of components and pre-set controls

Means shall be provided for securing components and pre-set controls to which access or adjustment is prohibited. National legislation may specify the sealing that is required.

On a class **I** instrument, devices to adjust sensitivity may remain unsecured.

Acceptable solution

For application of the control marks, the sealing area should have a diameter of at least 5 mm.

Components and pre-set controls may be secured by passwords or similar software means provided that any access to the secured controls or functions becomes automatically evident, e.g. by automatically updating a code number the value of which at the time of the last verified set-up had been durably marked on the data plate.

4.1.2.5 Adjustment

An instrument may be fitted with an automatic or a semi-automatic span adjustment device. This device shall be incorporated inside the instrument. External influence upon this device shall be practically impossible after sealing.

4.1.2.6 Gravity compensation

A gravity sensitive instrument may be equipped with a device for compensating the effects of gravity variations. After securing, external influence on or access to this device shall be practically impossible.

4.2 Indication of weighing results

4.2.1 Quality of reading

Reading of the results shall be reliable, easy and unambiguous under conditions of normal use:

- the overall inaccuracy of reading of an analogue indicating device shall not exceed $0.2 e$,
- the figures forming the results shall be of a size, shape and clarity for reading to be easy.

The scales, numbering and printing shall permit the figures which form the results to be read by simple juxtaposition.

4.2.2 Form of the indication

4.2.2.1 Weighing results shall contain the names or symbols of the units of mass in which they are expressed.

For any one indication of weight, only one unit of mass may be used.

The scale interval shall be in the form 1×10^k , 2×10^k or 5×10^k units in which the result is expressed, the index k being a positive or negative whole number or equal to zero.

All indicating, printing and tare weighing devices of an instrument shall, within any one weighing range, have the same scale interval for any given load.

4.2.2.2 A digital indication shall display at least one figure beginning at the extreme right.

Where the scale interval is changed automatically the decimal sign shall maintain its position in the display.

A decimal fraction shall be separated from its integer by a decimal sign (comma or dot), with the indication showing at least one figure to the left of the sign and all figures to the right.

Zero may be indicated by one zero to the extreme right, without a decimal sign.

The unit of mass shall be chosen so that weight values have not more than one non-significant zero to the right. For values with decimal sign, the non-significant zero is allowed only in the third position after the decimal sign.

4.2.3 Limits of indication

There shall be no indication above $\text{Max} + 9 e$.

4.2.4 Approximate indicating device

The scale interval of an approximate indicating device shall be greater than $\text{Max}/100$ without being smaller than $20 e$. This approximate device is considered as giving secondary indications.

4.2.5 Extending the range of self-indication on a semi-self-indicating instrument

The extension interval of the range of self-indication shall not be greater than the value of the self-indication capacity.

Acceptable solutions

- a) The scale interval of extension of the range of self-indication should be equal to the capacity of self-indication (comparator instruments are excluded from this provision).
- b) An extension device with accessible sliding poises is subject to the requirements of 6.2.2.
- c) On an extension device with enclosed sliding poises or mass switching mechanisms, each extension should involve an adequate change in the numbering. It should be possible to seal the housing and the adjusting cavities of the weights or masses.

4.3 Analogue indicating device

The following requirements apply in addition to those in 4.2.1 through 4.2.4.

4.3.1 Scale marks; length and width

Scales shall be designed and numbered so that reading the weighing result is easy and unambiguous.

Acceptable solutions


a) Form of scale marks

Scale marks should consist of lines of equal thickness; this thickness should be constant and be between $1/10$ and $1/4$ of the scale spacing, without being less than 0.2 mm. The length of the shortest scale mark should be at least equal to the scale spacing.

b) Arrangement of scale marks

Scale marks should be arranged in accordance with one of the sketches in Figure 5 (the line joining the end of the scale marks is optional).

Figure 5
Examples of application to rectilinear scales



c) Numbering

On one scale, the scale interval of numbering should be:

- constant,
- in the form 1×10^k , 2×10^k , 5×10^k units (k being a positive or negative whole number or equal to zero),
- not greater than 25 times the scale interval of the instrument.

If the scale is projected on a screen, at least two numbered scale marks should appear wholly in the projected zone.

The height of the numbers (real or apparent) expressed in millimetres should be not less than 3 times the minimum reading distance expressed in metres, without being less than 2 mm.

This height should be proportional to the length of the scale mark to which it relates.

The width of a number, measured parallel to the base of the scale, should be less than the distance between two consecutive numbered scale marks.

d) Indicating component

The width of the pointer of the indicating component should be approximately equal to that of the scale marks and of a length such that the tip is at least level with the middle of the shortest mark.

The distance between the scale and the pointer should be at most equal to the scale spacing, without being greater than 2 mm.

4.3.2 Scale spacing

The minimum value i_o of the scale spacing is equal to:

- on an instrument of class **I** or **II**:
1 mm for indicating devices,
0.25 mm for complementary indicating devices; in this case i_o is the relative movement between the indicating component and the projected scale corresponding to the verification scale interval of the instrument,
- on an instrument of class **III** or **III**:
1.25 mm for dial indicating devices,
1.75 mm for optical projection indicating devices.

Acceptable solution

The scale spacing (real or apparent) i , in millimetres, should be at least equal to $(L + 0.5) i_o$, where:

i_o is the minimum scale spacing, in millimetres

L is the minimum reading distance, in metres; at least, $L = 0.5$ m.

The greatest scale spacing should not exceed 1.2 times the smallest scale spacing of the same scale.

4.3.3 Limits of indication

Stops shall limit the movement of the indicating component whilst allowing it to travel below zero and above the capacity of self-indication. This requirement does not apply to multi-revolution dial instruments.

Acceptable solution

The stops limiting the movement of the indicating component should permit it to travel across zones of at least 4 scale spacings below zero and above the capacity of self-indication (these zones are not provided with a scale on fan charts and on dials with a single revolution pointer; they are called "blank zones").

4.3.4 Damping

The damping of the oscillations of the indicating component or of the movable scale shall be adjusted to a value slightly below "critical damping", whatever the influence factors.

Acceptable solution

Damping should achieve a stable indication after three, four or five half periods of oscillation.

Hydraulic damping elements sensitive to variations in temperature should be provided with a automatic regulating device or an easily accessible manual regulating device.

It should be impossible for the fluid of hydraulic damping elements on portable instruments to spill when the instrument is inclined at 45°.

4.4 Digital indicating and printing devices

The following requirements apply in addition to those in 4.2.1 through 4.2.5.

4.4.1 Change of indication

After a change in load, the previous indication shall not persist for longer than 1 second.

4.4.2 Stable equilibrium

Equilibrium is deemed to be stable when:

- in case of printing and/or data storage, the requirements in 4.4.5, last paragraph, are met,
- in case of zero or tare operations (4.5.4, 4.5.6, 4.5.7 and 4.6.8), it is sufficiently close to the final equilibrium to allow a correct operation of the device within relevant accuracy requirements.

4.4.3 Extended indicating device

An extended indicating device shall not be used on an instrument with a differentiated scale division.

When an instrument is fitted with an extended indicating device, displaying the indication with a scale interval smaller than e shall be possible only:

- during pressing a key, or
- for a period not exceeding 5 seconds after a manual command.

In any case printing shall not be possible.

4.4.4 Multiple use of indicating devices

Indications other than primary indications may be displayed in the same indicating device, provided that:

- quantities other than weight values are identified by the appropriate unit of measurement, or symbol thereof, or a special sign,
- weight values that are not weighing results (T.5.2.1 through T.5.2.3) shall be clearly identified, or they may be displayed only temporarily on manual command and shall not be printed.

No restrictions apply if the weighing mode is made inoperative by a special command.

4.4.5 Printing device

Printing shall be clear and permanent for the intended use. Printed figures shall be at least 2 mm high.

If printing takes place, the name or the symbol of the unit of measurement shall be either to the right of the value or above a column of values.

Printing shall be inhibited when the equilibrium is not stable.

Stable equilibrium is considered to be achieved when over a period of 5 seconds following print-out no more than two adjacent values are indicated, one of which being the printed value (*).

(*) Note: In case of instruments with $d < e$ the differentiated scale divisions are not considered.

4.4.6 Memory storage device

The storage of primary indications for subsequent indication, data transfer, totalizing, etc. shall be inhibited when the equilibrium is not stable. The criterion of stable equilibrium is the same as in 4.4.5.

4.5 Zero-setting and zero-tracking devices

An instrument may have one or more zero-setting devices and shall have not more than one zero-tracking device.

4.5.1 Maximum effect

The effect of any zero-setting device shall not alter the maximum weighing capacity of the instrument.

The overall effect of zero-setting and zero-tracking devices shall be not more than 4 %, and of the initial zero-setting device not more than 20 %, of the maximum capacity (*).

A wider range is possible for the initial zero-setting device if tests show that the instrument complies with 3.5, 3.6, 3.8 and 3.9 for any load compensated by this device within the specified range.

4.5.2 Accuracy

After zero setting the effect of zero deviation on the result of the weighing shall be not more than 0.25 e; however, on an instrument with auxiliary indicating devices this effect shall be not more than 0.5 d.

4.5.3 Multiple range instrument

Zero setting in any weighing range shall be effective also in the greater weighing ranges, if switching to a greater weighing range is possible while the instrument is loaded.

4.5.4 Control of the zero-setting device

An instrument – except an instrument according to 4.14 and 4.15 – whether or not equipped with an initial zero-setting device, may have a combined semi-automatic zero-setting device, and a semi-automatic tare-balancing device operated by the same key.

If an instrument has a zero-setting device and a tare-weighing device the control of the zero-setting device shall be separate from that of the tare-weighing device.

A semi-automatic zero-setting device shall function only:

- when the instrument is in stable equilibrium,
- if it cancels any previous tare operation.

(*) This provision does not affect an instrument of class , except if it is used for commercial transactions.

4.5.5 Zero indicating device on an instrument with digital indication

An instrument with digital indication shall have a device that displays a special signal when the deviation from zero is not more than $0.25 e$. This device may also work when zero is indicated after a tare operation.

This device is not mandatory on an instrument that has an auxiliary indicating or a zero-tracking device provided that the rate of zero tracking is not less than $0.25 d/\text{second}$.

4.5.6 Automatic zero-setting device

An automatic zero-setting device shall operate only when:

- the equilibrium is stable, and
- the indication has remained stable below zero for at least 5 seconds.

4.5.7 Zero-tracking device

A zero-tracking device shall operate only when:

- the indication is at zero, or at a negative net value equivalent to gross zero, and
- the equilibrium is stable, and
- the corrections are not more than $0.5 d/\text{second}$.

When zero is indicated after a tare operation, the zero-tracking device may operate within a range of 4 % of Max around the actual zero value.

4.6 Tare device

4.6.1 General requirements

A tare device shall comply with the relevant provisions of 4.1 through 4.4.

4.6.2 Scale interval

The scale interval of a tare-weighing device shall be equal to the scale interval of the instrument for any given load.

4.6.3 Accuracy

A tare device shall permit setting the indication to zero with an accuracy better than:

- $\pm 0.25 e$ for electronic instruments and any instrument with analogue indication,
- $\pm 0.5 d$ for mechanical instruments with digital indication and instruments with auxiliary indicating devices.

On a multi-interval instrument e shall be replaced by e_1 .

4.6.4 Operating range

The tare device shall be such that it cannot be used at or below its zero effect or above its maximum indicated effect.

4.6.5 Visibility of operation

Operation of the tare device shall be visibly indicated on the instrument. In the case of instruments with digital indication this shall be done by marking the indicated net value with the sign "NET" (*).

Note: If an instrument is equipped with a device that allows the gross value to be displayed temporarily while a tare device is in operation, the "NET" symbol shall disappear while the gross value is displayed.

This is not required for an instrument with a combined semi-automatic zero-setting device and a semi-automatic tare-balancing device operated by the same key.

It is permitted to replace the symbols NET and T by complete words in an official language of the country where the instrument is used.

Acceptable solution

The use of a mechanical tare adding device should be shown by the indication of the tare value, or by the display on the instrument of a sign, e.g. letter "T".

4.6.6 Subtractive tare device

When the use of a subtractive tare device does not allow the value of the residual weighing range to be known, a device shall prevent the use of the instrument above its maximum capacity or indicate that this capacity has been reached.

4.6.7 Multiple range instrument

On a multiple range instrument the tare operation shall be effective also in the greater weighing ranges, if switching to a greater weighing range is possible while the instrument is loaded.

4.6.8 Semi-automatic or automatic tare devices

These devices shall operate only when the instrument is in stable equilibrium.

4.6.9 Combined zero-setting and tare-balancing devices

If the semi-automatic zero-setting device and the semi-automatic tare-balancing device are operated by the same key, 4.5.2, 4.5.5 and if appropriate 4.5.7 apply at any load.

4.6.10 Consecutive tare operations

Repeated operation of a tare device is permitted.

If more than one tare device is operative at the same time, tare weight values shall be clearly designated when indicated or printed.

(*) NET may be displayed as "NET", "Net" or "net".

4.6.11 Printing of weighing results

Gross weight values may be printed without any designation. For a designation by a symbol, only "G" or "B" are permitted.

If only net weight values are printed without corresponding gross or tare values, they may be printed without any designation. A symbol for designation shall be "N". This applies also where semi-automatic zero setting and semi-automatic tare balancing are initiated by the same key.

Gross, net, or tare values determined by a multiple range instrument or by a multi-interval instrument need not be marked by a special designation referring to the (partial) weighing range.

If net weight values are printed together with the corresponding gross and/or tare values, the net and tare values shall at least be identified by the corresponding symbols "N" and "T".

However, it is permitted to replace the symbols G, B, N and T by complete words in an official language of the country where the instrument is used.

If net weight values and tare values determined by different tare devices are printed separately, they shall be suitably identified.

4.7 Preset tare device

4.7.1 Scale interval

Regardless of how a preset tare value is introduced into the device, its scale interval shall be equal or automatically rounded to the scale interval of the instrument. On a multiple range instrument a preset tare value may only be transferred from one weighing range to another one with a larger verification scale interval but shall then be rounded to the latter. For a multi-interval instrument, the maximum preset tare value shall not be greater than Max_1 and the indicated or printed calculated net value shall be rounded to the scale interval of the instrument for the same net weight value.

4.7.2 Modes of operation

A preset tare device may be operated together with one or more tare devices provided that:

- 4.6.10 is respected, and
- a preset tare operation cannot be modified or cancelled as long as any tare device operated after the preset tare operation is still in use.

Preset tare devices may operate automatically only if the preset tare value is clearly identified with the load to be measured (e.g. by bar code identification on the container).

4.7.3 Indication of operation

For the indicating device 4.6.5 applies. It shall be possible to indicate the preset tare value at least temporarily.

4.6.11 applies accordingly provided that:

- if the calculated net value is printed at least the preset tare value is printed as well, with the exception of an instrument covered by 4.14, 4.15 or 4.17;
- preset tare values are designated by the symbol "PT"; however, it is permitted to replace the symbol "PT" by complete words in an official language of the country where the instrument is used.

4.8 Locking positions

4.8.1 Prevention of weighing outside the "weigh" position

If an instrument has one or more locking devices, these devices shall only have two stable positions corresponding to "locked" and "weigh" and weighing shall only be possible in the "weigh" position.

A "preweigh" position may exist on an instrument of class **Ⓘ** or **Ⓢ**, except those under 4.14, 4.15 and 4.17.

4.8.2 Indication of position

The "locked" and "weigh" positions shall be clearly shown.

4.9 Auxiliary verification devices (removable or fixed)

4.9.1 Devices with one or more platforms

The nominal value of the ratio between the weights to be placed on the platform to balance a certain load and this load shall not be less than 1/5000 (it shall be visibly indicated just above the platform).

The value of the weights needed to balance a load equal to the verification scale interval shall be an integer multiple of 0.1 gram.

4.9.2 Numbered scale devices

The scale interval of the auxiliary verification device shall be equal to or smaller than 1/5 of the verification scale interval for which it is intended.

4.10 Selection of weighing ranges on a multiple range instrument

The range which is actually in operation shall be clearly indicated.

Manual selection of the weighing range is allowed:

- from a smaller to a greater weighing range, at any load,
- from a greater to a smaller weighing range, when there is no load on the load receptor; and the indication is zero or at a negative net value; the tare operation shall be cancelled, and zero shall be set to $\pm 0.25 e_1$, both automatically.

Automatic change over is allowed:

- from a smaller to the following greater weighing range when the load exceeds the maximum gross weight of the range being operative,

- only from a greater to the smallest weighing range when there is no load on the load receptor, and the indication is zero or at a negative net value; the tare operation shall be cancelled and zero shall be set to $\pm 0.25 e_1$, both automatically.

4.11 Devices for selection (or switching) between various load receptors-load transmitting devices and various load measuring devices

4.11.1 Compensation of no-load effect

The selection device shall ensure compensation for the unequal no-load effect of the various load receptors – load transmitting devices in use.

4.11.2 Zero-setting

Zero-setting of an instrument with any multiple combination of various load measuring devices and various load receptors shall be possible without ambiguity and in accordance with the provisions of 4.5.

4.11.3 Impossibility of weighing

Weighing shall not be possible while selection devices are being used.

4.11.4 Identification of the combinations used

Combinations of load receptors and load measuring devices used shall be readily identifiable.

4.12 Requirements for load cells

These requirements replace 3.5.4 for load cell(s) of an instrument, that have been tested separately according to the International Recommendation OIML R 60 "Metrological regulations for load cells" which assigns a fraction $p_i = 0.7$ of the maximum permissible error of the complete instrument to a load cell.

3.9.2.3, 3.9.4.1 and 3.9.4.2 are deemed to be satisfied if the load cell(s) meet(s) the following requirements.

4.12.1 Maximum capacity of the load cell

The maximum capacity of the load cell shall satisfy the condition:

$$E_{\max} \geq Q \cdot \text{Max} \cdot R/N$$

where:

E_{\max} = maximum capacity of the load cell

N = number of load cells

R = reduction ratio (see T.3.3)

Q = correction factor

The correction factor $Q > 1$ considers the possible effects of eccentric loading, dead load of the load receptor, initial zero setting range and non uniform distribution of the load.

4.12.2 Maximum number of load cell intervals

For each load cell the maximum number of load cell intervals n_{LC} (see OIML R 60) shall not be less than the number of verification scale intervals n of the instrument:

$$n_{LC} \geq n$$

On a multiple range or multi-interval instrument, this applies to any individual weighing range or partial weighing range:

$$n_{LC} \geq n_i$$

On a multi-interval instrument, the minimum dead load output return DR (see OIML R 60) shall satisfy the condition:

$$DR \leq 0.5 \cdot e_1 \cdot R/N$$

Acceptable solution

Where DR is not known, the condition $n_{LC} \geq \text{Max}_r/e_1$ should be satisfied.

Furthermore on a multiple range instrument where the same load cell(s) is (are) used for more than one range, the minimum dead load output return DR of the load cell (see OIML R 60) shall satisfy the condition

$$DR \leq e_1 \cdot R/N$$

Acceptable solution

Where DR is not known, the condition $n_{LC} \geq 0.4 \cdot \text{Max}_r/e_1$ should be satisfied.

4.12.3 Minimum load cell verification interval

The minimum load verification interval v_{\min} (see OIML R 60) shall not be greater than the verification scale interval e multiplied by the reduction ratio R of the load transmitting device and divided by the square root of the number N of load cells, as applicable:

$$v_{\min} \leq e \cdot R/\sqrt{N}$$

On a multiple range instrument where the same load cell(s) is (are) used for more than one range, or a multi-interval instrument, e is to be replaced by e_1 .

4.13 "Plus" and "minus" comparator instrument

For the purposes of verification, a "plus" and "minus" comparator instrument is considered to be a semi-self-indicating instrument.

4.13.1 Distinction between "plus" and "minus" zones

On an analogue indicating device the zones situated on either side of zero shall be distinguished by "+" and "-" signs.

On a digital indicating device an inscription near the indicating device shall be given:

- range $\pm \dots u_m$, or
- range - $\dots u_m$ / + $\dots u_m$,

where u_m represents the unit of measurement as per 2.1.

4.13.2 Form of scale

The scale of a comparator instrument shall have at least one scale division $d = e$ on either side of zero. The corresponding value shall be shown at either end of the scale.

4.14 Instrument for direct sales to the public (*)

The following requirements apply to an instrument of class **II**, **III**, or **III**, with a maximum capacity not greater than 100 kg designed to be used for direct sales to the public.

4.14.1 Primary indications

On an instrument for direct sales to the public the primary indications are the weighing result and information about correct zero position, tare and preset tare operations.

4.14.2 Zero-setting device

An instrument for direct sales to the public shall not be fitted with a nonautomatic zero-setting device unless operated only with a tool.

4.14.3 Tare device

A mechanical instrument with a weight receptor shall not be fitted with a tare device.

An instrument with one platform may be fitted with tare devices if they allow the public to see:

- whether they are in use, and
- whether their setting is altered.

Only one tare device shall be in operation at any given time.

Note: The restrictions in use are included under 4.14.3.2, 2nd indent.

An instrument shall not be fitted with a device which can recall the gross value while a tare or preset tare device is in operation.

4.14.3.1 Nonautomatic tare device

A displacement of 5 mm of a point of the control shall be at most equal to one verification scale interval.

(*) Note: Interpretation of what is included in "direct sales to the public" is left up to national legislation.

4.14.3.2 Semi-automatic tare device

An instrument may be fitted with semi-automatic tare devices if:

- the action of the tare devices does not permit a reduction of the value of the tare, and
- their effect can only be cancelled when there is no load on the load receptor.

In addition, the instrument shall comply with at least one of the following requirements:

1. the tare value is indicated permanently in a separate display,
2. the tare value is indicated with a sign "-" (minus), when there is no load on the load receptor, or
3. the effect of the device is cancelled automatically and the indication returns to zero when unloading the load receptor after a stable net weighing result greater than zero has been indicated.

4.14.3.3 Automatic tare device

An instrument shall not be fitted with an automatic tare device.

4.14.4 Preset tare device

A preset tare device may be provided if the preset tare value is indicated as a primary indication on a separate display which is clearly differentiated from the weight display. 4.14.3.2, first paragraph, applies.

It shall not be possible to operate a preset tare device if a tare device is in use.

Where a preset tare is associated with a price look up (PLU) the preset tare value may be cancelled at the same time as the PLU is cancelled.

4.14.5 Impossibility of weighing

It shall be impossible to weigh or to guide the indicating element during the normal locking operation or during the normal operation of adding or subtracting weights.

4.14.6 Visibility

All primary indications shall be displayed clearly and simultaneously to both the vendor and the customer.

On digital devices that display primary indications, the numerical figures on either set shall be of the same dimension and at least 10 mm high, with a tolerance of 0.5 mm.

On an instrument to be used with weights it shall be possible to distinguish the value of the weights.

Acceptable solution

The primary indications should be grouped together in two sets of scales or displays.

4.14.7 Auxiliary and extended indicating devices

An instrument shall not be fitted with an auxiliary indicating device nor an extended indicating device.

4.14.8 Instruments of class ②

An instrument of class ② shall comply with the requirements given in 3.9 for an instrument of class ③.

4.14.9 Significant fault

When a significant fault has been detected, a visible or audible alarm shall be provided for the customer, and data transmission to any peripheral equipment shall be prevented. This alarm shall continue until such time as the user takes action or the cause disappears.

4.14.10 Counting ratio

The counting ratio on a mechanical counting instrument shall be 1/10 or 1/100.

4.15 Additional requirements for an instrument for direct sales to the public with price indication

The following requirements are to be applied in addition to 4.14.

4.15.1 Primary indications

On a price-indicating instrument the supplementary primary indications are unit price and price to pay and, if applicable, number, unit price and price to pay for non-weighed articles, prices for non-weighed articles and price totals. Price charts, such as fan charts, are not subject to the requirements of this Recommendation.

4.15.2 Instrument with price scales

For unit price and price-to-pay scales, 4.2 and 4.3.1 through 4.3.3 apply accordingly; however, decimal fractions shall be indicated according to national regulations.

Reading from price scales shall be so possible that the absolute value of the difference between the product of the indicated weight W and unit price U and the indicated price to pay P is not greater than the product of e and the unit price for that scale:

$$|W \cdot U - P| \leq e \cdot U$$

4.15.3 Price computing instrument

The price to pay shall be calculated and rounded to the nearest interval of price to pay, by multiplication of weight and unit price, both as indicated by the instrument. The device which performs the calculation is in any case considered a part of the instrument.

The interval of price to pay shall comply with the national regulations applicable to trade.

The unit price is restricted to Price/100 g or Price/kg.

Notwithstanding the provision in 4.4.1, the indications of weight, unit price and price to pay shall remain visible after the weight indication is stable, and after any introduction of the unit price, for at least one second and while the load is on the load receptor.

Notwithstanding the provision in 4.4.1, these indications may remain visible for no more than 3 seconds after removing the load, provided that the weight indication has been stable before and the indication would otherwise be zero. As long as there is a weight indication after removing the load, it shall not be possible to introduce or change a unit price.

If transactions performed by the instrument are printed, weight, unit price and price to pay shall all be printed.

The data may be stored in a memory of the instrument before printing. The same data shall not be printed twice on the ticket for the customer.

Instruments that can be used for price labelling purposes must comply with 4.17 as well.

4.15.4 Special applications of a price computing instrument

Only if all transactions performed by the instrument or by connected peripherals are printed on a ticket or label intended for the customer, a price computing instrument may perform additional functions which facilitate trade and management. These functions shall not lead to confusion about the results of weighing and price computing.

Other operations or indications not covered by the following provisions may be performed, provided that no indication which could possibly be misunderstood as a primary indication is presented to the customer.

4.15.4.1 Non-weighed articles

An instrument may accept and record positive or negative prices to pay of one or several non-weighed articles, provided that the weight indication is zero or the weighing mode is made inoperative. The price to pay for one or more of such articles shall be shown in the price-to-pay display.

If the price to pay is calculated for more than one equal articles, the number of articles shall be shown on the weight display, without being possibly taken for a weight, and the price for one article on the unit price display, unless supplementary displays are used to show the number of articles and the article price.

4.15.4.2 Totalization

An instrument may totalize transactions on one or several tickets; the price total shall be indicated on the price-to-pay display, and printed accompanied by a special word or symbol, either at the end of the price-to-pay column, or on a separate label or ticket with appropriate reference to the commodities whose prices to pay have been totalized; all prices to pay that are totalized shall be printed, and the price total shall be the algebraic sum of all these prices as printed.

An instrument may totalize transactions performed on other instruments linked to it, directly or over metrologically controlled peripherals, under the provisions of 4.15.4 and if the price-to-pay scale intervals of all connected instruments are identical.

4.15.4.3 Multi-vendor operation

An instrument may be designed to be used by more than one vendor or to serve more than one customer at the same time, provided that the connection between the transactions and the relevant vendor or customer is appropriately identified.

4.15.4.4 Cancellation

An instrument may cancel previous transactions. Where the transaction has already been printed, the relevant price to pay cancelled shall be printed with an appropriate comment. If the transaction to be cancelled is displayed to the customer it shall be clearly differentiated from normal transactions.

4.15.4.5 Additional information

An instrument may print additional information if this is clearly correlated to the transaction and does not interfere with the assignment of the weight value to the unit symbol.

4.15.5 Self-service instrument

A self-service instrument need not have two sets of scales or displays.

If a ticket or a label is printed, the primary indications shall include a designation of the product when the instrument is used to sell different products.

4.16 Instrument similar to one normally used for direct sales to the public

An instrument similar to one normally used for direct sales to the public which does not comply with the provisions of 4.14 and 4.15 shall carry, near the display, the indelible marking:

"Not to be used for direct sales to the public".

4.17 Price-labelling instrument

4.14.8, 4.15.3 (paragraphs 1 and 5), 4.15.4.1 (paragraph 1) and 4.15.4.5 apply.

A price-labelling instrument shall have at least one display for the weight. It may be used temporarily for set-up purposes such as supervision of setting weight limits, unit prices, preset tare values, commodity names.

It shall be possible to verify, during use of the instrument, the actual values of unit price and preset tare value.

Printing below minimum capacity shall not be possible.

Printing of labels with fixed values of weight, unit price and price to pay is allowed provided that the weighing mode is made inoperative.

4.18 Mechanical counting instrument with unit-weight receptor

For the purpose of verification a counting instrument is considered to be a semi-self-indicating instrument.

4.18.1 Indicating device

To permit verification, a counting instrument shall have a scale with at least one scale division $d = e$ on either side of zero; the corresponding value shall be shown on the scale.

4.18.2 Counting ratio

The counting ratio shall be shown clearly just above each counting platform or each counting scale mark.

5 Requirements for electronic instruments

In addition to clauses 3 and 4, an electronic instrument shall comply with the following requirements.

5.1 General requirements

5.1.1 An electronic instrument shall be designed and manufactured such that, when it is exposed to disturbances:

- either, (a) significant faults do not occur, or
- (b) significant faults are detected and acted upon. The indication of significant faults in the display should not be confusing with other messages that appear in the display.

Note: A fault equal to or smaller than e is allowed irrespective of the value of the error of indication.

5.1.2 The requirements in 3.5, 3.6, 3.8, 3.9 and 5.1.1 shall be met durably, in accordance with the intended use of the instrument.

5.1.3 A pattern of an electronic instrument is presumed to comply with the requirements in 5.1.1, 5.1.2 and 5.3.2 if it passes the examinations and tests specified in 5.4.

5.1.4 The requirements in 5.1.1 may be applied separately to:

- a) each individual cause of significant fault, and/or
- b) each part of the electronic instrument.

The choice, whether 5.1.1 (a) or 5.1.1 (b) is applied, is left to the manufacturer.

5.2 Acting upon significant faults

When a significant fault has been detected, the instrument shall either be made inoperative automatically or a visual or audible indication shall be provided automatically and shall continue until such time as the user takes action or the fault disappears.

5.3 Functional requirements

5.3.1 Upon switch-on (switch-on of indication), a special procedure shall be performed that shows all relevant signs of the indicator in their active and non-active state sufficiently long to be checked by the operator.

5.3.2 In addition to 3.9, an electronic instrument shall comply with the requirements under a relative humidity of 85 % at the upper limit of the temperature range. This is not applicable to an electronic instrument of class **I**, and of class **II** if e is less than 1 g.

5.3.3 Electronic instruments, class **I** instruments excepted, shall be subjected to the span stability test specified in 5.4.4. The error near maximum capacity shall not exceed the maximum permissible error and the absolute value of the difference between the errors obtained for any two measurements shall not exceed half the verification scale interval or half the absolute value of the maximum permissible error, whichever is greater.

5.3.4 When an electronic instrument is subjected to the disturbances specified in 5.4.3, the difference between the weight indication due to the disturbance and the indication without the disturbance (intrinsic error), shall not exceed e or the instrument shall detect and react to a significant fault.

5.3.5 During the warm-up time of an electronic instrument there shall be no indication or transmission of the weighing result.

5.3.6 An electronic instrument may be equipped with interfaces permitting the coupling of the instrument to any peripheral devices or other instruments.

An interface shall not allow the metrological functions of the instrument and its measurement data to be inadmissibly influenced by the peripheral devices (for example computers), by other interconnected instruments, or by disturbances acting on the interface.

Functions that are performed or initiated via an interface shall meet the relevant requirements and conditions of clause 4.

Note: An "interface" comprises all mechanical, electrical and logic properties at the data interchange point between an instrument and peripheral devices or other instruments.

5.3.6.1 It shall not be possible to introduce into an instrument, through an interface, instructions or data intended or suitable to:

- display data that are not clearly defined and could be mistaken for a weighing result,

- falsify displayed, processed or stored weighing results,
- adjust the instrument or change any adjustment factor; however instructions may be given through the interface to carry out an adjustment procedure using a span adjustment device incorporated inside the instrument or, for instruments in class **I**, using an external standard-mass,
- falsify primary indications displayed in case of direct sales to the public.

5.3.6.2 An interface through which the functions mentioned in 5.3.6.1 cannot be performed or initiated, need not be secured. Other interfaces shall be secured as per 4.1.2.4.

5.3.6.3 An interface intended to be connected to a peripheral device to which the requirements of this Recommendation apply, shall transmit data relating to primary indications in such a manner that the peripheral device can meet the requirements.

5.3.7 A battery-operated electronic instrument shall either continue to function correctly or not indicate any weight values whenever the voltage is below the manufacturer's specified value.

5.4 Performance and span stability tests

5.4.1 Test considerations

All electronic instruments of the same category, whether or not equipped with checking facilities, shall be subjected to the same performance test programme.

5.4.2 State of instrument under test

Performance tests shall be carried out on fully operational equipment in its normal operational state or in a status as similar as possible thereto. When connected in other than a normal configuration, the procedure shall be mutually agreed by the approval authority and the applicant and shall be described in the test document.

If an electronic instrument is equipped with an interface permitting the coupling of the instrument to external equipment, the instrument shall, during the tests B.3.2, B.3.3 and B.3.4, be coupled to external equipment, as specified by the test procedure.

5.4.3 Performance tests

Performance tests shall be performed according to B.2 and B.3.

Table 8

Test	Characteristic under test
Static temperatures	Influence factor
Damp heat, steady state	Influence factor
Power voltage variations	Influence factor
Short time power reductions	Disturbance
Bursts (transients)	Disturbance
Electrostatic discharge	Disturbance
Electromagnetic susceptibility	Disturbance

5.4.4 Span stability test

Span stability test shall be performed according to B.4.

6 Technical requirements for a non-self-indicating instrument

A non-self-indicating instrument shall comply with clauses 3 and 4, as far as applicable. This clause gives complementary provisions corresponding to some of the requirements of clause 4.

While the provisions of 6.1 are mandatory, those of 6.2 contain "acceptable solutions" as introduced in clause 4.

Provisions for certain simple instruments that may be submitted directly for initial verification are given in 6.3 through 6.9. These simple instruments are:

- simple equal arm and 1/10 ratio beams,
- simple steelyards with sliding poises,
- Roberval and Béranger instruments,
- instruments with ratio platforms,
- instruments of the steelyard type with accessible sliding poises.

6.1 Minimum sensitivity

An extra load equivalent to the absolute value of the maximum permissible error for the applied load, shall be placed on the instrument at equilibrium and shall cause a permanent displacement of the indicating element of at least:

- 1 mm for an instrument of class **I** or **II**,
- 2 mm for an instrument of class **III** or **III** with $Max \leq 30$ kg,
- 5 mm for an instrument of class **III** or **III** with $Max > 30$ kg.

The sensitivity tests shall be carried out by placing extra loads with a slight impact, in order to eliminate the effects of discrimination threshold.

6.2 Acceptable solutions for indicating devices

6.2.1 General provisions

6.2.1.1 Equilibrium indicating component

Indicating component of relative displacement in relation to another indicating component: the two indices should be of the same thickness and the distance between them shall not exceed this thickness.

However, this distance may be equal to 1 mm, if the thickness of the indices is less than this value.

6.2.1.2 Securing

It should be possible to secure the sliding poises, the removable masses and the adjusting cavities or the housings of such devices.

6.2.1.3 Printing

If the device permits printing, this should be possible only if sliding bars or poises or a mass switching mechanism are each in a position corresponding to a whole number of scale divisions. Except for accessible sliding poises or bars, printing should be possible only if the equilibrium indicating component is in the reference position to within the nearest half scale interval.

6.2.2 Sliding poise device

6.2.2.1 Form of scale marks

On bars on which the scale interval is the verification scale interval of the instrument, the scale marks should consist of lines of constant thickness. On other major (or minor) bars the scale marks should consist of notches.

6.2.2.2 Scale spacing

The distance between scale marks should not be less than 2 mm and be of sufficient length so that the normal machining tolerances for notches or scale marks do not cause an error in the weighing result exceeding 0.2 of the verification scale interval.

6.2.2.3 Stops

The displacement of sliding poises and minor bars should be limited to the graduated part of major and minor bars.

6.2.2.4 Indicating component

Each sliding poise should be provided with an indicating component.

6.2.2.5 Accessible sliding poise device

There should be no moving parts in sliding poises, except sliding minor bars.

There should be no cavity on sliding poises that could accidentally hold foreign bodies.

It should be possible to secure parts that are detachable.

The displacement of sliding poises and minor bars should require a certain effort.

6.2.3 Indication by use of metrologically controlled weights

The reduction ratios should be in the form 10^k , k being an integer or zero.

On an instrument intended for direct sales to the public, the height of the raised edge of the weights receptor platform should not exceed one tenth of the greatest dimension of the platform, without being more than 25 mm.

6.3 Conditions of construction

6.3.1 Equilibrium indicating component

An instrument shall be provided with two moving indices or one moving indicating component and a fixed datum mark, the respective position of which indicates the reference position of equilibrium.

On an instrument of class ③ or ④ designed to be used for direct sales to the public, the indices and scale marks shall allow equilibrium to be seen from the opposite sides of the instrument.

6.3.2 Knives, bearings and friction plates

6.3.2.1 Types of connection

Levers shall be fitted with knives only; these shall be pivoted on bearings.

The line of contact between knives and bearings shall be a straight line.

Counter-beams shall be pivoted on knife-edges.

6.3.2.2 Knives

The knives shall be fitted to the levers in such a way that the invariability of the ratios of the lever arms is assured. They shall not be welded or soldered.

The edges of the knives of one and the same lever shall be practically parallel and shall be situated in one plane.

6.3.2.3 Bearings

The bearings shall not be welded or soldered to their supports or in their mountings.

It shall be possible for bearings of an instrument with ratio platforms and steelyards to oscillate in all directions on their supports or in their mountings. On such instruments anti-disconnection devices shall prevent the disconnection of articulated parts.

6.3.2.4 Friction plates

The longitudinal play of the knives shall be limited by friction plates. There shall be point contact between knife and friction plates and it shall be situated on the extension of the line(s) of contact between knife and bearing(s).

The friction plate shall form a plane through the point of contact with the knife and its plane shall be perpendicular to the line of contact between knife and bearing. It shall not be welded or soldered to the bearings or their support.

6.3.3 Hardness

Contact parts of knives, bearings, friction plates, interlevers, interlever supports and links shall have a hardness of at least 58 Rockwell C.

6.3.4 Protective coating

A protective coating may be applied to the parts in contact of jointed components, provided that this does not lead to changes of metrological properties.

6.3.5 Tare devices

No instrument shall be fitted with a tare device.

6.4 Simple equal arm beam

6.4.1 Symmetry of the beams

The beam shall have two planes of symmetry, longitudinal and transversal. It shall be in equilibrium with or without the pans. Detachable parts which may be used equally well on either end of the beam shall be interchangeable and of equal mass.

6.4.2 Zero setting

If an instrument of class $\textcircled{\text{III}}$ or $\textcircled{\text{III}}$ is provided with a zero-setting device, this shall be a cavity below one of the pans.

This cavity may be secured.

6.5 Simple 1/10 ratio beam

6.5.1 Indication of the ratio

The ratio shall be indicated legibly and permanently on the beam in the form 1:10 or 1/10.

6.5.2 Symmetry of the beam

The beam shall have a longitudinal plane of symmetry.

6.5.3 Zero setting

The provisions of 6.4.2 apply.

6.6 Simple sliding poise instrument (steelyard)

6.6.1 General

6.6.1.1 Scale marks

The scale marks shall be lines or notches, either on the edge, or on the flat of the graduated shank.

The minimum scale spacing is 2 mm between notches and 4 mm between lines.

6.6.1.2 Pivots

The load per unit length on the knives shall be not more than 10 kg/mm.

The bores of bearings in the form of an annulus shall have a diameter at least equal to 1.5 times the largest dimension of the cross section of the knife.

6.6.1.3 Equilibrium indicating component

The length of the equilibrium indicating component, taken from the edge of the fulcrum knife-edge of the instrument, shall be not less than 1/15 of the length of the graduated part of the major sliding poise bar.

6.6.1.4 Distinctive mark

The head and the sliding poise of an instrument with detachable sliding poises shall bear the same distinctive mark.

6.6.2 Instrument with single capacity

6.6.2.1 Minimum distance between knife-edges

The minimum distance between knife-edges is:

25 mm for maximum capacities less than or equal to 30 kg,
20 mm for maximum capacities exceeding 30 kg.

6.6.2.2 Graduation

The graduation shall extend from zero to the maximum capacity.

6.6.2.3 Zero-setting

If an instrument of class  or  is provided with a zero-setting device, this shall be a captive screw or nut arrangement with a maximum effect of 4 verification scale intervals per revolution.

6.6.3 Instrument with dual capacity

6.6.3.1 Minimum distance between knife-edges

The minimum distance between the knife-edges is:

45 mm for the lower capacity,
20 mm for the higher capacity.

6.6.3.2 Differentiation of suspension mechanisms

The suspension mechanism of an instrument shall be differentiated from the load suspension mechanism.

6.6.3.3 Numbered scales

The scales corresponding to each of the capacities of the instrument shall permit weighing from zero to maximum capacity, without a break in continuity:

- either without the two scales having a common part,
- or with a common part of not more than 1/5 of the highest value of the lower scale.

6.6.3.4 Scale intervals

The scale intervals of each of the scales shall have a constant value.

6.6.3.5 Zero-setting devices

Zero-setting devices are not permitted.

6.7 Roberval and Béranger instruments

6.7.1 Symmetry

Detachable symmetrical parts occurring in pairs shall be interchangeable and of equal mass.

6.7.2 Zero-setting

If an instrument is provided with a zero-setting device, this shall be a cavity below the support of one of the pans. This cavity may be secured.


6.7.3 Length of the knife-edges

On an instrument having a simple beam:

- the distance between the outward ends of the load knife-edges shall be at least equal to the diameter of the bottom of the pan,
- the distance between the outward ends of the centre knife-edge shall be at least equal to 0.7 times the length of the load knife-edges.

A double beam instrument shall have a stability of the mechanism equal to that obtained with a simple beam instrument.

Figure 6



6.8 Instruments with ratio platforms

6.8.1 Maximum capacity

The maximum capacity of the instrument shall be greater than 30 kg.

6.8.2 Indication of the ratio

The ratio between the weighed load and the equilibrium load shall be indicated legibly and permanently on the beam in the form 1:10 or 1/10.

6.8.3 Zero-setting

An instrument shall have a zero-setting device consisting:

- either of a cup with greatly convex cover,
- or of a captive screw or nut arrangement, with a maximum effect of 4 verification scale intervals per revolution.

6.8.4 Complementary balancing device

If an instrument is provided with a complementary device that avoids the use of weights which are of low value in relation to the maximum capacity, this device shall be a graduated steelyard with a sliding poise, the effect being additive and not more than 10 kg.

6.8.5 Locking of the beam

An instrument shall have a manual device for locking the beam, the action of which prevents the equilibrium indices coinciding when at rest.

6.8.6 Provisions relating to wooden parts

If certain parts of an instrument, such as the frame, the platform or the board are of wood, this shall be dry and free from defects. It shall be covered with a paint or an effective protective varnish.

No nails shall be used for the final assembly of wooden parts.

6.9 Instrument with a load-measuring device with accessible sliding poises (of the steelyard type)

6.9.1 General

The provisions of 6.2 relating to load measuring devices with accessible sliding poises shall be observed.

6.9.2 Range of numbered scale

The numbered scale of the instrument shall permit continuous weighing from zero to the maximum capacity.

6.9.3 Minimum scale spacing

The scale spacing i_x of the different bars ($x = 1, 2, 3...$) corresponding to the scale interval d_x of these bars, shall be:

$$i_x \geq \frac{d_x}{e} \cdot 0.05 \text{ mm} \quad \text{but } i_x \geq 2 \text{ mm}$$

6.9.4 Ratio platform

If an instrument is provided with a ratio platform for extending the indicating range of the numbered scale, the ratio between the value of the weights placed on the platform to balance a load and the load itself shall be 1/10 or 1/100.

This ratio shall be indicated legibly and permanently on the beam in a position close to the ratio platform, in the form: 1:10, 1:100, or 1/10, 1/100.

6.9.5 Zero-setting

The provisions of 6.8.3 apply

6.9.6 Locking of the beam

The provisions of 6.8.5 apply.

6.9.7 Wooden parts

The provisions of 6.8.6 apply.

7 Marking of an instrument

7.1 Descriptive markings (*)

An instrument shall carry, in order, the following markings.

7.1.1 Compulsory in all cases:

- manufacturer's mark, or name written in full
- indication of accuracy class in the form of a Roman number in an oval (**):
 - for special accuracy 
 - for high accuracy 
 - for medium accuracy 
 - for ordinary accuracy 
- maximum capacity in the form Max ...
- minimum capacity in the form Min ...
- verification scale interval in the form e =

7.1.2 Compulsory if applicable:

- name or mark of manufacturer's agent for an imported instrument
- serial number
- identification mark on each unit of an instrument consisting of separate but associated units
- pattern approval mark

(*) Markings given by way of example but variable according to national regulations.

(**) See footnote to 3.1.1.

- scale interval, if $d < e$, in the form $d =$
- maximum additive tare effect in the form $T = + \dots$
- maximum subtractive tare effect if different from Max in the form $T = - \dots$ (*)
- maximum safe load in the form $Lim = \dots$
(if the manufacturer has provided for a maximum safe load of more than $Max + T$)
- the special temperature limits within which the instrument complies with the prescribed conditions of correct operation in the form: $\dots \text{ }^\circ\text{C}/\dots \text{ }^\circ\text{C}$
- counting ratio on a counting instrument according to 4.18 in the form $1:\dots$ or $1/\dots$
- ratio between weight platform and load platform in the specified in 6.5.1, 6.8.2 and 6.9.4,
- range of plus/minus indication of a digital comparator instrument, in the form $\pm \dots u_m$ or $-\dots u_m / + \dots u_m$, u_m standing for the unit of mass as per 2.1.

7.1.3 Additional markings

Additional markings may, if necessary, be required on an instrument according to its particular use or to certain special characteristics, such as e.g.:

- not to be used for direct sales to the public/commercial transactions
- to be used exclusively for:
- the stamp does not guarantee/guarantees only
- to be used only as follows:

7.1.4 Presentation of descriptive markings

The descriptive markings shall be indelible and of a size, shape and clarity allowing easy reading.

They shall be grouped together in a clearly visible place either on a descriptive plate fixed to an instrument, or on a part of the instrument itself.

The markings: Max ...
Min ...
e ...
and d if $d \neq e$

shall also be shown near the display of the result if they are not already located there.

It shall be possible to seal the plate bearing the descriptive markings unless its removal will result in its destruction. If the data plate is sealed, it shall be possible to apply a control mark to it.

Acceptable solutions

a) Markings in special cases

In special cases, some of the markings should be in the form of a table; see examples in Figure 7.

(*) Max may be also interpreted as the actual range of indication, as per 4.2.3.

Figure 7

For a multi-interval instrument	For an instrument with more than one weighing range (W_1, W_2)		For an instrument with weighing ranges in different classes	
	W_1	W_2	W_1 Ⓓ	W_2 Ⓓ
Max 2/5/15 kg	Max 20 kg	100 kg	Max 1 000 g	5 000 g
Min 20 g	Min 200 g	1 kg	Min 1 g	40 g
e = 1/2/5 g	e = 10 g	50 g	e = 0.1 g	2 g
			d = 0.02 g	2 g

b) Dimensions

When several plates are placed one above the other (as for example in the case of an instrument consisting of several separate devices), they should be of the same width. This common width is fixed at 80 mm.

c) Fixing

The plate should be fixed by rivets or screws with one of the rivets of red copper or material having qualities recognised as similar.

It should be possible to secure the head of one of the screws by means of a lead cap inserted in a device that cannot be dismantled. The diameter of the rivet head or of the lead cap should be able to accommodate a stamp 4 mm in diameter.

The plate may be glued or consist of a transfer provided that its removal results in its destruction.

d) Dimensions of the letters

The height of capital letters should be at least 2 mm.

7.1.5 Specific cases

7.1.1 through 7.1.4 apply in their entirety to a simple instrument made by one manufacturer.

When a manufacturer builds a complex instrument or when several manufacturers are involved in making a simple or complex instrument, the following additional provisions shall be applied.

7.1.5.1 Instrument having several load receptors and load measuring devices

Each load measuring device which is connected or can be connected to one or more load receptors, shall bear the descriptive markings relating to these, viz:

- identification mark,
- maximum capacity,
- minimum capacity,
- verification scale interval,

and, if appropriate, maximum safe load and maximum additive tare effect.

7.1.5.2 Instrument consisting of separately-built main parts

If main parts cannot be exchanged without altering the metrological characteristics of the instrument, each unit shall have an identification mark which shall be repeated in the descriptive markings.

7.2 Verification marks

7.2.1 Position

An instrument shall have a place for the application of verification marks.

This place shall:

- be such that the part on which it is located cannot be removed from the instrument without damaging the marks,
- allow easy application of the marks without changing the metrological qualities of the instrument,
- be visible without the instrument having to be moved when it is in service.

7.2.2 Mounting

An instrument required to bear verification marks shall have a verification mark support, at the place provided for above, which ensures the conservation of the marks:

- a) when the mark is made with a stamp, this support may consist of a strip of lead or any other material with similar qualities, inserted into a plate fixed to the instrument, or a cavity bored in the instrument.
- b) when the mark is of the self-adhesive type, a space shall be provided on the instrument for the application of this mark.

Acceptable solution

For application of the verification marks a stamping area of at least 200 mm² is required.

If transfers are used as verification marks the space for these transfers shall have a diameter of at least 25 mm.

8 Metrological controls

8.1 Liability to metrological controls

States may, through legislation, impose controls to assure that instruments used in specific applications comply with the requirements of this Recommendation. Controls may consist of pattern approval, initial verification, subsequent – e.g. periodic – verifications and in-service inspections. However, instruments according to 6.4 through 6.9 of this Recommendation shall not be subject to pattern approval, and national legislation may provide for initial verification without pattern approval for particular instrument applications.

8.2 Pattern approval

8.2.1 Application for pattern approval

The application for pattern approval shall include the submission to the approving authority of normally one instrument representative of the submitted pattern. Subject to agreement with the approving authority, the manufacturer may define and submit modules to be examined separately. This is particularly relevant in the following cases:

- where testing the instrument as a whole is difficult or impossible,
- where modules are manufactured and/or placed on the market as separate units to be incorporated in a complete instrument,
- where the applicant wants to have a variety of modules included in the approved pattern.

Acceptable solution

Typical modules are:

- load cells,
- electronic indicator, and
- connecting elements, both mechanical and electrical.

The following information and documents shall be provided by the applicant, as far as applicable.

8.2.1.1 Metrological characteristics:

- characteristics of the instrument, as per 7.1
- specifications of the modules or components of the measuring system, and when modules are submitted to be examined separately the fractions p_i of the error limits.

8.2.1.2 Descriptive documents:

- drawings of general arrangement and details of metrological interest including details of any interlocks, safeguards, restrictions, limits, etc.,
- a short functional description of the instrument,
- a short technical description including, if necessary, schematic diagrams of the method of operation in particular for internal processing and exchange via interface, of data and instructions. Adherence to requirements for which no test is available, such as software-based operations, may be demonstrated by a specific declaration of the manufacturer (e.g., for interfaces as per 5.3.6.1, and for password protected access to set-up and adjustment operations as per 4.1.2.4).

8.2.2 Pattern evaluation

The submitted documents shall be examined to verify compliance with the requirements of this Recommendation.

Suitable spotchecks shall be performed to establish confidence that the functions are performed correctly in accordance with the submitted documents. Reactions to significant faults need not be triggered.

The instruments shall be submitted to the testing procedures of Annex A, and of Annex B if applicable. If testing of the complete instrument is not possible, tests may, as agreed by the approving authority and the applicant, be performed:

- on a simulated set-up,
- on modules or main devices separately.

Where load cells are tested separately, the test equipment and preloading shall follow OIML R 60.

It may be feasible to perform the tests on premises other than those of the authority.

The approving authority may, in special cases, require the applicant to supply test loads, equipment and personnel to perform the tests.

The approving authorities are advised to consider the possibility of accepting, with the consent of the applicant, test data obtained from other national authorities, without repeating these tests (*).

Peripheral recipient devices need to be examined and tested only once while being connected to a weighing instrument, and may be declared as suitable for connection to any verified weighing instrument having an appropriate interface.

They may, at their discretion and under their responsibility, accept test data provided by the applicant for the submitted pattern, and reduce their own tests accordingly.

8.3 Initial verification

Initial verification shall not be performed unless conformity of the instrument to the approved pattern and/or the requirements of this Recommendation is established under the responsibility of either the manufacturer or the verifying authority. The instrument shall be tested at the time of installation and ready for use, unless it can be readily shipped and installed after initial verification.

8.3.1 Conformity

A declaration of conformity to the approved pattern and/or this Recommendation shall cover:

- correct functioning of all devices, e.g. zero-setting, tare, and calculating devices,
- construction material and design, as far as they are of metrological relevance.

8.3.2 Visual inspection

Before testing, the instrument shall be visually inspected for:

- metrological characteristics, i.e. accuracy class, Min, Max, e, d,
- prescribed inscriptions and positions for verification and control marks.

If location and conditions of use of the instrument are known, it should be considered whether they are appropriate.

8.3.3 Tests

Tests shall be carried out to verify compliance with the following requirements:

- 3.5.1, 3.5.3.3 and 3.5.3.4: errors of indication (refer to A.4.4 to A.4.6, but 5 loading steps are normally sufficient),

(*) This follows a Resolution adopted by the International Committee of Legal Metrology (CML) in 1986, recognizing the interest which knowledge of the results of tests, that may have been carried out by other national metrology services, could have for certain national metrology services to which a submission for pattern approval of the same instrument was made, and urging the CML Members to facilitate the exchange of such information, which in principle may only be provided with the agreement of the manufacturer, its representative or the importer of the instrument concerned.

- 4.6.2 and 4.7.3: accuracy of zero-setting and tare devices (refer to A.4.2.3 and A.4.6.2),
- 3.6.1: repeatability (refer to A.4.10, but normally no more than 3 weighings on classes Ⅲ and Ⅳ or 6 weighings on classes Ⅰ and Ⅱ are necessary),
- 3.6.2: eccentric loading (refer to A.4.7),
- 3.8: discrimination (refer to A.4.8).

Other tests may be performed in special cases, e.g. extraordinary construction or doubtful results.

The approving authority may, in special cases, require from the applicant to supply test loads, equipment and personnel to perform the tests.

For all tests, the error limits to be respected shall be the maximum permissible errors upon initial verification. If the instrument is to be shipped to another location after initial verification, the difference in local gravity acceleration between the locations of testing and use shall be considered if appropriate.

8.3.4 Stamping

According to national legislation, initial verification may be testified by verification marks. These marks may indicate the month or year when initial verification took place, or when reverification is due. National legislation may also require securing of components whose dismantling or maladjustment might alter the metrological characteristics of the instrument without the alterations being clearly visible.

8.4 Subsequent metrological control

8.4.1 Subsequent verification

Upon subsequent verification, normally only inspection and tests according to 8.3.2 and 8.3.3 shall be performed, the error limits being those on initial verification. Stamping and securing may take place according to 8.3.4, the date being that of the subsequent verification.

8.4.2 In-service inspection

Upon in-service inspection normally only inspection and tests according to 8.3.2 and 8.3.3 shall be performed, the error limits being twice those on initial verification. Stamping and securing may remain unchanged, or renewed as per 8.4.1.

ANNEX A
(mandatory)

TESTING PROCEDURES FOR
NONAUTOMATIC WEIGHING INSTRUMENTS

A.1 Administrative examination (8.2.1)

Review the documentation that is submitted, including necessary photographs, drawings, relevant technical specifications of main components, etc., to determine if it is adequate and correct. Consider the operational manual.

A.2 Compare construction with documentation (8.2.2)

Examine the various devices of the instrument to ensure compliance with the documentation.

A.3 Initial examination

A.3.1 Metrological characteristics

Note metrological characteristics according to the "Evaluation Report" (see R 76-2).

A.3.2 Descriptive markings (7.1)

Check the descriptive markings according to the check-list given in the Evaluation Report.

A.3.3 Stamping and securing (4.1.2.4 and 7.2)

Check the arrangements for stamping and securing according to the checklist given in the Evaluation Report.

A.4 Performance tests

A.4.1 General conditions

A.4.1.1 Normal test conditions (3.5.3.1)

Errors shall be determined under normal test conditions. When the effect of one factor is being evaluated, all other factors are to be held relatively constant, at a value close to normal.

A.4.1.2 Temperature

The tests shall be performed at a steady ambient temperature, usually normal room temperature unless otherwise specified.

The temperature is deemed to be steady when the difference between the extreme temperatures noted during the test does not exceed one-fifth of the temperature range of the given instrument without being greater than 5 °C (2 °C in the case of a creep test), and the rate of change does not exceed 5 °C per hour.

A.4.1.3 Power supply

Instruments using electric power shall normally be connected to the power supply and "on" throughout the tests.

A.4.1.4 Reference position before tests

For an instrument liable to be tilted, the instrument shall be levelled at its reference position.

A.4.1.5 Automatic zero-setting and zero-tracking

During the tests, the effect of the automatic zero-setting device or the zero-tracking device may be switched off or suppressed by starting the test with a load equal to say $10 e$.

In certain tests where the automatic zero-setting or zero-tracking must be in operation (or not), specific mention of this is made in those test descriptions.

A.4.1.6 Indication with a scale interval smaller than e

If an instrument with digital indication has a device for displaying the indication with a smaller scale interval (not greater than $1/5 e$), this device may be used to determine the error. If a device is used it should be noted in the Evaluation Report.

A.4.1.7 Using a simulator to test modules (3.5.4 and 3.7.1)

If a simulator is used to test a module, the repeatability and stability of the simulator should make it possible to determine the performance of the module with at least the same accuracy as when a complete instrument is tested with weights, the mpe to be considered being those applicable to the module. If a simulator is used, this shall be noted in the Evaluation Report and its traceability referenced.

A.4.1.8 Adjustment (4.1.2.5)

A semi-automatic span adjustment device shall be initiated only once before the first test.

An instrument of class **I** shall, if applicable, be adjusted prior to each test following the instructions in the operating manual.

Note: The temperature test A.5.3.1 is considered as one test.

A.4.1.9 Recovery

After each test the instrument should be allowed to recover sufficiently before the following test.

A.4.1.10 Preloading

Before each weighing test the instrument shall be pre-loaded once to Max or to Lim if this is defined, except for the tests in A.5.2 and A.5.3.2.

Where load cells are tested separately, the pre-loading shall follow OIML R 60.

A.4.1.11 Multiple range instrument

In principle, each range should be tested as a separate instrument.

A.4.2 Checking of zero

A.4.2.1 Range of zero-setting (4.5.1)

A.4.2.1.1 Initial zero-setting

With the load receptor empty, set the instrument to zero. Place a test load on the load receptor and switch the instrument off and then back on. Continue this process until, after placing a load on the load receptor and switching the instrument on and off, it does not re-zero. The maximum load that can be re-zeroed is the positive portion of the initial zero-setting range.

Remove any load from the load receptor and set the instrument to zero. Then remove the load receptor (platform) from the instrument. If, at this point, the instrument can be reset to zero by switching it off and back on, the mass of the load receptor is used as the negative portion of the initial zero-setting range.

If the instrument cannot be reset to zero with the load receptor removed, add weights to any live part of the scale (e.g. on the parts where the load receptor rests) until the instrument indicates zero again.

Then remove weights and, after each weight is removed, switch the instrument off and back on. The maximum load that can be removed while the instrument can still be reset to zero by switching it off and on is the negative portion of the initial zero-setting range.

The initial zero-setting range is the sum of the positive and negative portions. If the load receptor cannot readily be removed, only the positive part of the initial zero-setting range need be considered.

A.4.2.1.2 Nonautomatic and semi-automatic zero-setting

This test is performed in the same manner as described in A.4.2.1.1, except that the zero-setting means is used rather than switching the instrument off and on.

A.4.2.1.3 Automatic zero-setting

Remove the load receptor as described in A.4.2.1.1 and place weights on the instrument until it indicates zero.

Remove weights in small amounts and after each weight is removed allow time for the automatic zero-setting device to function so as to see if the instrument is reset to zero automatically. Repeat this procedure until the instrument will not reset to zero automatically.

The maximum load that can be removed so that the instrument can still be reset to zero, is the zero-setting range.

If the load receptor cannot readily be removed, a practical approach can be to add weights to the instrument and use another zero-setting device, if provided, to set the instrument to zero. Then remove weights and check whether the automatic zero-setting still sets the instrument to zero. The maximum load that can be removed so that the instrument can still be reset to zero is the zero-setting range.

A.4.2.2 Zero indicating device (4.5.5)

For instruments with digital indication and without a zero-tracking device, adjust the instrument to about one scale interval below zero; then by adding weights equivalent to say 1/10 of the scale interval, determine the range over which the zero indicating device indicates the deviation from zero.

A.4.2.3 Accuracy of zero-setting (4.5.2)

A.4.2.3.1 Nonautomatic and semi-automatic zero-setting

The accuracy of the zero-setting device is tested by first loading the instrument to an indication as close as possible to a changeover point, and then by initiating the zero-setting device and determining the additional load at which the indication changes from zero to one scale interval above zero. The error at zero is calculated according to the description in A.4.4.3.

A.4.2.3.2 Automatic zero-setting or zero-tracking

The indication is brought out of the automatic range (e.g. by loading with 10 e). Then the additional load at which the indication changes from one scale interval to the next above is determined and the error is calculated according to the description in A.4.4.3. It is assumed that the error at zero load would be equal to the error at the load in question.

A.4.3 Setting to zero before loading

For instruments with digital indication, the adjustment to zero, or the determination of the zero point is carried out as follows:

- a) for instruments with nonautomatic zero-setting, weights equivalent to half a scale interval are placed on the load receptor, and the instrument is adjusted until the indication alternates between zero and one scale interval. Then weights equivalent to half a scale interval are removed from the load receptor to attain a center of zero reference position;
- b) for instruments with semi-automatic or automatic zero-setting or zero-tracking, the deviation from zero is determined as described in A.4.2.3.

A.4.4 Determination of weighing performance

A.4.4.1 Weighing test

Apply test loads from zero up to and including Max, and similarly remove the test loads back to zero. When determining the initial intrinsic error, at least 10 different test loads shall be selected, and for other weighing tests at least 5 shall be selected. The test loads selected shall include Max and Min, and values at or near those at which the maximum permissible error (mpe) changes.

It should be noted that when loading or unloading weights the load shall be progressively increased or progressively decreased.

If the instrument is provided with an automatic zero-setting or zero-tracking device, it may be in operation during the tests, except for the temperature test. The error at zero point is then determined according to A.4.2.3.2.

A.4.4.2 Supplementary weighing test (4.5.1)

For instruments with an initial zero-setting device with a range greater than 20 % of Max, a supplementary weighing test shall be performed using the upper limit of the range as zero point.

A.4.4.3 Evaluation of error (A.4.1.6)

For instruments with digital indication and without a device for displaying the indication with a smaller scale interval (not greater than $1/5 e$), the changeover points are to be used to determine the indication of the instrument, prior to rounding, as follows.

At a certain load, L , the indicated value, I , is noted. Additional weights of say $1/10 e$ are successively added until the indication of the instrument is increased unambiguously by one scale interval ($I + e$). The additional load ΔL added to the load receptor gives the indication P , prior to rounding by using the following formula:

$$P = I + 1/2 e - \Delta L$$

The error prior to rounding is:

$$E = P - L = I + 1/2 e - \Delta L - L$$

The corrected error prior to rounding is:

$$E_c = E - E_o \leq mpe$$

where E_o is the error calculated at zero or at a load close to zero (e.g. $10 e$).

Example: an instrument with a scale interval, e , of 5 g is loaded with 1 kg and thereby indicates 1 000 g. After adding successive weights of 0.5 g, the indication changes from 1 000 g to 1 005 g at an additional load of 1.5 g. Inserted in the above formula these observations give:

$$P = (1\ 000 + 2.5 - 1.5) \text{ g} = 1\ 001 \text{ g}$$

Thus the true indication prior to rounding is 1 001 g, and the error is:

$$E = (1\ 001 - 1\ 000) \text{ g} = + 1 \text{ g}$$

If the changeover point at zero as calculated above was $E_o = + 0.5 \text{ g}$, the corrected error is:

$$E_c = + 1 - (+ 0.5) = + 0.5 \text{ g}$$

In the tests A.4.2.3 and A.4.11.1, the error shall be determined with a sufficient accuracy in view of the tolerance in question.

Note: The above description and formulae are valid also for multi-interval instruments. Where the load L and the indication I are in different partial weighing ranges:

- the additional weights ΔL are to be in steps of $1/10$ of e_i ,
- in the equation " $E = P - L = \dots$ " above, the term " $1/2 e$ " is to be $1/2 e_i$ or $1/2 e_{i+1}$ according to the partial weighing range in which the indication ($I + e$) is appearing.

A.4.4.4 Testing of modules

When testing modules separately, it shall be possible to determine the errors with a sufficiently small uncertainty considering the chosen fractions of the mpe either by using a device for displaying the indication with a scale interval smaller than $(1/5) \cdot p_i \cdot e$ or by evaluating the change-over point of the indication with an uncertainty better than $(1/5) \cdot p_i \cdot e$.

A.4.4.5 Weighing test using substitution material (3.7.3)

The test shall be carried out taking A.4.4.1 into account.

Check the repeatability error at a load of 50 % of Max and determine the allowed number of substitutions according to 3.7.3.

Apply test loads from zero up to and including maximum portion of standard weights.

Determine the error (A.4.4.3) and then remove the weights so that the no-load indication, or, in the case of an instrument with a zero-tracking device, the indication of say 10 e, is reached.

Substitute the previous weights with substitution material until the same changeover point, as used for the determination of the error, is reached. Repeat the above procedure until Max of the instrument is reached.

Unload in reverse order to zero, i.e. unload the weights and determine the changeover point. Place the weights back and remove the substitution material until the same changeover point is reached. Repeat this procedure until no-load indication.

Similar equivalent procedures may be applied.

A.4.5 Instrument with more than one indicating device (3.6.3)

If the instrument has more than one indicating device, the indications of the various devices shall be compared during the tests described in A.4.4.

A.4.6 Tare

A.4.6.1 Weighing test (3.5.3.3)

Weighing tests (loading and unloading according to A.4.4.1) shall be performed with at least two different tare values. At least 5 load steps shall be selected. The steps shall include values close to Min, the values at which the mpe changes and the value close to the maximum possible net load.

If the instrument is equipped with an additive tare device one of the weighing tests shall be performed with a tare value close to the maximum additive tare effect.

If the instrument is provided with automatic zero-setting or zero-tracking device it may be in operation during the test, in which case the error at zero point shall be determined according to A.4.2.3.2.

A.4.6.2 Accuracy of tare setting (4.6.3)

The accuracy of the tare device shall be established in a manner similar to the test described in A.4.2.3 with the indication set to zero using the tare device.

A.4.6.3 Tare weighing device (3.5.3.4 and 3.6.3)

If the instrument has a tare weighing device, the results obtained for the same load (tare), by the tare weighing device and the indicating device, shall be compared.

A.4.7 Eccentricity tests (3.6.2)

Large weights should be used in preference to several small weights. Smaller weights shall be placed on top of larger weights, but unnecessary stacking should be avoided within the segment to be tested. The load shall be applied centrally in the segment if a single weight is used, but applied uniformly over the segment, if several small weights are used.

The location of the load shall be marked on a sketch in the Evaluation Report.


The error at each measurement is determined according to A.4.4.3. The zero error E_0 used for the correction is the value determined prior to each measurement.

If the instrument is provided with automatic zero-setting or zero-tracking, it shall not be in operation during the following tests.

A.4.7.1 Instrument with a load receptor having not more than four points of support

The four quarter segments roughly equal to 1/4 of the surface of the load receptor (as per the sketches in Figure 8 or similar sketches) shall be loaded in turn.

Figure 8



A.4.7.2 Instrument with a load receptor having more than four points of support

The load shall be applied over each support on an area of the same order of magnitude as the fraction $1/n$ of the surface area of the load receptor, where n is the number of points of support.

Where two points of support are too close together for the above-mentioned test load to be distributed as indicated above, the load shall be doubled and distributed over twice the area on both sides of the axis connecting the two points of support.

A.4.7.3 Instrument with special load receptors (tank, hopper, etc.)

The load shall be applied to each point of support.

A.4.7.4 Instrument used for weighing rolling loads (3.6.2.4)

A rolling load shall be applied at different positions on the load receptor. These positions shall be at the beginning, the middle and at the end of the load receptor in the normal driving direction. The positions shall then be repeated in the reverse direction.

A.4.8 Discrimination test (3.8)

The following tests shall be performed with three different loads, e.g. Min, 1/2 Max and Max.


A.4.8.1 Non-self-indication and analogue indication

An extra load shall be placed gently on or removed from the load receptor while the instrument is at equilibrium. For certain extra load the equilibrium mechanism shall assume a different position of equilibrium, as specified.

A.4.8.2 Digital indication

A load plus sufficient additional weights (say 10 times $1/10 d$) shall be placed on the load receptor. The additional weights shall then be removed successively until the indication, I , is decreased unambiguously by one actual scale interval, $I - d$. One of the additional weights shall be replaced and a load equal to $1.4 d$ shall then be gently placed on the load receptor and give a result increased by one actual scale interval above the initial indication, $I + d$. See example in Figure 9.

Figure 9



The indication at the start is $I = 200 \text{ g}$. Remove additional weights until the indication changes to $I - d = 190 \text{ g}$. Add $1/10 d = 1 \text{ g}$ and thereafter $1.4 d = 14 \text{ g}$. The indication shall then be $I + d = 210 \text{ g}$.

A.4.9 Sensitivity of a non-self-indicating instrument (6.1)

During this test the instrument shall oscillate normally, and an extra load equal to the value of the mpe for the applied load shall be placed on the instrument while the load receptor is still oscillating. For damped instruments the extra load shall be

applied with a slight impact. The linear distance between the middle points of this reading and the reading without the extra load shall be taken as the permanent displacement of the indication. The test shall be performed with a minimum of two different loads (e.g. zero and Max).

A.4.10 Repeatability test (3.6.1)

Two series of weighings shall be performed, one with a load of about 50 % and one with a load close to 100 % of Max. For instruments with Max less than 1 000 kg each series shall consist of 10 weighings. In other cases each series shall consist of at least 3 weighings. Readings shall be taken when the instrument is loaded, and when the unloaded instrument has come to rest between weighings. In the case of a zero deviation between the weighings, the instrument shall be reset to zero, without determining the error at zero. The true zero position need not be determined between the weighings.

If the instrument is provided with automatic zero-setting or zero-tracking, it shall be in operation during the test.

A.4.11 Variation of indication with time (for instruments of class Ⅱ, Ⅲ or Ⅳ only)

A.4.11.1 Creep test (3.9.4.1)

Load the instrument close to Max. Take one reading as soon as the indication has stabilized and then note the indication while the load remains on the instrument for a period of four hours. During this test the temperature should not vary more than 2 °C.

The test may be terminated after 30 minutes if the indication differs less than 0.5 e during the first 30 minutes and the difference between 15 and 30 minutes is less than 0.2 e.

A.4.11.2 Zero return test (3.9.4.2)

The deviation in the zero indication before and after a period of loading with a load close to Max for half an hour, shall be determined. The reading shall be taken as soon as the indication has stabilized.

For multiple range instruments, continue to read the zero indication during the following 5 minutes after the indication has stabilized.

If the instrument is provided with automatic zero-setting or zero-tracking, it shall not be in operation.

A.4.12 Test for the stability of equilibrium (4.4.2)

Load the instrument up to 50 % of Max or up to a load included in the range of operation of the relevant function. Manually disturb the equilibrium by one single action and initiate the command for printing, data storage, or other function, as soon as possible. In the case of printing or data storage, read the indicated value 5 seconds after printing. In the case of zero-setting or tare balancing, check the accuracy as per A.4.2.3/A.4.6.2. Perform the test 5 times.

A.5 Influence factors

A.5.1 Tilting

The instrument shall be tilted both forwards and backwards longitudinally, and from side to side, transversely.

In the text that follows, class ② instruments intended for direct sales to the public are designated class ②* and class ② instruments not intended for direct sales to the public are designated class ②.

In practice the tests (no-load and loaded) described in A.5.1.1.1 and A.5.1.1.2 can be combined as follows.

After zero-setting in the reference position, the indication (prior to rounding) is determined at no load and at the two test loads. The instrument is then unloaded and tilted (without a new zero-setting), after which the indications at no load and at the two test loads are determined. This procedure is repeated for each of the tilting directions.

In order to determine the influence of tilting on the loaded instrument, the indication obtained at each tilt shall be corrected for the deviation from zero which the instrument had prior to loading.

If the instrument is provided with automatic zero-setting or zero-tracking, it shall not be in operation.

A.5.1.1 Tilting class ②, ③ and ④ instruments (3.9.1)

A.5.1.1.1 Tilting at no-load (class ②*, ③ and ④)

The instrument shall be set to zero in its reference position (not tilted). The instrument shall then be tilted longitudinally up to 2/1000 or the limiting value of the level indicator, whichever is greater. The zero indication is noted. The test shall be repeated with transverse tilting.

A.5.1.1.2 Tilting when loaded (class ②, ②*, ③ and ④)

The instrument shall be set to zero in its reference position and two weighings shall be carried out at a load close to the lowest load where the maximum permissible error changes, and at a load close to Max. The instrument is then unloaded and tilted longitudinally and set to zero. The tilting shall be 2/1000 or the limiting value of the level indicator, whichever is greater. Weighing tests as described above shall be performed. The test shall be repeated with transverse tilting.

A.5.1.2 Tilting class ① instrument (3.9.1.2)

The instrument shall be tilted longitudinally up to the limiting value of the level indicator. Check the tilt. Repeat with transverse tilting.

If the tilt is not greater than 2/1000, no further testing is required. Otherwise test as per A.5.1.1.2.

A.5.1.3 Instrument without level indicator

For an instrument liable to be tilted and not fitted with a level indicator the tests in A.5.1.1 shall be performed except that the instrument shall be tilted 5 % instead of 0.2 %.

A.5.2 Warm-up time test (5.3.5)

An instrument using electric power shall be disconnected from the supply for a period of at least 8 hours prior to the test. The instrument shall then be connected and

switched on and as soon as the indication has stabilized, the instrument shall be set to zero and the error at zero shall be determined. Calculation of error shall be made according to A.4.4.3. The instrument shall be loaded with a load close to Max. These observations shall be repeated after 5, 15 and 30 minutes. Every individual measurement performed after 5, 15, and 30 minutes, shall be corrected for the zero error at that time.

For instruments of class **I**, the provisions of the operating manual for the time following connection to the mains shall be observed.

A.5.3 Temperature tests

(see Figure 10 as a practical approach to performing the temperature tests)

A.5.3.1 Static temperatures (3.9.2.1 and 3.9.2.2)

The test consists of exposure of the equipment under test (EUT) to constant (*) temperatures within the range stated in 3.9.2, under free air conditions, for a 2 hour period after the EUT has reached temperature stability.

The weighing tests (loading and unloading) shall be carried out according to A.4.4.1:

- at a reference temperature (normally 20 °C but for class **I** instruments the mean value of the specified temperature limits),
- at the specified high temperature,
- at the specified low temperature,
- at a temperature of 5 °C, if the specified low temperature is below 10 °C, and
- at the reference temperature.

The change of temperature shall not exceed 1 °C/min during heating and cooling down.

For class **I** instruments, changes in barometric pressure shall be taken into account.

The absolute humidity of the test atmosphere shall not exceed 20 g/m³, unless the operating manual gives different specifications.

Reference to IEC Publications: see Bibliography /1/ (**)

A.5.3.2 Temperature effect on the no-load indication (3.9.2.3)

The instrument shall be set to zero and then changed to the prescribed highest and lowest temperatures as well as at 5 °C if applicable. After stabilisation the error of the zero indication shall be determined. The change in zero indication per 1 °C (class **I** instruments) or per 5 °C (other instruments) shall be calculated. The changes of these errors per 1 °C (class **I** instruments) or per 5 °C (other instruments) shall be calculated for any two consecutive temperatures of this test.

This test may be performed together with the temperature test (A.5.3.1). The errors at zero shall then be additionally determined immediately before changing to the next temperature and after the 2 hour period after the instrument has reached stability at this temperature.

(*) See A.4.1.2.

(**) See preliminary note to Annex B.

Note: Preloading is not allowed before these measurements.

If the instrument is provided with automatic zero-setting or zero-tracking, it shall not be in operation.

A.5.4 Voltage variations (3.9.3)

Stabilize the EUT under constant environmental conditions.

The test consists of subjecting the EUT to variations of AC mains voltage.

The test shall be performed with test loads of 10 e and a load between 1/2 Max and Max.

Test severity: Voltage variations: upper limit $V + 10\%$
lower limit $V - 15\%$

where V is the value marked on the instrument; if a range of voltages (V_{\min} , V_{\max}) is marked then the test shall be performed at $V_{\max} + 10\%$ and $V_{\min} - 15\%$.

Maximum allowable variations: All functions shall operate as designed.

All indications shall be within the maximum permissible errors.

Note: Where an instrument is powered by a three phase supply, the voltage variations shall apply for each phase successively.

If the instrument is provided with an automatic zero-setting device or a zero-tracking device, it may be in operation during the test, in which case the error at zero point shall be determined according to A.4.2.3.2.

A.6 Endurance test (3.9.4.3)

(applicable only to instruments of class **II**, **III** and **III** with $\text{Max} \leq 100 \text{ kg}$)


The endurance test shall be performed after all other tests.

Under normal conditions of use, the instrument shall be subjected to the repetitive loading and unloading of a load approximately equal to 50 % of Max. The load shall be applied 100 000 times. The frequency and speed of application shall be such that the instrument attains an equilibrium when loaded and when unloaded. The force of the load applied shall not exceed the force attained in a normal loading operation.

A weighing test in accordance with the procedure in A.4.4.1 shall be performed before the endurance test is started to obtain the intrinsic error. A weighing test shall be performed after the completion of the loadings to determine the durability error due to wear and tear.

If the instrument is provided with automatic zero-setting or zero-tracking device it may be in operation during the test, in which case the error at zero point shall be determined according to A.4.2.3.2.

Figure 10
 Proposed test sequence for test A.5.3.1 combined with A.5.3.2
 (temperature test where the temperature limits are + 40 °C / - 10 °C)



ANNEX B
(mandatory)

ADDITIONAL TESTS FOR ELECTRONIC INSTRUMENTS

Preliminary note: The tests which are specific to electronic instruments, as described in this Annex, have been taken as far as possible from the work of the International Electrotechnical Commission (IEC).

B.1 General requirements for electronic instruments under test (EUT)

Energize the EUT for a time period equal to or greater than the warm-up time specified by the manufacturer and maintain the EUT energized for the duration of the test.

Adjust the EUT as closely as practicable to zero prior to each test, and do not readjust it at any time during the test, except to reset it if a significant fault has been indicated. The deviation of the no-load indication due to any test condition shall be recorded, and any load indication shall be corrected accordingly to obtain the weighing result.

The handling of the instrument shall be such that no condensation of water occurs on the instrument.

B.2 Performance tests for influence factors

B.2.1 Static temperatures: see A.5.3.

B.2.2 Damp heat, steady state

(not applicable to class **I** instruments or class **II** instruments where e is less than 1 gram)

Test procedure in brief:

The test consists of exposure of the EUT to a constant (*) temperature and a constant relative humidity. The EUT shall be tested with at least five different test loads (or simulated loads):

- at the reference temperature (20 °C or the mean value of the temperature range whenever 20 °C is outside this range) and a relative humidity of 50 % following conditioning,
- at the high temperature of the range specified in 3.9.2 and a relative humidity of 85 %, two days following temperature and humidity stabilization, and
- at the reference temperature and relative humidity of 50 %.

(*) See A.4.1.2.

Maximum allowable variations: All functions shall operate as designed.
All indications shall be within maximum permissible errors.

Reference to IEC Publications: see Bibliography /2/

B.2.3 Power voltage variations: see A.5.4.

B.3 Performance tests for disturbances

Prior to any test, the rounding error shall be set as close as possible to zero.

If there are interfaces on the instrument, an appropriate peripheral device shall be connected to each different type of interface during the tests.

B.3.1 Short time power reductions

Test procedure in brief: Stabilize the EUT under constant environmental conditions.

A test generator capable of reducing the amplitude of one or more half cycles (at zero crossings) of the AC mains voltage shall be used. The test generator shall be adjusted before connecting the EUT. The mains voltage reductions shall be repeated ten times with an interval of at least 10 seconds.

The test shall be performed with one small test load.

Test severity:	Reduction	100 %	50 %
	Number of half cycles	1	2

Maximum allowable variations: The difference between the weight indication due to the disturbance and the indication without the disturbance shall either not exceed e or the instrument shall detect and react to a significant fault.

B.3.2 Bursts

The test consists in exposing the EUT to specified bursts of voltage spikes.

Test instrumentation: See IEC 801-4 (1988), N° 6

Test set-up: See IEC 801-4 (1988), N° 7

Test procedure: See IEC 801-4 (1988), N° 8

Before any test stabilize the EUT under constant environmental conditions.

The test shall be applied separately to:

- power supply lines,
- I/O circuits and communication lines, if any.

The test shall be performed with one small test load.

Test severity: Level 2 (see IEC 801-4 (1988), N° 5)
Open circuit output test voltage for:
– power supply lines: 1 kV,
– I/O signal, data and control lines: 0.5 kV.

Maximum allowable variations: The difference between the weight indication due to the disturbance and the indication without the disturbance shall either not exceed e or the instrument shall detect and react to a significant fault.

Reference to IEC Publications: see Bibliography /3/

B.3.3 Electrostatic discharge

The test consists in exposing the EUT to specified, direct and indirect, electrostatic discharges.

Test generator: See IEC 801-2 (1991), N° 6
Test set-up: See IEC 801-2 (1991), N° 7
Test procedure: See IEC 801-2 (1991), N° 8

This test includes the point penetration method, if appropriate. For direct discharges the air discharge shall be used where the contact discharge method cannot be applied.

Before any test stabilize the EUT under constant environmental conditions.

At least 10 direct discharges and 10 indirect discharges shall be applied. The time interval between successive discharges shall be at least 10 seconds.

The test shall be performed with one small test load.

Test severity: Level 3 (see IEC 801-2 (1991), N° 5)
DC voltage up to and including 6 kV for contact discharges and 8 kV for air discharges.

Maximum allowable variations: The difference between the weight indication due to the disturbance and the indication without the disturbance shall either not exceed e or the instrument shall detect and react to a significant fault.

Reference to IEC Publications: see Bibliography /4/

B.3.4 Immunity to radiated electromagnetic fields

Note: At the time of printing this Recommendation the IEC Publication 801-3 is still being revised (latest draft: 65A/77B (Secretariat) 121/88 dated July 1991). When performing this test, metrology authorities shall refer to the final text of IEC 801-3 when available, and before that to the most recent draft. A corrigendum to this Recommendation will be issued as soon as the final text of IEC 801-3 is available.

The test consists in exposing the EUT to specified electromagnetic fields.

Test equipment: See IEC..., N° 6
Test set-up: See IEC..., N° 7
Test procedure: See IEC..., N° 8

Before any test, stabilize the EUT under constant environmental conditions.

The EUT shall be exposed to electromagnetic fields of the strength and character as specified by the severity level.

The test shall be performed with one small test load only.

Test severity: Level 2 (see IEC..., N° 6)

Frequency range :	26 - 1 000	MHz
Field strength :	3	V/m
Modulation	:80 % AM, 1 kHz sine wave	

Maximum allowable variations: The difference between the weight indication due to the disturbance and the indication without the disturbance either shall not exceed e or the instrument shall detect and react to a significant fault.

Reference to IEC Publications: see Bibliography /5/

B.4 Span stability test

(not applicable to class **I** instruments)

test procedure in brief:

The test consist in observing the variations of the error of the EUT under sufficiently constant ambient conditions (reasonably constant conditions in a normal laboratory environment) at various intervals before, during and after the EUT has been subjected to performance tests.

The performance tests shall include the temperature test and, if applicable, the damp heat test; they shall not include any endurance test; other performance tests in Annexes A and B may be performed.

The EUT shall be disconnected from the mains power supply, or battery supply where fitted, two times for at least 8 hours during the period of the test. The number of disconnections may be increased if the manufacturer specifies so or at the discretion of the approval authority in the absence of any such specification.

For the conduct of this test the manufacturer's operating instructions shall be considered.

The EUT shall be stabilized at sufficiently constant ambient conditions after switch-on for at least 5 hours, but at least 16 hours after the temperature and damp heat tests have been performed.

Test duration:	28 days or the period necessary for the performance tests to be carried out, whichever is shorter.
Time between measurements:	Between 1/2 and 10 days, with a fairly even distribution of the measurements over the total duration of the test.
Test load:	Near Max; the same test weights shall be used throughout this test.
Number of measurements:	At least 8.
Test sequence:	<p>Stabilize all factors at sufficiently constant ambient conditions.</p> <p>Ajust the EUT as close to zero as possible.</p> <p>Automatic zero-tracking shall be made inoperative and automatic built-in span adjustment device shall be made operative.</p> <p>Apply the test weight(s) and determine the error.</p> <p>At the first measurement immediately repeat zeroing and loading four times to determine the average value of the error. For the next measurements perform only one, unless either the result is outside the specified tolerance or the range of the five readings of the initial measurement is more than 0.1 e.</p> <p>Record the following data:</p> <ul style="list-style-type: none"> a) date and time, b) temperature, c) barometric pressure, d) relative humidity, e) test load, f) indication, g) errors, h) changes in test location, <p>and apply all necessary corrections resulting from variations of temperature, pressure, etc. between the various measurements.</p> <p>Allow full recovery of the EUT before any other tests are performed.</p>
Maximum allowable variations:	<p>The variation in the errors of indication shall not exceed half the verification scale interval or half the absolute value of the maximum permissible error on initial verification for the test load applied, whichever is greater, on any of the n measurements.</p> <p>Where the differences of the results indicate a trend more than half the allowable variation specified above, the test shall be continued until the trend comes to rest or reverses itself, or until the error exceeds the maximum allowable variation.</p>

BIBLIOGRAPHY

References are given below to Publications of the International Electrotechnical Commission, IEC, of which mention is made in some of the tests in Annexes A and B.

- /1/ IEC Publication 68-2-1 (1974): Basic environmental testing procedures, Part 2: Tests, Test Ad: Cold, for heat dissipating equipment under test (EUT), with gradual change of temperature.
- IEC Publication 68-2-2 (1974): Basic environmental testing procedures, Part 2: Tests, Test Bd: Dry heat, for heat dissipating equipment under test (EUT), with gradual change of temperature.
- IEC Publication 68-3-1 (1974): Background information, Section 1: Cold and dry heat tests.
- /2/ IEC Publication 68-2-3 (1969): Basic environmental testing procedures, Part 2: Tests, Test Ca: Damp heat, steady state.
- IEC Publication 68-2-28 (1980): Guidance for damp heat tests.
- /3/ IEC Publication 801-4 (1988): Electromagnetic compatibility for industrial-process measurement and control equipment, Part 4: Electrical fast transients requirements.
- /4/ IEC Publication 801-2 (1991): Electromagnetic compatibility for industrial-process measurement and control equipment, Part 2: Electrostatic discharge requirements.
- /5/ IEC Publication...:
(see introductory note in B.3.4) Electromagnetic compatibility for industrial-process measurement and control equipment, Part 3: Radiated electromagnetic field requirements.

International Recommendation OIML R 76, edition 1992 (E)
NONAUTOMATIC WEIGHING INSTRUMENTS

INDEX OF TERMS DEFINED IN THE TERMINOLOGY

Actual scale interval	T.3.2.2
Analogue indication	T.5.1.2
Automatic zero-setting device	T.2.7.2.3
Auxiliary indicating devices	T.2.5
Auxiliary verification device	T.2.7.7
Balancing by weights	T.5.1.1
Calculated net value	T.5.3.2
Calculated total weight value	T.5.3.3
Complementary indicating device	T.2.5.3
Device for interpolation of reading	T.2.5.2
Digital indication	T.5.1.3
Discrimination	T.4.2
Disturbance	T.6.1.2
Durability	T.4.4
Durability error	T.5.5.7
Electronic component	T.2.3.3
Electronic device	T.2.3.1
Electronic instrument	T.1.2.6
Electronic sub-assembly	T.2.3.2
Error (of indication)	T.5.5.1
Extended indicating device	T.2.6
Extension interval of self-indication	T.3.1.5
Fault	T.5.5.5
Graduated instrument	T.1.2.1
Gross value	T.5.2.1
Indicating component	T.2.4.1
Indicating device	T.2.4
Indicating device with a differentiated scale division	T.2.5.4
Indication stabilizing device	T.2.7.9
Influence factor	T.6.1.1
Influence quantity	T.6.1
Initial intrinsic error	T.5.5.3
Initial zero-setting device	T.2.7.2.4
Instrument with price scales	T.1.2.7
Intrinsic error	T.5.5.2
Levelling device	T.2.7.1
Load-measuring device	T.2.1.3
Load receptor	T.2.1.1
Load-transmitting device	T.2.1.2
Locking device	T.2.7.6
Maximum capacity	T.3.1.1
Maximum permissible error	T.5.5.4
Maximum safe load	T.3.1.7
Maximum tare effect	T.3.1.6
Minimum capacity	T.3.1.2
Minimum reading distance	T.5.4.4
Module	T.2.2
Multi-interval instrument	T.3.2.6
Multiple range instrument	T.3.2.7

Net value	T.5.2.2
Nonautomatic weighing instrument	T.1.2
Nonautomatic zero-setting device	T.2.7.2.1
Non-graduated instrument	T.1.2.2
Non-self-indicating instrument	T.1.2.5
Number of verification scale intervals	T.3.2.5
Overall inaccuracy of reading	T.5.4.2
Performance test	T.7
Preset tare device	T.2.7.5
Preset tare value	T.5.3.1
Price-computing instrument	T.1.2.8
Price-labelling instrument	T.1.2.9
Primary indications	T.1.3.1
Rated operating conditions	T.6.2
Reading by simple juxtaposition	T.5.4.1
Reduction ratio	T.3.3
Reference conditions	T.6.3
Reference position	T.6.4
Repeatability	T.4.3
Rider	T.2.5.1
Rounding error or digital indication	T.5.4.3
Scale base	T.2.4.3
Scale interval of numbering	T.3.2.4
Scale mark	T.2.4.2
Scale spacing	T.3.2.1
Secondary indications	T.1.3.2
Selection device for load receptors and load-measuring devices	T.2.7.8
Self-indicating instrument	T.1.2.3
Self-indication capacity	T.3.1.3
Self-service instrument	T.1.2.10
Semi-automatic zero-setting device	T.2.7.2.2
Semi-self-indicating instrument	T.1.2.4
Sensitivity	T.4.1
Significant durability error	T.5.5.8
Significant fault	T.5.5.6
Span stability	T.5.5.9
Tare-balancing device	T.2.7.4.1
Tare device	T.2.7.4
Tare value	T.5.2.3
Tare-weighing device	T.2.7.4.2
Verification scale interval	T.3.2.3
Warm-up time	T.4.5
Weighing instrument	T.1.1
Weighing range	T.3.1.4
Weighing results	T.5.2
Zero-setting device	T.2.7.2
Zero-tracking device	T.2.7.3