



OIML CERTIFICATION SYSTEM (OIML-CS)

Test Laboratory Capability

Non-automatic weighing instruments – OIML R 76:2006

Name of Test Laboratory:	NMIJ/AIST			
	Class I	Class II	Class III	Class IIII
Maximum test capacity (kg)	10	100	300	300
Maximum number of verification scale intervals (Max/e)	1 000 000	100 000	10 000	1000
Minimum verification scale intervals (g)	0.001	0.001	0.1	5

Note: The use of a modular approach for testing may lead to higher capabilities.