



OIML CERTIFICATION SYSTEM (OIML-CS)

Test Laboratory Capability

Non-automatic weighing instruments – OIML R 76:1992

Name of Test Laboratory:	NMIA			
	Class I	Class II	Class III	Class IIII
Maximum test capacity (kg)	-	12	3000	3000
Maximum number of verification scale intervals (Max/e)	-	100 000	4000	1000
Minimum verification scale intervals (g)	-	0.1	1	10

Note: The use of a modular approach for testing may lead to higher capabilities.

Restrictions:

--