



OIML CERTIFICATION SYSTEM (OIML-CS)

Test Laboratory Capability

Automatic catchweighers – OIML R 51:2006

Test Laboratory	NMi Certin B.V.			
Accuracy class(es)	XI <input checked="" type="checkbox"/> Y(I) <input checked="" type="checkbox"/>	XII <input checked="" type="checkbox"/> Y(II) <input checked="" type="checkbox"/>	XIII <input checked="" type="checkbox"/> Y(a) <input checked="" type="checkbox"/>	XIII <input checked="" type="checkbox"/> Y(b) <input checked="" type="checkbox"/>
Maximum test capacity (kg)	50	50	600	600
Maximum number of verification scale intervals (Max/e)	500 000	100 000	10 000	1 000
Minimum verification scale interval (g)	0.001	0.001	0.1	5
Maximum speed of the belt (m/s)	Dependent upon the EUT			
Temperature range	Lower limit: -40 °C Upper limit: +60 °C			

Note: The use of a modular approach for testing may lead to higher capabilities.