



OIML CERTIFICATION SYSTEM (OIML-CS)

OIML Test Laboratory Capability

Non-automatic weighing instruments – OIML R 76:2006

Name of Test Laboratory:		EVPU		
	Class I	Class II	Class III	Class IIII
Maximum test capacity (kg)	N/A	1,000	5000	5000
Maximum number of verification scale intervals (Max/e)	N/A	100,000	10,000	1,000
Minimum verification scale intervals (g)	N/A	0.001	0.1	5

Note: The use of a modular approach for testing may lead to higher capabilities.