



OIML CERTIFICATION SYSTEM (OIML-CS)

Test Laboratory Capability

Water meters - OIML R 49:2013

Name of Test Laboratory	Endress+Hauser Flowtec AG			
	Class 1		Class 2	
Minimum flow rate (l/h)	0,0018 m ³ /h		0,0018 m ³ /h	
Maximum flow rate (m ³ /h)	6 300		6 300	
Minimum pipe size (mm)	DN2		DN2	
Maximum pipe size (mm)	DN2400		DN2400	
Water temperature classes	<input checked="" type="checkbox"/> T30 <input type="checkbox"/> T30/70 <input checked="" type="checkbox"/> T50 <input type="checkbox"/> T30/90 <input type="checkbox"/> T70 <input type="checkbox"/> T30/130 <input type="checkbox"/> T90 <input type="checkbox"/> T30/180 <input type="checkbox"/> T130 <input type="checkbox"/> T180	Water temperature classes	<input checked="" type="checkbox"/> T30 <input checked="" type="checkbox"/> T50 <input type="checkbox"/> T70 <input type="checkbox"/> T90 <input type="checkbox"/> T130 <input type="checkbox"/> T180	
Environmental tests (climatic, electromagnetic compatibility, magnetic field, etc.)	E2, H3, M3 acc. OIML D11:2013 class B, O and E2 acc. OIML R49:2013			