

OIML CERTIFICATION SYSTEM (OIML-CS)

OIML Test Laboratory Capability

Non-automatic weighing instruments – OIML R 76:1992

Name of Test Laboratory:		National Institute of Metrology, China		
	CI. I	CI W	CI III	CI III
	Class I	Class II	Class III	Class IIII
Maximum test capacity (kg)	60	3000	3000	3000
Maximum number of verification scale intervals (Max/e)	1,000,000	100,000	10,000	1,000
Minimum verification scale intervals (g)	0.001	0.001	0.1	5

Note: The use of a modular approach for testing may lead to higher capabilities.