



# OIML CERTIFICATION SYSTEM (OIML-CS)

## OIML Issuing Authority Capability

### Water meters - OIML R 49:2013

OIML Issuing Authority	CMI			
	Class 1		Class 2	
Minimum flow rate (l/h)	2		2	
Maximum flow rate (m <sup>3</sup> /h)	1 500		1 500	
Minimum pipe size (mm)	DN6		DN6	
Maximum pipe size (mm)	DN600		DN600	
Water temperature classes	<input checked="" type="checkbox"/> T30	<input checked="" type="checkbox"/> T30/70	<input checked="" type="checkbox"/> T30	<input checked="" type="checkbox"/> T30/70
	<input checked="" type="checkbox"/> T50	<input checked="" type="checkbox"/> T30/90	<input checked="" type="checkbox"/> T50	<input checked="" type="checkbox"/> T30/90
	<input checked="" type="checkbox"/> T70	<input type="checkbox"/> T30/130	<input checked="" type="checkbox"/> T70	<input type="checkbox"/> T30/130
	<input checked="" type="checkbox"/> T90	<input type="checkbox"/> T30/180	<input checked="" type="checkbox"/> T90	<input type="checkbox"/> T30/180
	<input type="checkbox"/> T130		<input type="checkbox"/> T130	
	<input type="checkbox"/> T180		<input type="checkbox"/> T180	
Environmental tests (climatic, electromagnetic compatibility, magnetic field, etc.)	Full range according to OIML R 49:2013			