



OIML CERTIFICATION SYSTEM (OIML-CS)

Test Laboratory Capability

Automatic catchweighers – OIML R 51:2006

| Test Laboratory | FORCE Technology (Horsholm) | | | |
|--|--|--|--|--|
| Accuracy class(es) | XI <input type="checkbox"/> Y(I) <input type="checkbox"/> | XII <input type="checkbox"/> Y(II) <input type="checkbox"/> | XIII <input checked="" type="checkbox"/> Y(a) <input checked="" type="checkbox"/> | XIII <input checked="" type="checkbox"/> Y(b) <input checked="" type="checkbox"/> |
| Maximum test capacity (kg) | N/A | N/A | 40 000 | 40 000 |
| Maximum number of verification scale intervals (Max/e) | N/A | N/A | 10 000 | 1 000 |
| Minimum verification scale interval (g) | N/A | N/A | 0.1 | 5 |
| Maximum speed of the belt (m/s) | Dependent upon the EUT | | | |
| Temperature range | Lower limit: - 10 °C Upper limit: + 55 °C | | | |