

Title: Revision of R 148:2020 *Non-invasive non-automated sphygmomanometers*

Part 3: Test report format

Project Group: OIML TC 18/SC 1/p 5

Convenership: P.R. China

Conveners: Ms Can Wang

Contents

Foreword	3
Introduction	4
Applicability of this test report format	5
Test report format	8
1 Test review.....	8
1.1 Summary of test results for type approval	8
1.2 Summary of test results for verification.....	9
2 Maximum permissible errors of the cuff pressure indication <u>under ambient conditions</u> ..	11
3 Maximum permissible error of the cuff pressure indication under storage conditions.....	13
4 Maximum permissible errors of the cuff pressure indication under varying temperature conditions	15
5 Air leakage of the pneumatic system.....	17
6 Pressure reduction rate for deflation valves.....	18
7 Rapid exhaust.....	18
8 Pressure indicating devices	19
8.1 Nominal range and measurement range.....	19
8.2 Analogue indication.....	19
8.3 Digital indication	21
9 Additional technical requirements for mercury manometers.....	22
9.1 Portable devices	22
9.2 Device to prevent mercury from being spilled during use and transport.....	22
9.3 Quality of the mercury	23
9.4 Graduation of the mercury tube	23
10 Additional requirements for aneroid manometers	24
10.1 Scale mark at zero.....	24
10.2 Zero.....	25
10.3 Pointer.....	25
10.4 Hysteresis error	25
10.5 Durability of the manometer.....	26
11 Mechanical safety	26
11.1 Resistance to shock for handheld sphygmomanometers.....	26
11.2 Non-automated sphygmomanometers used during patient transport.....	27
11.3 Non-automated sphygmomanometers containing a mercury manometer.....	27
12 Aborting a measurement	28
13 Unauthorised access and tamper proofing	28
14 Tubing connectors.....	29
15 Electrical safety.....	29
16 Durability of markings.....	30

Foreword

The International Organisation of Legal Metrology (OIML) is a worldwide, intergovernmental organisation whose primary aim is to harmonise the regulations and metrological controls applied by the national metrological services, or related organisations, of its Member States.

The main categories of OIML publications are:

- **International Recommendations (OIML R)**, which are model regulations that establish the metrological characteristics required of certain measuring instruments and which specify methods and equipment for checking their conformity. OIML Member States shall implement these Recommendations to the greatest possible extent;
- **International Documents (OIML D)**, which are informative in nature and which are intended to harmonise and improve work in the field of legal metrology;
- **International Guides (OIML G)**, which are also informative in nature and which are intended to give guidelines for the application of certain requirements to legal metrology; and
- **International Basic Publications (OIML B)**, which define the operating rules of the various OIML structures and systems.

OIML Draft Recommendations, Documents and Guides are developed by Project Groups linked to Technical Committees or Subcommittees which comprise representatives from the Member States. Certain international and regional institutions also participate on a consultation basis. Cooperative agreements have been established between the OIML and certain institutions, such as ISO and the IEC, with the objective of avoiding contradictory requirements. Consequently, manufacturers and users of measuring instruments, test laboratories, etc. may simultaneously apply OIML publications and those of other institutions.

International Recommendations, Documents, Guides and Basic Publications are published in English (E) and translated into French (F) and are subject to periodic revision.

Additionally, the OIML participates in Joint Committees with other Institutions for the development of **Vocabularies (OIML V)** and **Joint Guides (G)** and periodically commissions legal metrology experts to write **Expert Reports (OIML E)**. Expert Reports are intended to provide information and advice, and are written solely from the viewpoint of their author, without the involvement of a Technical Committee or Subcommittee, nor that of the CIML. Thus, they do not necessarily represent the views of the OIML.

This publication - reference OIML R 148-3, edition ~~2020-202x~~ (E) - was developed by OIML Technical Subcommittee TC 18/SC 1 *Blood pressure instruments*. It was approved for final publication by the International Committee of Legal Metrology in ~~2020-202x~~ and supersedes OIML R ~~148-3+6-1:2002-2020~~ (E).

OIML Publications may be downloaded from the OIML website in the form of PDF files. Additional information on OIML Publications may be obtained from the Organisation's headquarters:

Bureau International de Métrologie Légale
11, rue Turgot - 75009 Paris – France
Telephone: 33 (0)1 48 78 12 82
Fax: 33 (0)1 42 82 17 27
E-mail: biml@oiml.org
Internet: www.oiml.org

Non-invasive non-automated sphygmomanometers

Part 3: Test report format

Introduction ~~Explanatory notes on the test report format:~~

Formatted: R Head-1, Level 1, Space After: 18 pt

i) General

This test report format, which is informative with regard to the implementation of R 148-1 in national regulations, presents a standardised format for the results of the various tests and examinations to which a type of sphygmomanometer shall be submitted with a view to its approval as well as for the results of verification tests. The tests are listed in R 148-2.

It is recommended that all metrology services or laboratories evaluating types of sphygmomanometers according to OIML R 148 or to national or regional regulations based on OIML R 148 use this test report format, directly or after translation into a language other than English or French.

It is also recommended that this test report format in English or in French (or in both languages) be transmitted by the country performing these tests to the relevant authorities of another country, under bi- or multi-lateral cooperation agreements.

In the framework of the OIML Certification System (OIML-CS), use of the test report format is mandatory.

ii) Page numbering and the use of report page formats

In addition to the sequential numbering at the bottom of each page, a space has been left at the top of each page for numbering the pages of reports established following this model. In particular, each test is reported individually on a separate page following the relevant format. For a given report, it is advisable to complete the sequential numbering of each page by indicating the total number of pages in the report.

Where required, pressure values in the Tables can be replaced by values expressed in kPa.

Where required, these forms can be copied and used several times in cases where the test in question has to be repeated under varying conditions.

iii) Definitions and formula

For the purposes of this test report format, the following definitions and formula, taken from OIML V 2-200:2012 *International Vocabulary of Basic and General Terms in Metrology* (VIM) are used.

~~Non-invasive non-automated sphygmomanometers~~

~~Test report~~

Applicability of this test report format

Type approval test report	<input type="checkbox"/>
Verification test report	<input type="checkbox"/>

Note: For verification purposes, tick those fields which are appropriate for verification according to your national regulations or which are listed in 1.2 under the heading: Summary of test results for verification.

Application No.: Number of report:	
Object:	
Type:	
Serial number:	
Manufacturer's name and address:	

Applicant Customer's	name	and	
address:			

Date of receipt:	
Date/period of measurement:	
Date of report:	Number of pages:

Issuing Test laboratory Institute's	name	
and address:		

Formatted: Left

Formatted: Font: (Default) Times New Roman, Bold, English (United Kingdom)

Formatted: Font: (Default) Times New Roman, Bold, English (United Kingdom)

Formatted: Font: (Default) Times New Roman, Bold, English (United Kingdom)

Formatted: Left, Level 1

Formatted Table

Reference (number and year of edition)
to the relevant OIML Recommendation


Characteristic values (principle of measurement, measuring unit, measurement range,
range of display):

Additional devices (printer, interface, etc.):

Reference manometer (model, serial number, expanded uncertainty, calibration certificate):

Time measuring device (model, serial number, expanded uncertainty, calibration certificate):

Stamp/signature:



Formatted Table

Non-invasive non-automated sphygmomanometers

Test report format Non-invasive non-automated sphygmomanometers

Test report

1 Test review

1.1 Summary of test results for type approval

Formatted: Space After: 6 pt

Formatted: Level 1, Space After: 6 pt

Formatted: Level 1

Formatted: R Title, Space After: 0 pt, Line spacing: single

Formatted: Level 2

Formatted: Level 3

R 148-1	Clause from		Subject	Test result	OIML requirement	Passed	Failed
	R 148-2	R 148-3					
5.1	1	2	Maximum permissible errors of the cuff pressure indication <u>under ambient conditions</u>				
5.2	<u>32</u>	3	Maximum permissible errors of the cuff pressure indication under storage conditions				
5.3	<u>23</u>	4	Maximum permissible errors of the cuff pressure indication under varying temperature conditions				
6.2.1	4	5	Air leakage of the pneumatic system				
6.2.2	5	6	Pressure reduction rate for deflation valves				
6.2.3	6	7	Rapid exhaust				
6.3		8	Pressure indicating devices				
6.3.1		8.1	Nominal range and measurement range				
6.3.2		8.2	Analogue indication				
6.3.2.1		8.2.1	Scale				
6.3.2.2		8.2.2	First scale mark				
6.3.2.3		8.2.3	Scale interval				
6.3.2.4	7	8.2.4	Scale spacing and thickness of scale marks				
6.3.3		8.3	Digital indication				
6.4		9	Additional technical requirements for mercury manometers				
6.4.1		9.1	Portable devices				
6.4.2	8,9	9.2	Device to prevent mercury from being spilled during use and				

Formatted Table

			transport				
6.4.3		9.3	Quality of the mercury				
6.4.4		9.4	Graduation of the mercury tube				
6.5		10	Additional requirements for aneroid manometers				
6.5.1		10.1	Scale mark at zero				
6.5.2		10.2	Zero				
6.5.3		10.3	Pointer				
6.5.4	10	10.4	Hysteresis error				
6.5.5	11	10.5	Durability of manometer				
6.6.1	12	11	Mechanical safety				
6.6.1.1	12.1	11.1	Resistance to shock for handheld sphygmomanometers				
6.6.1.2	12.2	11.2	Non-automated sphygmomanometers used during patient transport				
6.6.1.3	12.3	11.3	Non-automated sphygmomanometers containing a mercury manometer				
6.6.2		12	Aborting a measurement				
6.6.3		13	Unauthorised access and tamper proofing				
6.6.5		14	Tubing connectors				
6.6.4		15	Electrical safety				
6.7	13	16	Durability of markings				

1.2 Summary of test results for verification

Formatted: Level 3

Clause from			Subject	Test result	OIML requirement	Pass ed	Failed ed
R 148-1	R 148-2	R 148-3					
5.1	1	2	Maximum permissible errors of the cuff pressure indication <u>under ambient conditions</u>				
6.2.1	4	5	Air leakage				

Note 1: The sequence of the different tests is arbitrary; it follows the sequence of the different clauses in the text. The sequence of testing is at the discretion of the person conducting the tests.

Note 2: To be considered as approved or verified, an instrument must have successfully passed all the applicable tests.

2 Maximum permissible errors of the cuff pressure indication under ambient conditions

Formatted: Level 2

For the limits of temperature and humidity see R 148-1, 5.1: the temperature shall be between 15 °C and 25 °C, the relative humidity shall be between 15 % and 85 %.

To determine the error of the cuff pressure indication, ~~conduct up and down runs~~ at three different temperatures: e.g. 15 °C and 15 % relative humidity, 20 °C and 60 % relative humidity, and 25 °C and 85 % relative humidity.

Table 1

Example: Temperature 20 °C and 60 % relative humidity Unit (mmHg)

Pressure	1st reading		2nd reading		Mean		Deviation Error		Hysteresis	
	Up	Down	Up	Down	Up	Down	Up	Down	1st reading	2nd reading
0	2	0	0	4	1	2	1	2	2	4
50	52	54	54	54	53	54	3	4	2	0
100	106	100	104	104	105	102	5	2	6	0
150										
200										
250										
column 1	column 2	column 3	column 4	column 5	column 6	column 7	column 8	column 9	column 10	column 11

maximum ~~deviation~~Error: 5 mmHg maximum hysteresis: 6 mmHg

Column 1 = values measured by the reference manometer

Column 2, 3, 4 and 5 = results of the measurement of the instrument under test

Column 6 = (column 2 + column 4) / 2

Column 7 = (column 3 + column 5) / 2

Column 8 = column 6 – column 1

Column 9 = column 7 – column 1

Column 10 = | column 2 – column 3 |

Column 11 = | column 4 – column 5 |

Table 2

Temperature °C and % relative humidity Unit (.....)

Pressure	1st reading		2nd reading		Mean		ErrorDeviation		Hysteresis	
	Up	Down	Up	Down	Up	Down	Up	Down	1st reading	2nd reading

Formatted: Centered

maximum
~~error~~deviation: maximum hysteresis:

Note 4: Up runs are conducted for the hysteresis error of aneroid manometers with the requirement of R 148-1, 6.5.4, or enter "/" if not necessary. The hysteresis error is the absolute value of the difference between the indications of the instrument when the same pressure is reached by increasing or decreasing the pressure.

Note 2: The time between the up and down runs should not be less than five minutes at the maximum pressure. A time difference from the first run to the second run of one hour is recommended.

Is the maximum ~~deviation error~~ of all the cuff pressure indicationall the readings of the instrument under test and the reference manometer less than or equal to ±0.4 kPa (±3 mmHg)?

Yes ☐ Passed ☐
No ☐ Failed ☐

Date of the test						
Serial number						
Test apparatus used for the test						
The environmental conditions	At start			At end		
	T (°C)	RH %		T (°C)	RH %	

Name of the technician						

Formatted Table

3 **Maximum permissible error of the cuff pressure indication under storage conditions**

Formatted: Level 2

Determine the error after storage for 24 h at a low temperature ~~of -20 °C~~ and for 24 h at a high temperature ~~of 70 °C~~ and a relative humidity of 85 % (non-condensing).

☐ after storage at - 20 °C and 70 °C (without electronic components)

Formatted: List Paragraph, Bulleted + Level: 1 + Aligned at: 0 cm + Indent at: 0.63 cm

☐ after storage at - 5 °C and 50 °C (with electronic components)

Table 3

Measurement at °C and % relative humidity ~~after storage at -20 °C and 70 °C~~ Unit
t (.....)

Formatted

Formatted: Right, Right 2.56 ch

<u>Maximum deviation:</u>				
<div>Pressure 1st reading 2nd reading Mean Error</div>				
<u>Pressure</u>	<u>1st reading</u>	<u>2nd reading</u>	<u>Mean</u>	<u>Error</u>

Formatted Table

Formatted Table

maximum error:

Formatted: Indent: First line: 3.22 ch, Space Before: 0 pt

Is the maximum ~~error of all the cuff pressure indication~~~~deviation of all the readings of the instrument~~
~~under test and the reference manometer~~ less than or equal to ± 0.4 kPa (± 3 mmHg)?

Yes ☐Passed ☐No ☐Failed ☐

Date of the test						
Serial number						
Test apparatus used for the test						
The environmental conditions	At start			At end		
	<i>T</i> (°C)	<i>RH</i> %		<i>T</i> (°C)	<i>RH</i> %	
Name of the technician						

4 Maximum permissible errors of the cuff pressure indication under varying temperature conditions

For each of the following combinations of temperature and humidity, condition the device for at least 3 h in the climatic chamber to allow the device to reach steady conditions.

10 °C ambient temperature, 85 % relative humidity (non-condensing);

20 °C ambient temperature, 85 % relative humidity (non-condensing);

40 °C ambient temperature, 85 % relative humidity (non-condensing).

Table 4

Temperature °C and % relative humidity ____ Unit (.....)

Pressure	1st-reading		2nd-reading		Mean		Deviation	
	Up	Down	Up	Down	Up	Down	Up	Down

Maximum deviation:

Pressure	1st reading	2nd reading	Mean	DifferenceE FFOF

maximum differenceerror:

Table 5

Temperature °C and % relative humidity ____ Unit (.....)

Formatted Table

Formatted: Indent: First line: 3.22 ch

Pressure	1st-reading		2nd-reading		Mean		Deviation	
	Up	Down	Up	Down	Up	Down	Up	Down

Maximum deviation:

Note: Repeat the same procedure for the set of conditions: 20 °C ambient temperature, 85 % relative humidity (non-condensing) and 40 °C ambient temperature, 85 % relative humidity (non-condensing).

Is the maximum difference between the indicated pressure of the manometer of the device to be tested and the corresponding readings of the reference manometer at the relevant temperature value less than or equal to ± 0.4 kPa (± 3 mmHg), or ± 2 % of the reading, whichever is greater?

Yes ☐

Passed ☐

No ☐

Failed ☐

Date of the test						
Serial number						
Test apparatus used for the test						
The environmental conditions	At start			At end		
	<i>T</i> (°C)	<i>RH</i> %		<i>T</i> (°C)	<i>RH</i> %	
Name of the technician						

5 **Air leakage of the pneumatic system**

Carry out the test over the whole measurement range and at at least three equally spaced pressure steps (e.g. 6.7 kPa (50 mmHg), 20.0 kPa (150 mmHg), and 33.3 kPa (250 mmHg)). Test the air leakage rate over a period of five minutes (see R 148-2, 4.2), and determine the measured value from this. Wait at least 60 s before reading each value.

Table 65

Air leakage of the pneumatic system Unit (.....)

Pressure	First reading	Reading after 5 min	Difference between the readings

Does the pressure drop over a period of five minutes correspond to an air leakage rate less than or equal to 0.5 kPa/min (4 mmHg/min)?

Yes ☐

Passed ☐

No ☐

Failed ☐

Date of the test						
Serial number						
Test apparatus used for the test						
The environmental conditions	At start			At end		
	<i>T</i> (°C)	<i>RH</i> %		<i>T</i> (°C)	<i>RH</i> %	
Name of the technician						

6 **Pressure reduction rate for deflation valves**

Formatted: Level 2

Are the pneumatic system deflation valves capable of adjustment to a deflation rate of 0.3 kPa/s to 0.4 kPa/s (2 mmHg/s to 3 mmHg/s)?

Yes ☐ Passed ☐
No ☐ Failed ☐

Are the pneumatic system deflation valves easily adjusted to these values?

Yes ☐ Passed ☐
No ☐ Failed ☐

Date of the test						
Serial number						
Test apparatus used for the test						
The environmental conditions	At start			At end		
	<i>T</i> (°C)	<i>RH</i> %		<i>T</i> (°C)	<i>RH</i> %	
	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>
Name of the technician						

7 **Rapid exhaust**

Formatted: Level 2

Time for the pressure reduction from 34.7 kPa to 2.0 kPa (260 mmHg to 15 mmHg):

t_{Re} = s

Is t_{Re} less than or equal to 10 s?

Yes ☐Passed ☐No ☐Failed ☐

Date of the test						
Serial number						
Test apparatus used for the test						
The environmental conditions	At start			At end		
	<i>T</i> (°C)	<i>RH</i> %		<i>T</i> (°C)	<i>RH</i> %	
Name of the technician						

8 Pressure indicating devices

8.1 Nominal range and measurement range

Nominal range:

measurement range:

Is the nominal range equal to the measurement range?

Does the nominal range extend from 0 kPa to at least 34.7 kPa (0 mmHg to at least 260 mmHg)?

Yes ☐Passed ☐No ☐Failed ☐

8.2 Analogue indication

8.2.1 Scale

Formatted: Level 2

Formatted: Level 3

Formatted: Space Before: 0 pt

Formatted: Level 3

Is the scale designed and arranged so that the measuring values can be read clearly and easily recognised?

Yes ☐Passed ☐No ☐Failed ☐

8.2.2 First scale mark

Does graduation begin with the first scale mark at 0 kPa (0 mmHg)?

Yes ☐Passed ☐No ☐Failed ☐

8.2.3 Scale interval

Is the scale interval 0.5 kPa or 2 mmHg for a scale graduated in kPa or mmHg, respectively?

Yes ☐Passed ☐No ☐Failed ☐

Is each fourth scale mark indicated by a greater length and is each eighth scale mark numbered in the case of a scale graduated in kPa?

Yes ☐Passed ☐No ☐Failed ☐

Is each fifth scale mark indicated by a greater length and is each tenth scale mark numbered in the case of a scale graduated in mmHg?

Yes ☐Passed ☐No ☐Failed ☐**8.2.4 Scale spacing and thickness of scale marks**

Is the distance between adjacent scale marks not less than 1.0 mm and does the thickness of the scale marks not exceed 20 % of the smallest scale spacing?

Yes ☐Passed ☐No ☐Failed ☐

Are the scale marks of equal thickness?

Yes ☐Passed ☐No ☐Failed ☐**8.3 Digital indication**

Is the digital scale interval 0.1 kPa (1 mmHg)?

If the measured value of a parameter is indicated on more than one display, do all the displays indicate the same numerical value?

Are the numbers and characters clearly legible?

Yes ☐Passed ☐No ☐Failed ☐

Formatted: Level 3

Date of the test						
Serial number						
Test apparatus used for the test						
The environmental conditions	At start			At end		
	<i>T</i> (°C)	<i>RH</i> %		<i>T</i> (°C)	<i>RH</i> %	
Name of the technician						

9 Additional technical requirements for mercury manometers

Formatted: Level 2

9.1 Portable devices

Formatted: Level 3

Does the portable device have an adjusting or locking mechanism to secure it in a specified position of use?

Yes ☐Passed ☐No ☐Failed ☐

9.2 Device to prevent mercury from being spilled during use and transport

Formatted: Level 3

Does the tube have a device to prevent the mercury from being spilled during transport and use and is this device efficient?

Yes ☐Passed ☐No ☐Failed ☐

Is the exhaust time taken for the mercury column to fall from 26.6 kPa to 5.3 kPa (from 200 mmHg to 40 mmHg) less than or equal to 1.5 s when the pressure in the system drops rapidly from 26.6 kPa to 0 kPa (from 200 mmHg to 0 mmHg)?

Yes ☐Passed ☐No ☐Failed ☐

9.3 Quality of the mercury

For type approval only:

Does the supplier confirm that the purity of the mercury is not less than 99.99 %?

Yes ☐Passed ☐No ☐Failed ☐

For type approval and for verification:

Does a visual inspection show a clean meniscus and no bubbles of air?

Yes ☐Passed ☐No ☐Failed ☐

9.4 Graduation of the mercury tube

Are the graduations permanently marked on the tube containing mercury?

Yes ☐Passed ☐No ☐Failed ☐

Formatted: Level 3

Formatted: Level 3

Date of the test						
Serial number						
Test apparatus used for the test						
The environmental conditions	At start			At end		
	<i>T</i> (°C)	<i>RH</i> %		<i>T</i> (°C)	<i>RH</i> %	
Name of the technician						

10 Additional requirements for aneroid manometers

Formatted: Level 2

10.1 Scale mark at zero

Formatted: Level 3

If a tolerance zone is shown at zero, is it less than or equal to ±0.4 kPa (±3 mmHg)? Is it clearly marked? Is a scale mark at zero indicated?

Yes	<input type="checkbox"/>	Passed	<input type="checkbox"/>
No	<input type="checkbox"/>	Failed	<input type="checkbox"/>

Note: Graduations within the tolerance zone are optional.

10.2 Zero

Formatted: Level 3

Is the movement of the elastic sensing element including the pointer unobstructed within 0.8 kPa (6 mmHg) below zero?

Yes ☐Passed ☐No ☐Failed ☐

Note: Neither the dial nor the pointer shall be adjustable by the user.

10.3 Pointer

Formatted: Level 3

Does the pointer cover between 1/3 and 2/3 of the length of the shortest scale mark of the scale?

Yes ☐Passed ☐No ☐Failed ☐

Is the pointer thinner than the scale mark at the place of indication and is the distance between the pointer and the dial less than or equal to 2 mm?

Yes ☐Passed ☐No ☐Failed ☐**10.4 Hysteresis error**

Formatted: Level 3

Is the maximum hysteresis error throughout the pressure range less than or equal to 0.5 kPa (4 mmHg) according to Table 2?

Yes ☐Passed ☐

No ☐Failed ☐

Note: The purpose of this test is to determine if the elastic sensing element has been exposed to a tension within the elastic range (i.e. the “Hooke’s” range) or not throughout the whole pressure range.

10.5 Durability of the manometer

Formatted: Level 3

Is the change in the pressure indication after 10 000 alternating pressure cycles less than or equal to 0.4 kPa (3 mmHg) throughout the pressure range?

Yes ☐Passed ☐No ☐Failed ☐

Date of the test						
Serial number						
Test apparatus used for the test						
The environmental conditions	At start			At end		
	<i>T</i> (°C)	<i>RH</i> %		<i>T</i> (°C)	<i>RH</i> %	
Name of the technician						

11 Mechanical safety

Formatted: Level 2

11.1 Resistance to shock for handheld sphygmomanometers

Formatted: Level 3

Does the instrument have adequate mechanical strength when subjected to mechanical stress caused by normal use, pushing, impact, dropping, and rough handling?

Yes ☐

Passed ☐

No ☐

Failed ☐

11.2 Non-automated sphygmomanometers used during patient transport

Formatted: Level 3

Does the instrument have adequate mechanical strength when subjected to mechanical stress caused by normal use, pushing, impact, dropping, and rough handling?

Yes ☐

Passed ☐

No ☐

Failed ☐

11.3 Non-automated sphygmomanometers containing a mercury manometer

Formatted: Level 3

Does the instrument comply with the requirements of R 148-1, 5.1 after testing?

Yes ☐

Passed ☐

No ☐

Failed ☐

Date of the test						
Serial number						
Test apparatus used for the test						
The environmental conditions	At start			At end		
	<i>T</i> (°C)	<i>RH</i> %		<i>T</i> (°C)	<i>RH</i> %	
Name of the technician						

12 **Aborting a measurement**

Formatted: Level 2

Is it possible to abort the blood pressure measurement at any time by activating the manual rapid exhaust valve, which shall be easily accessible?

Yes ☐ Passed ☐
No ☐ Failed ☐

13 **Unauthorised access and tamper proofing**

Formatted: Level 2

Are means provided to prevent tampering and unauthorised access?

Yes ☐ Passed ☐
No ☐ Failed ☐

Is it clear to an operator if tampering or authorised access has happened?

Yes ☐ Passed ☐
No ☐ Failed ☐

Date of the test						
Serial number						
Test apparatus used for the test						
The environmental conditions	At start			At end		
	<i>T</i> (°C)	<i>RH</i> %		<i>T</i> (°C)	<i>RH</i> %	
Name of the technician						

14 Tubing connectors

Are Luer locks used?

Yes ☐

Failed ☐

No ☐

Passed ☐

Date of the test						
Serial number						
Test apparatus used for the test						
The environmental conditions	At start			At end		
	<i>T</i> (°C)	<i>RH</i> %		<i>T</i> (°C)	<i>RH</i> %	
Name of the technician						

Formatted: Level 2

15 Electrical safety

Note: This test is optional within the OIML Certification System (OIML-CS)

If the instrument is equipped with electrical devices, do these devices comply with regional and national safety regulations?

Yes ☐

Passed ☐

No ☐

Failed ☐

Formatted: Level 2

16 Durability of markings

Formatted: Level 2

Do the required markings fulfil the requirements that they shall be removable only with a tool or by appreciable force and shall be sufficiently durable to remain clearly legible during the expected service life of the sphygmomanometer?

Yes ☐

Passed ☐

No ☐

Failed ☐

Date of the test						
Serial number						
Test apparatus used for the test						
The environmental conditions	At start			At end		
	<i>T</i> (°C)	<i>RH</i> %		<i>T</i> (°C)	<i>RH</i> %	
Name of the technician						