

INTERNATIONAL RECOMMENDATION

OIML R xxx-1
Edition 202x (E)

~~Contact~~-clinical thermometers

Part 1: Metrological and technical requirements

Thermomètres médicaux ~~à contact~~

Partie 1: Exigences métrologiques et techniques



Contents

1	Scope	5
2	Terminology	5
2.1	clinical thermometer.....	5
2.2	non-automated measurement mode	5
2.3	automated measurement mode	5
2.4	body site	5
2.5	probe.....	5
2.6	sensor.....	5
2.7	test mode.....	65
3	Description of the category of instrument	6
4	Units of measurement	6
5	Metrological requirements.....	6
5.1	Minimum measuring range.....	6
5.2	Maximum permissible errors.....	6
5.3	Thermal shock	7
5.4	Influence of immersion time.....	7
5.5	Storage.....	7
5.6	Relative humidity	7
5.7	Mechanical shock	7
5.8	Effect of voltage variations of the power source	7
5.9	Electromagnetic interference	87
5.10	Indication.....	8
6	Technical requirements	98
6.1	General	98
6.2	Test mode	98
6.3	Water	98
6.4	Alarms	98
6.5	Stamping.....	9
6.6	Cleaning and disinfection	9
6.7	Signal input and output ports.....	9
7	Metrological controls	9
7.1	Type approval.....	109
7.2	Verification.....	109
7.3	Sealing	109
7.4	Marking of the device.....	10
7.4.1	Markings required on the indicating unit	10
7.4.2	Markings required on the interchangeable probe	10
7.4.3	Markings required on the interchangeable probe	1140
7.5	Manufacturer's information.....	1140

Foreword

The International Organisation of Legal Metrology (OIML) is a worldwide, intergovernmental organisation whose primary aim is to harmonise the regulations and metrological controls applied by the national metrological services, or related organisations, of its Member States.

The main categories of OIML publications are:

International Recommendations (OIML R), which are model regulations that establish the metrological characteristics required of certain measuring instruments and which specify methods and equipment for checking their conformity. OIML Member States shall implement these Recommendations to the greatest possible extent;

International Documents (OIML D), which are informative in nature and which are intended to harmonise and improve work in the field of legal metrology;

International Guides (OIML G), which are also informative in nature and which are intended to give guidelines for the application of certain requirements to legal metrology; and

International Basic Publications (OIML B), which define the operating rules of the various OIML structures and systems.

OIML Draft Recommendations, Documents and Guides are developed by Project Groups linked to Technical Committees or Subcommittees which comprise representatives from the Member States. Certain international and regional institutions also participate on a consultation basis. Cooperative agreements have been established between the OIML and certain institutions, such as ISO and the IEC, with the objective of avoiding contradictory requirements. Consequently, manufacturers and users of measuring instruments, test laboratories, etc. may simultaneously apply OIML publications and those of other institutions.

International Recommendations, Documents, Guides and Basic Publications are published in English (E) and translated into French (F) and are subject to periodic revision.

Additionally, the OIML participates in Joint Committees with other Institutions for the development of **Vocabularies (OIML V)** and **Joint Guides (G)** and periodically commissions legal metrology experts to write **Expert Reports (OIML E)**. Expert Reports are intended to provide information and advice, and are written solely from the viewpoint of their author, without the involvement of a Technical Committee or Subcommittee, nor that of the CIML. Thus, they do not necessarily represent the views of the OIML.

This publication - reference OIML R XXX-1, edition 202x (E) - was developed by the OIML Technical Subcommittee TC 18/SC 2 *Medical thermometers*. It was approved for final publication by the International Committee of Legal Metrology in 202x and supersedes OIML R 7:1979 (E), R 114:1995(E) and R 115:1995(E).

OIML Publications may be downloaded from the OIML website in the form of PDF files. Additional information on OIML Publications may be obtained from the Organisation's headquarters:

Bureau International de Métrologie Légale
11, rue Turgot - 75009 Paris – France
Telephone: 33 (0)1 48 78 12 82
Fax: 33 (0)1 42 82 17 27
E-mail: biml@oiml.org Internet: www.oiml.org

~~Contact~~-eClinical thermometers

Part 1 - Metrological and technical requirements

1 Scope

This Recommendation specifies the metrological and technical requirements for clinical thermometers intended to measure temperature ~~through contact with the human or animal body~~of humans or animals.

This Recommendation applies to clinical thermometers with a maximum device or intended for continuous measurement.

This Recommendation applies to mechanical or electrical clinical thermometers but does not apply to secondary indicators, printing devices or other auxiliary devices.

This Recommendation does not apply to thermometers for premature babies, ~~or~~ ovulation thermometers or screening thermographs intended to be used for the individual non-invasive human febrile temperature screening of humans under indoor environmental conditions.

2 Terminology

2.1 clinical thermometer

medical measuring instrument used for indicating the temperature of a body indicated site

2.2 ~~non-automated measurement mode~~direct measurement mode

operating mode in which the clinical thermometer indicates the temperature of the ~~probe~~measurement site

Note: in this mode, the body indicated site is the same as the measurement site.

2.3 ~~automated~~indirect measurement mode

operating mode in which the clinical thermometer applies criteria or algorithms to estimate ~~body~~measurement site temperature without human intervention

Note 1: in this mode, the body indicated site may not be the same as the measurement site.

Note 2: Infrared clinical thermometers or clinical thermometers with claimed measurement time shorter than 60 seconds (for methods such as oral or rectal measurement), or shorter than 5 minutes (for methods such as axillary measurement) are typically in indirect measurement mode

2.4 ~~body~~measurement site

~~part of the patient~~body site that establishes thermal coupling with the clinical thermometer

2.5 probe

part of the clinical thermometer that provides a thermal coupling between the sensor and the patient

2.6 sensor

part of the clinical thermometer that converts thermal energy into an electrical signal

2.7 test mode

operating mode in which a clinical thermometer operates in ~~non-automated~~direct measurement mode and does not ~~turn-off without human intervention~~perform automatic switch off

2.8 body indicated site

body site to which the indication of clinical thermometer refers

3 Description of the category of instrument

Any contact clinical thermometer contains at least a sensing part and an output means. For mechanical thermometers, the sensor is usually a glass filled with liquid, and the output is a scale that can be marked directly on the glass (solid-stem type), or on a strip placed inside it (enclosed-scale type). When the thermometer meets the ~~body-measurement~~ site, the liquid moves inside the glass until thermal equilibrium is reached. In this situation, the temperature indication corresponds to the position of the meniscus on the scale.

Note: Some countries do not allow mercury to be used as a thermometric liquid.

For electrical thermometers, the sensor (e.g., thermistor or thermopile) is typically integrated into a probe that can be either interchangeable or permanently connected to the rest of the thermometer. To perform measurements, the probe is positioned at the ~~body-measurement~~ site to establish thermal coupling with the sensor, which then converts thermal energy into an electrical signal. An electronic circuit processes this signal and transmits it to an indicating unit appropriately. The temperature indication may either provide an estimate of the ~~body-measurement~~ site temperature with which the probe is or not in contact (~~automated-indirect measurement~~ mode), or the immediate ~~body-measurement~~ site temperature (~~non-automated~~direct measurement mode) or thermodynamic system (e.g. the thermostatic water bath ~~is an example~~or blackbody) with which the probe is in contact (test mode).

4 Units of measurement

The temperature shall be indicated in degree Celsius (°C).

Note: An alternative means for indicating temperature in degrees Fahrenheit (°F), may be used where permitted by national regulations.

5 Metrological requirements

5.1 Minimum measuring range

The measuring range shall be a minimum of 35.5 °C to 42.0 °C. Greater measuring ranges may be subdivided into partial ranges; however, the range 35.5 °C to 42.0 °C shall be continuous.

5.2 Maximum permissible errors

For any set of conditions within an ambient temperature range from 10 °C to 40 °C, relative humidity from 15 % to 85 % and, for electrical thermometers, voltage supply within $\pm 10\%$ of the nominal value of the mains voltage (or within the specified range of battery voltage), the maximum permissible error for the indication of ~~contact~~ clinical thermometer in test mode at any point of the minimum measuring range shall be as follows:

Table 1 – Maximum permissible errors

Instrument	Maximum permissible error
Mechanical contact -clinical thermometer	$\pm 0.15\text{ }^{\circ}\text{C}$
Electrical contact -clinical thermometer	$\pm 0.20\text{ }^{\circ}\text{C}$

Outside the minimum measuring range, the maximum permissible error may be twice the values indicated above.

5.3 Thermal shock

For electrical ~~contact~~-clinical thermometers, the maximum permissible error for the indication when ambient temperature changes abruptly from $-5\text{ }^{\circ}\text{C}$ to $+50\text{ }^{\circ}\text{C}$ shall comply with [Error! Reference source not found.5.2](#).

5.4 Influence of immersion time

For mechanical contact clinical thermometers at temperature t_1 ($15\text{ }^{\circ}\text{C} \leq t_1 \leq 30\text{ }^{\circ}\text{C}$), which is suddenly immersed in a well stirred water bath having a constant temperature t_2 ($35.5\text{ }^{\circ}\text{C} \leq t_2 \leq 42\text{ }^{\circ}\text{C}$) and is withdrawn after 20 seconds, the thermometer reading (R_1), after cooling to ambient temperature ($15\text{ }^{\circ}\text{C}$ to $30\text{ }^{\circ}\text{C}$) must comply with maximum permissible error requirements ([Error! Reference source not found.5.2](#))

Repeat the procedure at the same temperature set for the thermostatic bath (t_2), but withdrawn the thermometers after 60 seconds. The thermometer reading (R_2), after cooling to ambient temperature ($15\text{ }^{\circ}\text{C}$ to $30\text{ }^{\circ}\text{C}$) must comply with maximum permissible error requirements ([Error! Reference source not found.5.2](#)).

The module-absolute value of the difference between thermometer readings ($R_2 - R_1$) must be less than or equal to $0.005 \times (t_2 - t_1)$.

5.5 Storage

For electrical ~~contact~~-clinical thermometers, the maximum permissible error for the indication after storage for 24 h at $-20\text{ }^{\circ}\text{C} \pm 2\text{ }^{\circ}\text{C}$, followed by additional storage for 24 h at $60\text{ }^{\circ}\text{C} \pm 2\text{ }^{\circ}\text{C}$ (non-condensing) and cooling to ambient temperature shall be comply with [Error! Reference source not found.5.2](#).

5.6 Relative humidity

For electrical ~~contact~~-clinical thermometer, the maximum permissible error for the indication after 4 h at temperature range $\pm 2\text{ }^{\circ}\text{C}$ at an initial ambient temperature between $20\text{ }^{\circ}\text{C}$ and $32\text{ }^{\circ}\text{C}$ an ambient temperature between $20\text{ }^{\circ}\text{C}$ and $32\text{ }^{\circ}\text{C}$ and a relative humidity of 50 %, followed by 48 h at a relative humidity between 91 % and 95 %, and followed by 48 h in reference conditions shall comply with [Error! Reference source not found.5.2](#).

5.7 Mechanical shock

For electrical contact clinical thermometers, the maximum permissible error for the indication after falling on to a hard surface from a height of 1 m from three different orientations shall comply with [Error! Reference source not found.5.2](#).

A fixed contact clinical thermometer (e.g., wall mounted) is exempt from the requirements of this subclause.

5.8 Effect of voltage variations of the power source

Changes in the voltage within the working-nominal range of power source specified by the manufacturer shall not influence the temperature indication.

Outside this working range no temperature indication shall be displayed.

5.9 Electromagnetic interference

For electrical ~~contact~~ clinical thermometers, the maximum permissible error for the indication shall comply with **Error! Reference source not found.5.2** under each of the following conditions:

- when the thermometer is subjected to an electromagnetic field having a frequency between 150 kHz and 500 MHz with a field strength of 10 V/m;
- when the main supply of the thermometer is subjected to short duration power reductions, spikes and bursts. In this situation, if performance assessment during application of the transient is not possible, assessing performance before and after the test is acceptable;
- when at least ten repeated electrostatic discharges of 8 kV are applied to the thermometer casing or other accessible parts. In this situation, if performance assessment during application of the transient is not possible, assessing performance before and after the test is acceptable;
- when the thermometer is subjected to the field produced by high frequency surgical equipment (only applicable to thermometer intended to be used with high frequency surgical equipment).

5.10 Indication

The unit of measurement must accompany the indication.

The resolution shall be 0.1 °C at least.

For mechanical contact clinical thermometers, the scale must be clear and uniform, and must be engraved or printed clearly and indelibly.

For analogue scales, the distance between the marks must be at least 0.5 mm. The scale marks must be clear, straight and of thickness less than 0.25 times the distance between two consecutive scale marks. The marks lines must be perpendicular to the axis of the thermometer. The lines corresponding to degrees must be numbered, and the numbers must be clear and legible. The line corresponding to 37 °C may be indicated specially, using a different colour and/or by additional marking such as a dot, asterisk or arrow.

For digital indication, the display of temperature, excluding any transient displays during measurement, shall last at least one second, and shall be at least 4 mm in height, or shall be optically magnified to appear at least 4 mm in height.

The indicating unit shall include a self-checking device that meets the requirements of **Error! Reference source not found.5.2**. This device, which may be manual or automatic, shall input a predetermined electrical signal. Failure shall be clearly indicated.

Note: This device checks only the operation of the indicating unit and does not ensure that a temperature measurement is correct. It provides a means of detecting a faulty operation caused by a defective component or other disturbance.

Where the indicating device is fitted with a remote measuring transducer, its self-checking device shall periodically and automatically test the entire indicating unit at two or more values within the specified measuring range. The remote measuring transducer shall be fitted with a digitized output signal, and data transmission shall be verified by a checking device contained within the indicating unit.

6 Technical requirements

6.1 General

The thermometer must be free from any defects which might prevent it from operating normally, or which might lead to errors by its users.

6.2 Test mode

For ~~contact~~-clinical thermometers with an ~~automated-indirect~~ measurement mode, the instrument shall have a test mode which shall be accessed by the operator without additional tools, or shall have a direct measurement mode.

6.3 Water

Electrical ~~contact~~-clinical thermometers with non-interchangeable probes shall be water-resistant only if the manual states that they are resistant to immersion.

~~that probe is not interchangeable, shall be water resistant.~~

6.4 Alarms

Audible or visual alarms built into ~~contact~~-clinical thermometers shall not confuse the user.

~~Contact-e~~Clinical thermometers shall provide a clear indication or warning signal when the measured temperature is outside the specified measuring range.

For ~~contact~~-clinical thermometers with power supplied by a battery, the instrument shall provide a clear indication or warning signal when the battery voltage is outside the specified limits.

6.5 Stamping

Space for stamping must be provided on the stem of solid-stem thermometers and on the sheath of enclosed-scale thermometers.

6.6 Cleaning and disinfection

The clinical thermometer shall meet the requirements for maximum permissible errors when it has been subjected to the cleaning and disinfecting procedures specified by the manufacturer.

6.7 Signal input and output ports

The construction of the signal input and output ports relevant to the temperature measurement shall ensure that incorrectly fitted or defective accessories shall not result in erroneous indication of temperature.

7 Metrological controls

Regional or national regulations may prescribe type approval, initial and/or subsequent verification for ~~contact~~ clinical thermometers.

7.1 Type approval

At least three samples of a new type of ~~contact~~-clinical thermometer or its parts shall be tested, but national regulations may prescribe more samples.

The tests to verify conformity with metrological and technical requirements shall be carried out in accordance with R XXX-2. A test report shall be prepared according to R XXX-3.

For ~~contact~~-clinical thermometers with automated indirect measurement mode, the manufacturer (or responsible organization) must present the clinical investigation report carried out in accordance with ISO 12487:2026 for each type of interchangeable probe that is compatible.

7.2 Verification

After type approval has been granted, initial verification shall be carried out before the ~~contact~~-clinical thermometer or its parts are placed on the market.

Each instrument of an approved type of ~~contact~~-clinical thermometer in use shall be verified periodically in accordance with applicable metrological laws and regulations of a member state, or after repair.

During verification, testing can be conducted at any set of climatic conditions within the temperature range from 15 °C to 40 °C and a relative humidity range from 15 % to 85 %. At least the requirements of 5.2 shall be fulfilled.

7.3 Sealing

For ~~contact~~-clinical thermometers that are part of a patient-monitor, metrological control marks should prevent the manipulation of the metrologically relevant parts used for measuring temperature.

The metrological control marks shall be in place to prevent the opening of the casing. If the clinical thermometer is too small, the metrological control marks shall be put on the package.

7.4 Marking of the device

7.4.1 Markings required on the indicating unit

The indicating unit of the ~~contact~~-clinical thermometer shall be marked with the following information:

- name and/or trademark of the manufacturer;
- name and/or trademark of the responsible organisation;
- model designation of ~~contact~~-clinical thermometer;
- units of measurement, positioned close to the displayed values;
- measurement range;
- type approval number (if applicable);
- serial number (or lot number);
- country of origin.

7.4.2 Markings required on the interchangeable probe

The interchangeable probe shall be marked with the following information:

- name and/or trademark of the manufacturer;
- name and/or trademark of the responsible organisation;
- type of interchangeable probe;
- type of ~~contact~~-clinical thermometers that are compatible;
- type approval number (if applicable);
- serial number (or lot number);

7.4.3 Markings required on the interchangeable probe

The individual package of a ~~contact~~-clinical thermometer or its parts shall be marked with the following information:

- name and/or trademark of the manufacturer;
- name and/or trademark of the responsible organisation;
- type of ~~contact~~-clinical thermometer;
- thermometric liquid (if applicable);
- units of measurement, positioned close to the displayed values;
- measurement range;
- type approval number (if applicable);
- year of fabrication;
- country of origin.

7.5 Manufacturer's information

Information supplied by the manufacturer shall comply with the specifications and requirements given in this Recommendation. The manufacturer's instruction manual shall contain the following information:

- reference to **OIML R XXX** including the complete title;
- name and address of manufacturer;
- name and address of responsible organization;
- maximum permissible errors and results of clinical investigation;
- explanation of the operating procedures which are important for correct application (such as use of probe cover or the procedure to change the probe);
- description of all symbols, abbreviations and error codes used on the instrument;
- measurement range;
- methods for cleaning or disinfection;
- proper identification of interchangeable components and parts, such as temperature probe and batteries, including rated voltage, if applicable;
- explanation about self-checking device;
- information on suitable environmental conditions for use, storage and transport;
- information on the use of a single-use thermometer or temperature probe and the risks involved in case of reuse;
- nature and frequency of the maintenance which is required to ensure that the device always operates correctly and safely;
- disclosure that applicable national or regional metrological laws and regulations must be considered, if national regulation requires the disclosure;
- warm-up time, if applicable;
- troubleshooting, if applicable;
- specification of the signal input/output port(s);
- specification of the rated voltage of power source, if applicable;

Application software manuals and others manuals shall also comply with the specifications and requirements given in this Recommendation.